

Certificate of Conformity

Certificate Number: B-S171216288
Medical Devices Directive 93/42/EEC

Beide
Compliance Laboratory

Holder.....: SHENZHEN HONGXINYUAN ELECTRONIC CO.,LTD
Address.....: NO.1 -95 ZHIWUYUAN ROAD, NANYUE COMMUNITY
LONGGANG STREET, LONGGANG DISTRICT,
SHENZHEN CITY,CHINA
Manufacturer.....: SAME AS HOLDER
Address.....: SAME AS HOLDER
Product.....: oxygen concentrator
Model No.....: XY-1,XY-1S,XY-3,XY-3C,XY-3S,XY-5,XY-5S,XY-6,XY-6S
Technical Data....: 220-240V~,50Hz,0.8A,110W

The submitted products have been tested by us with the following standard(s) and found to be in compliance with the listed European Directives.

EN 60601-1:2006+A1:2013

The test results apply only to the particular sample tested and to the specific tests carried out. Technical Report and documentation are at the Holder's disposal.

This certificate applies specifically to the sample investigated in our test reference number only. The CE markings as shown below can be affixed on the product after preparation of necessary technical documentation. Other relevant Directives have to be observed.



Certification Manager
Date: Jan.08,2018



Company No.07113834



Beide (Shenzhen) Product Service Limited

China: 6F, Bldg E, Hourui 3rd Ind Zone, Xixiang, Bao'an Dist, Shenzhen, China

Http://www.szbeide.com E-mail: admin@szbeide.com