

STATEMENT

We, Shenzhen Biocare Bio-Medical Equipment Co., Ltd

Hereby declare that:

The Defibrillation Protection specification is 400 J for Electrocardiograph models iE 300, iE 6, and iE12A.
This information is confirmed by the manufacturer and can be provided upon request..

Duly authorized to sign this Authorization on behalf of: ***Shenzhen Biocare Bio-Medical Equipment Co., Ltd.***

Signed:

Name:

Title: Sales Director



Shenzhen Biocare Bio-Medical Equipment Co., Ltd.

Tel: +86-755-36615333

Fax: +86-755-27960643

Web: www.biocare.com.cn

Address: # 16-1, Jinhui Road, Jinsha Community, Kengzi Sub-District, Pingshan New District, 518122
Shenzhen, P.R. China.