

HiMedia Laboratories Private Limited

23, Vadhani Industrial Estate, L.B.S. Marg, Mumbai - 400086

Website : www.himedialabs.com, Email : info@himedialabs.com

Certificate of Analysis , Quality and Conformity

Material Code : SD155	Material Name : Vancomycin VA 5 mcg	Lot No : 0000411098
AR No.: 040000968913	Date of Report : 2019-11-15	Exp. Date : 2021-10

TEST	SPECIFICATIONS	RESULTS
<u>Appearance</u> 1 Appearance	Filter paper discs of 6 mm diameter with printed "VA 5" on centre of each side of the disc.	Complies
<u>Cultural response</u> 1 Cultural response	Average diameter of zone of inhibition observed on Mueller Hinton Agar (M173) after 18 hours incubation at 35-37°C for standard cultures.	
<u>Enterococcus faecalis ATCC 29212</u> 1 Zone of inhibition (mm)*	10mm - 16mm	15mm

The performance of this batch has been tested and standardized as per the EUCAST Standards. *QC Limit as per EUCAST Standard only.


STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets.

This document has been produced electronically and is valid



Pradnya G. Bhogle
Microbiologist/Sr.Executive Microbiologist



Asst./Dy/QC Manager



Dy/QA Manager