

HBcAb

**Competitive Enzyme Immunoassay for
the determination of antibodies
to Hepatitis B core Antigen
in human serum and plasma**

- for "in vitro" diagnostic use only -



DIA.PRO

**Diagnostic Bioprobes Srl
Via G. Carducci n° 27
20099 Sesto San Giovanni
(Milano) - Italy**

Phone +39 02 27007161

Fax +39 02 44386771

e-mail: info@diapro.it

REF. BCAB.CE
96 Tests

HBcAb

A. INTENDED USE

Competitive Enzyme ImmunoAssay (ELISA) for the determination of antibodies to Hepatitis B core Antigen in human plasma and sera.

The kit is intended for the screening of blood units and the follow-up of HBV-infected patients.

For "in vitro" diagnostic use only.

B. INTRODUCTION

The World Health Organization (WHO) defines Hepatitis B as follows:

"Hepatitis B is one of the major diseases of mankind and is a serious global public health problem. Hepatitis means inflammation of the liver, and the most common cause is infection with one of 5 viruses, called hepatitis A,B,C,D, and E. All of these viruses can cause an acute disease with symptoms lasting several weeks including yellowing of the skin and eyes (jaundice); dark urine; extreme fatigue; nausea; vomiting and abdominal pain. It can take several months to a year to feel fit again. Hepatitis B virus can cause chronic infection in which the patient never gets rid of the virus and many years later develops cirrhosis of the liver or liver cancer.

HBV is the most serious type of viral hepatitis and the only type causing chronic hepatitis for which a vaccine is available. Hepatitis B virus is transmitted by contact with blood or body fluids of an infected person in the same way as human immunodeficiency virus (HIV), the virus that causes AIDS. However, HBV is 50 to 100 times more infectious than HIV. The main ways of getting infected with HBV are: (a) perinatal (from mother to baby at the birth); (b) child-to-child transmission; (c) unsafe injections and transfusions; (d) sexual contact.

Worldwide, most infections occur from infected mother to child, from child to child contact in household settings, and from reuse of un-sterilized needles and syringes. In many developing countries, almost all children become infected with the virus. In many industrialized countries (e.g. Western Europe and North America), the pattern of transmission is different. In these countries, mother-to-infant and child-to-child transmission accounted for up to one third of chronic infections before childhood hepatitis B vaccination programmes were implemented. However, the majority of infections in these countries are acquired during young adulthood by sexual activity, and injecting drug use. In addition, hepatitis B virus is the major infectious occupational hazard of health workers, and most health care workers have received hepatitis B vaccine.

Hepatitis B virus is not spread by contaminated food or water, and cannot be spread casually in the workplace. High rates of chronic HBV infection are also found in the southern parts of Eastern and Central Europe. In the Middle East and Indian sub-continent, about 5% are chronically infected. Infection is less common in Western Europe and North America, where less than 1% are chronically infected.

Young children who become infected with HBV are the most likely to develop chronic infection. About 90% of infants infected during the first year of life and 30% to 50% of children infected between 1 to 4 years of age develop chronic

infection. The risk of death from HBV-related liver cancer or cirrhosis is approximately 25% for persons who become chronically infected during childhood.

Chronic hepatitis B in some patients is treated with drugs called *interferon* or *lamivudine*, which can help some patients. Patients with cirrhosis are sometimes given liver transplants, with varying success. It is preferable to prevent this disease with vaccine than to try and cure it.

Hepatitis B vaccine has an outstanding record of safety and effectiveness. Since 1982, over one billion doses of hepatitis B vaccine have been used worldwide. The vaccine is given as a series of three intramuscular doses. Studies have shown that the vaccine is 95% effective in preventing children and adults from developing chronic infection if they have not yet been infected. In many countries where 8% to 15% of children used to become chronically infected with HBV, the rate of chronic infection has been reduced to less than 1% in immunized groups of children. Since 1991, WHO has called for all countries to add hepatitis B vaccine into their national immunization programmes."

Hepatitis B core Antigen (or HBcAg) is the major component of the core particles of HBV.

HBcAg is composed of a single polypeptide of about 17 kD that is released upon disaggregating the core particles; the antigen contains at least one immunological determinant.

Upon primary infection, anti HBcAg antibodies are one of the first markers of HBV hepatitis appearing in the serum of the patient, slightly later than HBsAg, the viral surface antigen.

Anti HBcAg antibodies are produced usually at high titers and their presence is detectable even years after infection. Isolated HBcAb, in absence of other HBV markers, have been observed in infected blood units, suggesting the use of this test for screening HBV, in addition of HBsAg.

The determination of HBcAb has become important for the classification of the viral agent, together with the detection of the other markers of HBV infection, in sera and plasma.

C. PRINCIPLE OF THE TEST

The assay is based on the principle of competition where the antibodies in the sample compete with a monoclonal antibody for a fixed amount of antigen on the solid phase.

A purified recombinant HBcAg is coated to the microwells.

The patient's serum/plasma is added to the microwell together with an additive able to block interferences present in the sample.

In the second incubation after washing, a monoclonal antibody, conjugated with Horseradish Peroxidase (HRP) and specific for HBcAg is added and binds to the free rec-HBcAg coated on the plastic.

After incubation, microwells are washed to remove any unbound conjugate and then the chromogen/substrate is added. In the presence of peroxidase enzyme the colorless substrate is hydrolyzed to a colored end-product.

The color intensity is inversely proportional to the amount of antibodies to HBcAg present in the sample.

D. COMPONENTS

Each kit contains sufficient reagents to perform 96 tests.

1. Microplate **MICROPLATE**

8x12 microwell strips coated with recombinant HBcAg and sealed into a bag with desiccant. Allow the microplate to reach room temperature before opening; reseal unused strips in the bag with desiccant and store at 2.8°C.

2. Negative Control **CONTROL -**

1x1.0ml/vial. Ready to use. Contains 5% bovine serum albumin, 10 mM phosphate buffer pH 7.4 +/-0.1, 0.09% sodium azide and 0.045% ProClin 300 as preservatives. The negative control is pale yellow color coded.

3. Positive Control **CONTROL +**

1x1.0ml/vial. Ready to use. Contains 5% bovine serum albumin, anti HBcAg antibodies at a concentration of about 10 PEI U/ml, (calibrated on PEI HBc Reference Material 82), 10 mM phosphate buffer pH 7.4 +/-0.1, 0.09% sodium azide and 0.045% ProClin 300 as preservatives. The positive control is green color coded.

4. Calibrator **CAL ...**

n° 1 vial. Lyophilised. To be dissolved with EIA grade water as reported in the label. Contains fetal bovine serum, human antibodies to HBcAg at a concentration of 2 PEI U/ml +/-10% (calibrated on PEI HBc Reference Material 82) and 0.045% ProClin 300 as preservative.

Note: The volume necessary to dissolve the content of the vial may vary from lot to lot. Please use the right volume reported on the label.

5. Wash buffer concentrate **WASHBUF 20X**

1x60ml/bottle. 20x concentrated solution. Once diluted, the wash solution contains 10 mM phosphate buffer pH 7.0+/-0.2, 0.05% Tween 20 and 0.045% ProClin 300.

6. Enzyme Conjugate **CONJ**

1x16ml/vial. Ready-to-use solution. Contains 5% bovine serum albumine, 10 mM tris buffer pH 6.8 +/-0.1, Horseradish peroxidase conjugated mouse monoclonal antibody to HBcAg in presence of 0.3 mg/ml gentamicine sulphate and 0.045% ProClin 300. as preservatives. The component is red colour coded.

7. Chromogen/Substrate **SUBS TMB**

1x16ml/vial. Contains a 50 mM citrate-phosphate buffered solution at pH 3.6 +/-0.1, 0.03% tetra-methyl-benzidine (TMB), 0.02% hydrogen peroxide (H₂O₂) and 4% dimethylsulphoxide

Note: To be stored protected from light as sensitive to strong illumination.

8. Specimen Diluent **DILSPE**

4x3ml/vial. 10 mM tris buffered solution pH 8.0 +/-0.1 containing 0.045% ProClin 300 for the pre-treatment of samples and controls in the plate, blocking interference. The component is blue colour coded.

Note: Use all the content of one vial before opening a second one. The reagent is sensitive to oxidation.

9. Sulphuric Acid **H₂SO₄ 0.3 M**

1x15ml/vial. Contains 0.3 M H₂SO₄ solution. Attention: Irritant (H315; H319; P280; P302+P352; P332+P313; P305+P351+P338; P337+P313; P362+P363)

10. Plate sealing foil n° 2

11. Instruction manual n° 1

E. MATERIALS REQUIRED BUT NOT PROVIDED

1. Calibrated Micropipettes (100ul and 50ul) and disposable plastic tips.
2. EIA grade water (double distilled or deionised, charcoal treated to remove oxidizing chemicals used as disinfectants).
3. Timer with 60 minute range or higher.
4. Absorbent paper tissues.

5. Calibrated ELISA microplate thermostatic incubator (dry or wet) set at +37°C.
6. Calibrated ELISA microwell reader with 450nm (reading) and with 620-630nm (blinking) filters.
7. Calibrated ELISA microplate washer.
8. Vortex or similar mixing tools.

F. WARNINGS AND PRECAUTIONS

1. The kit has to be used by skilled and properly trained technical personnel only, under the supervision of a medical doctor responsible of the laboratory.
2. When the kit is used for the screening of blood units and blood components, it has to be used in a laboratory certified and qualified by the national authority in that field (Ministry of Health or similar entity) to carry out this type of analysis.
3. All the personnel involved in performing the assay have to wear protective laboratory clothes, talc-free gloves and glasses. The use of any sharp (needles) or cutting (blades) devices should be avoided. All the personnel involved should be trained in biosafety procedures, as recommended by the Center for Disease Control, Atlanta, U.S. and reported in the National Institute of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.
4. All the personnel involved in sample handling should be vaccinated for HBV and HAV, for which vaccines are available, safe and effective.
5. The laboratory environment should be controlled so as to avoid contaminants such as dust or air-borne microbial agents, when opening kit vials and microplates and when performing the test. Protect the Chromogen (TMB) from strong light and avoid vibration of the bench surface where the test is undertaken.
6. Upon receipt, store the kit at 2-8°C into a temperature controlled refrigerator or cold room.
7. Do not interchange components between different lots of the kits. It is recommended that components between two kits of the same lot should not be interchanged.
8. Check that the reagents are clear and do not contain visible heavy particles or aggregates. If not, advise the laboratory supervisor to initiate the necessary procedures.
9. Avoid cross-contamination between serum/plasma samples by using disposable tips and changing them after each sample. Do not reuse disposable tips.
10. Avoid cross-contamination between kit reagents by using disposable tips and changing them between the use of each one. Do not reuse disposable tips.
11. Do not use the kit after the expiration date stated on external (primary container) and internal (vials) labels.
12. Treat all specimens as potentially infective. All human serum specimens should be handled at Biosafety Level 2, as recommended by the Center for Disease Control, Atlanta, U.S. in compliance with what reported in the Institutes of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.
13. The use of disposable plastic-ware is recommended in the preparation of the washing solution or in transferring components into other containers of automated workstations, in order to avoid contamination.
14. Waste produced during the use of the kit has to be discarded in compliance with national directives and laws concerning laboratory waste of chemical and biological substances. In particular, liquid waste generated from the washing procedure, from residuals of controls and from samples has to be treated as potentially infective material and inactivated. Suggested procedures of inactivation are treatment with a 10% final concentration of household bleach for 16-18 hrs or heat inactivation by autoclave at 121°C for 20 min..
15. Accidental spills have to be adsorbed with paper tissues soaked with household bleach and then with water.

Tissues should then be discarded in proper containers designated for laboratory/hospital waste.

16. The Sulphuric Acid is an irritant. In case of spills, wash the surface with plenty of water.
17. Other waste materials generated from the use of the kit (example: tips used for samples and controls, used microplates) should be handled as potentially infective and disposed according to national directives and laws concerning laboratory wastes.

G. SPECIMEN: PREPARATION AND RECOMMENDATIONS

1. Blood is drawn aseptically by venepuncture and plasma or serum is prepared using standard techniques of preparation of samples for clinical laboratory analysis. No influence has been observed in the preparation of the sample with citrate, EDTA and heparin.
2. Avoid any addition of preservatives to samples; especially sodium azide as this chemical would affect the enzymatic activity of the conjugate.
3. Samples have to be clearly identified with codes or names in order to avoid misinterpretation of results. When the kit is used for the screening of blood units, bar code labeling and electronic reading is strongly recommended.
4. Haemolysed (red) and visibly hyperlipemic ("milky") samples have to be discarded as they could generate false results. Samples containing residues of fibrin or heavy particles or microbial filaments and bodies should be discarded as they could give rise to false results.
5. Sera and plasma can be stored at +2°...+8°C in primary collection tubes for up to five days after collection. Do not freeze primary tubes of collection. For longer storage periods, sera and plasma samples, carefully removed from the primary collection tube, can be stored frozen at -20°C for at least 12 months. Any frozen samples should not be frozen/thawed more than once as this may generate particles that could affect the test result.
6. If particles are present, centrifuge at 2.000 rpm for 20 min or filter using 0.2-0.8µ filters to clean up the sample for testing.

H. PREPARATION OF COMPONENTS AND WARNINGS

A study conducted on an opened kit has not pointed out any relevant loss of activity up to 6 re-uses of the device and up to 6 months.

1. Microplates:

Allow the microplate to reach room temperature (about 1 hr) before opening the container. Check that the desiccant has not turned dark green, indicating a defect in storage. In this case, call Dia.Pro's customer service. Unused strips have to be placed back inside the aluminum pouch, with the desiccant supplied, firmly zipped and stored at +2°...8°C. After first opening, remaining strips are stable until the humidity indicator inside the desiccant bag turns from yellow to green.

2. Negative Control:

Ready to use. Mix well on vortex before use.

3. Positive Control:

Ready to use. Mix well on vortex before use.

4. Calibrator:

Add the volume of ELISA grade water, reported on the label, to the lyophilised powder; let fully dissolve and then gently mix on vortex.

Note: The dissolved calibrator is not stable. Store it frozen in aliquots at -20°C.

5. Wash buffer concentrate:

The whole content of the concentrated solution has to be diluted 20x with bidistilled water and mixed gently end-over-end before use. During preparation avoid foaming as the presence of bubbles could impact on the efficiency of the washing cycles. **Note:** Once diluted, the wash solution is stable for 1 week at +2..8° C.

6. Enzyme conjugate:

Ready to use. Mix well on vortex before use.

Avoid contamination of the liquid with oxidizing chemicals, dust or microbes. If this component has to be transferred, use only plastic, and if possible, sterile disposable containers.

7. Chromogen/Substrate:

Ready to use. Mix well on vortex before use.

Avoid contamination of the liquid with oxidizing chemicals, air-driven dust or microbes. Do not expose to strong light, oxidizing agents and metallic surfaces.

If this component has to be transferred use only plastic, and if possible, sterile disposable container.

8. Specimen Diluent

Ready to use solution. Mix gently on vortex before use. Use all the content of one vial before opening a second one. The reagent is sensitive to oxidation.

9. Sulphuric Acid:

Ready to use. Mix well on vortex before use.

Attention: Irritant (H315; H319; P280; P302+P352; P332+P313; P305+P351+P338; P337+P313; P362+P363).

Legenda:

Warning H statements:

H315 – Causes skin irritation.

H319 – Causes serious eye irritation.

Precautionary P statements:

P280 – Wear protective gloves/protective clothing/eye protection/face protection.

P302 + P352 – IF ON SKIN: Wash with plenty of soap and water.

P332 + P313 – If skin irritation occurs: Get medical advice/attention.

P305 + P351 + P338 – IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

P337 + P313 – If eye irritation persists: Get medical advice/attention.

P362 + P363 – Take off contaminated clothing and wash it before reuse.

I. INSTRUMENTS AND TOOLS USED IN COMBINATION WITH THE KIT

1. Micropipettes have to be calibrated to deliver the correct volume required by the assay and must be submitted to regular decontamination (70% ethanol, 10% solution of bleach, hospital grade disinfectants) of those parts that could accidentally come in contact with the sample or the components of the kit. They should also be regularly maintained in order to show a precision of 1% and a trueness of ±2%.
2. The ELISA incubator has to be set at +37°C (tolerance of ±0.5°C) and regularly checked to ensure the correct temperature is maintained. Both dry incubators and water baths are suitable for the incubations, provided that the instrument is validated for the incubation of ELISA tests.
3. The **ELISA washer** is extremely important to the overall performances of the assay. The washer must be carefully validated in advance, checked for the delivery of the right

dispensation volume and regularly submitted to maintenance according to the manufacturer's instructions for use. In particular the washer, at the end of the daily workload, has to be extensively cleaned out of salts with deionized water. Before use, the washer has to be extensively primed with the diluted Washing Solution.

The instrument weekly has to be submitted to decontamination according to its manual (NaOH 0.1 M decontamination suggested).

5 washing cycles (aspiration + dispensation of 350ul/well of washing solution + 20 sec soaking = 1 cycle) are sufficient to ensure the assay with the declared performances. If soaking is not possible add one more cycle of washing.

An incorrect washing cycle or salt-blocked needles are the major cause of false positive reactions.

4. Incubation times have a tolerance of $\pm 5\%$.
5. The ELISA microplate reader has to be equipped with a reading filter of 450nm and with a second filter of (620-630nm, mandatory) for blanking purposes. Its standard performances should be (a) bandwidth ≤ 10 nm; (b) absorbance range from 0 to ≥ 2.0 ; (c) linearity to ≥ 2.0 ; repeatability $\geq 1\%$. Blanking is carried out on the well identified in the section "Assay Procedure". The optical system of the reader has to be calibrated regularly to ensure that the correct optical density is measured. It should be regularly maintained according to the manufacturer 's instructions.
6. When using an ELISA automated work station, all critical steps (dispensation, incubation, washing, reading, shaking, data handling) have to be carefully set, calibrated, controlled and regularly serviced in order to match the values reported in the sections "Validation of Test" and "Assay Performances". The assay protocol has to be installed in the operating system of the unit and validated as for the washer and the reader. In addition, the liquid handling part of the station (dispensation and washing) has to be validated and correctly set. Particular attention must be paid to avoid carry over by the needles used for dispensing samples and for washing. This must be studied and controlled to minimize the possibility of contamination of adjacent wells due to strongly reactive samples, leading to false positive results. The use of ELISA automated work stations is recommended for blood screening and when the number of samples to be tested exceed 20-30 units per run.
7. Dia.Pro's customer service offers support to the user in the setting and checking of instruments used in combination with the kit, in order to assure full compliance with the requirements described. Support is also provided for the installation of new instruments to be used with the kit.

L. PRE ASSAY CONTROLS AND OPERATIONS

1. Check the expiration date of the kit printed on the external label (primary container). Do not use if expired.
2. Check that the liquid components are not contaminated by visible particles or aggregates. Check that the Chromogen (TMB) is colourless or pale blue by aspirating a small volume of it with a sterile plastic pipette. Check that no breakage occurred in transportation and no spillage of liquid is present inside the box (primary container). Check that the aluminium pouch, containing the microplate, is not punctured or damaged.
3. Dilute all the content of the 20x concentrated Wash Solution as described above.
4. Dissolve the Calibrator as described above and gently mix.
5. Allow all the other components to reach room temperature (about 1 hr) and then mix gently on vortex all liquid reagents.
6. Set the ELISA incubator at $+37^{\circ}\text{C}$ and prepare the ELISA washer by priming with the diluted washing solution, according to the manufacturers instructions. Set the right number of washing cycles as reported in the specific section.

7. Check that the ELISA reader is turned on or ensure it will be turned on at least 20 minutes before reading.
8. If using an automated work station, turn on, check settings and be sure to use the right assay protocol.
9. Check that the micropipettes are set to the required volume.
10. Check that all the other equipment is available and ready to use.
11. In case of problems, do not proceed further with the test and advise the supervisor.

M. ASSAY PROCEDURE

The assay has to be performed according to the procedure given below, taking care to maintain the same incubation time for all the samples being tested.

1. Place the required number of strips in the plastic holder and carefully identify the wells for controls, calibrator and samples.
2. Leave the A1 well empty for blanking purposes.
3. Dispense 50 ul Specimen Diluent into all the control and sample wells.
4. Pipette 50 μl of the Negative Control in triplicate, 50 ul of the Calibrator in duplicate and then 50 ul of the Positive Control in single. Then dispense 50 ul of each of the samples.
5. Incubate the microplate for **60 min at $+37^{\circ}\text{C}$** .
Important note: *Strips have to be sealed with the adhesive sealing foil, only when the test is performed manually. Do not cover strips when using ELISA automatic instruments.*
6. When the first incubation is finished, wash the microwells as previously described (section I.3)
7. Pipette 100 μl Enzyme Conjugate in all the wells, except A1; incubate the microplate for **60 min at $+37^{\circ}\text{C}$** .

Important note: *Be careful not to touch the plastic inner surface of the well with the tip filled with the Enzyme Conjugate. Contamination might occur.*

8. When the second incubation is finished, wash the microwells as previously described (section I.3)
9. Pipette 100 μl Chromogen/Substrate into all the wells, A1 included.

Important note: *Do not expose to strong direct light. as a high background might be generated.*

10. Incubate the microplate protected from light at **room temperature ($18-24^{\circ}\text{C}$) for 20 minutes**. Wells dispensed with negative control and negative samples will turn from clear to blue (competitive method).
11. Pipette 100 μl Sulphuric Acid into all the wells using the same pipetting sequence as in step 9 to stop the enzymatic reaction. Addition of the stop solution will turn the negative control and negative samples from blue to yellow.
12. Measure the colour intensity of the solution in each well, as described in section I.5 using a 450nm filter (reading) and a 620-630nm filter (background subtraction, mandatory), blanking the instrument on A1.

Important notes:

1. *Ensure that no finger prints are present on the bottom of the microwell before reading. Finger prints could generate false positive results on reading.*
2. *Reading has should ideally be performed immediately after the addition of the Stop Solution but definitely no longer than 20 minutes afterwards. Some self oxidation of the chromogen can occur leading to a higher background.*
3. *The Calibrator (CAL) does not affect the cut-off calculation and therefore the test results calculation. The Calibrator*

may be used only when a laboratory internal quality control is required by the management.

N. ASSAY SCHEME

Specimen Diluent	50 ul
Controls&calibrator and samples	50 ul
1st incubation	60 min
Temperature	+37°C
Wash	n° 5 cycles with 20" of soaking OR n° 6 cycles without soaking
Enzyme Conjugate	100 ul
2nd incubation	60 min
Temperature	+37°C
Wash	n° 5 cycles with 20" of soaking OR n° 6 cycles without soaking
TMB/H2O2 mix	100 ul
3rd incubation	20 min
Temperature	r.t.
Sulphuric Acid	100 ul
Reading OD	450nm / 620-630nm

Problem	Check
Blank well > 0.050 OD450nm	that the Chromogen/Substrate solution has not become contaminated during the assay
Negative Control (NC) < 1.000 OD450nm after blanking coefficient of variation > 20%	1. that the washing procedure and the washer settings are as validated in the pre qualification study; 2. that the proper washing solution has been used and the washer has been primed with it before use; 3. that no mistake has been done in the assay procedure (dispensation of positive control instead of negative control); 4. that no contamination of the negative control or of the wells where the control was dispensed has occurred due to positive samples, to spills or to the enzyme conjugate; 5. that micropipettes have not become contaminated with positive samples or with the enzyme conjugate 6. that the washer needles are not blocked or partially obstructed.
Calibrator Co/S < 1	1. that the procedure has been correctly performed; 2. that no mistake has occurred during its distribution (ex.: dispensation of negative control instead of positive control); 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the calibrator has occurred.
Positive Control > 0.200 OD450nm	1. that the procedure has been correctly performed; 2. that no mistake has occurred during the distribution of the control (dispensation of negative control instead of positive control); 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the positive control has occurred.

If any of the above problems have occurred, report the problem to the supervisor for further actions.

An example of dispensation scheme is reported below:

Microplate

	1	2	3	4	5	6	7	8	9	10	11	12
A	BLK	S2										
B	NC	S3										
C	NC	S4										
D	NC	S5										
E	CAL	S6										
F	CAL	S7										
G	PC	S8										
H	S1	S9										

Legenda: BLK = Blank NC = Negative Control
CAL = Calibrator PC = Positive Control S = Sample

O. INTERNAL QUALITY CONTROL

A check is performed on the controls/calibrator any time the kit is used in order to verify whether the expected OD450nm/620-630nm or Co/S values have been matched in the analysis. Ensure that the following parameters are met:

Parameter	Requirements
Blank well	< 0.050 OD450nm value
Negative Control (NC)	> 1.000 OD450nm after blanking coefficient of variation < 20%
Calibrator (about 2 PEI U/ml)	Co/S > 1
Positive Control	< 0.200 OD450nm

If the results of the test match the requirements stated above, proceed to the next section.

If they do not, do not proceed any further and perform the following checks:

Important note:

The analysis must be done proceeding as the reading step described in the section M, point 12.

P. RESULTS

The results are calculated by means of a cut-off value determined with the following formula:

$$\text{Cut-Off} = (\text{NC} + \text{PC}) / 5$$

Important note: When the calculation of results is performed by the operating system of an ELISA automated work station, ensure that the proper formulation is used to calculate the cut-off value and generate the correct interpretation of results.

Q. INTERPRETATION OF RESULTS

Results are interpreted as ratio between the cut-off value and the sample OD450nm/620-630nm or Co/S.

Results are interpreted according to the following table:

Co/S	Interpretation
< 0.9	Negative
0.9 - 1.1	Equivocal
> 1.1	Positive

A negative result indicates that the patient has not been infected by HBV.

Any patient showing an equivocal result should be re-tested on a second sample taken 1-2 weeks after the initial sample.

The blood unit should not be transfused.

A positive result is indicative of HBV infection and therefore the patient should be treated accordingly or the blood unit should be discarded.

Important notes:

1. Interpretation of results should be done under the supervision of the laboratory supervisor to reduce the risk of judgement errors and misinterpretations.
2. When test results are transmitted from the laboratory to another facility, attention must be paid to avoid erroneous data transfer.
3. Diagnosis of viral hepatitis infection has to be taken by and released to the patient by a suitably qualified medical doctor.

An example of calculation is reported below (data obtained proceeding as the the reading step described in the section M, point 12):

The following data must not be used instead of real figures obtained by the user.

Negative Control: 2.000 – 2.200 – 2.000 OD450nm
 Mean Value: 2.100 OD450nm
 Higher than 1.000 – Accepted

Positive Control: 0.100 OD450nm
 Lower than 0.200 – Accepted

Cut-Off = (2.100 + 0.100) / 5 = 0.440

Calibrator: 0.400-0.360 OD450nm
 Mean value: 0.380 OD450nm
 Co/S > 1 – Accepted

Sample 1: 0.028 OD450nm
 Sample 2: 1.890 OD450nm
 Sample 1 Co/S > 1.1 positive
 Sample 2 Co/S < 0.9 negative

R. PERFORMANCES

Evaluation of Performances has been conducted in accordance to what reported in the Common Technical Specifications or CTS (art. 5, Chapter 3 of IVD Directive 98/79/EC).

1. LIMIT OF DETECTION:

The sensitivity of the assay has been calculated by means of the reference preparation for HBcAb supplied by Paul Erlich Institute (PEI HBc Reference Material 82). The assay shows a sensitivity of about 1.25 PEI U/ml. The table below reports the Co/S values shown by the PEI standard diluted as suggested by the manufacturer to prepare a limiting dilution curve in Fetal Calf Serum (FCS).

PEI U/ml	Lot 1001	Lot 0702	Lot 0702/2	Lot 1202
5	22.6	18.0	19.0	17.7
2.5	8.0	5.5	5.4	5.0
1.25	1.1	1.3	1.0	1.0
0.625	0.4	0.4	0.4	0.4

In addition Accurun 1 – series 3000 – supplied by Boston Biomedica Inc., USA, was tested to determine its Co/S value. Results are reported in the table below:

Accurun 1 – series 3000

Value	Lot 1001	Lot 0702	Lot 1202
Co/S	2.9	2.3	2.2

2. DIAGNOSTIC SPECIFICITY AND SENSITIVITY

The Performance Evaluation of the device was carried out in a trial conducted on more than total 6000 samples.

2.1 Diagnostic Specificity

It is defined as the probability of the assay of scoring negative in the absence of specific analyte. In addition to the first study, where a total of 5179 unselected donors, including 1st time donors, 206 samples from hospitalized patients and 164 potentially interfering specimen were examined, the diagnostic specificity was recently assessed by testing a total of 1498 negative samples on seven different lots. A value of specificity of 100% was observed. In addition to the above population, 189 potentially interfering samples (other liver diseases, pregnant women, hemolized, lipemic, RF positives) have been tested and found negative, confirming a 100% of specificity of the device. Finally, both human plasma, derived with different standard techniques of preparation (citrate, EDTA and heparin), and human sera have been used to determine the specificity. No false reactivity due to the method of specimen preparation has been observed.

2.2 Diagnostic Sensitivity

It defined as the probability of the assay of scoring positive in the presence of specific analyte. In addition to the first Performance Evaluation Study, in order to further evaluate the diagnostic sensitivity of the device, a total of 262 positive samples were recently evaluated. The respective results, collected from seven different lots of the device show a diagnostic sensitivity of 100%.

3. PRECISION

The mean values obtained from a study conducted on three lots and on two samples of different anti-HBcAg reactivity, examined in 16 replicates in three separate runs is reported below:

BCAB.CE: lot # 1202

Negative Control (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	1.943	1.939	1.924	1.935
Std.Deviation	0.081	0.078	0.103	0.087
CV %	4.2	4.0	5.3	4.5

Calibrator (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.143	0.147	0.148	0.146
Std.Deviation	0.014	0.017	0.018	0.016
CV %	9.8	11.4	12.1	11.1
Co/S	2.8	2.7	2.6	2.7

BCAB.CE: lot # 0702

Negative Control (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	2.163	2.110	2.106	2.126
Std.Deviation	0.105	0.088	0.139	0.111
CV %	4.9	4.2	6.6	5.2

Calibrator (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.182	0.193	0.195	0.190
Std.Deviation	0.018	0.023	0.019	0.020
CV %	10.0	12.0	9.9	10.6
Co/S	2.5	2.2	2.3	2.3

BCAB.CE: lot # 0702/2

Negative Control (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	2.278	2.098	2.130	2.169
Std.Deviation	0.135	0.126	0.159	0.140
CV %	5.9	6.0	7.5	6.5

Calibrator (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.193	0.190	0.199	0.134
Std.Deviation	0.023	0.023	0.027	0.025
CV %	12.1	12.3	13.5	12.6
Co/S	2.4	2.2	2.2	2.3

The variability shown in the tables did not result in sample misclassification.

Important note:

The performance data have been obtained proceeding as the reading step described in the section M, point 12.

S. LIMITATIONS OF THE PROCEDURE

Bacterial contamination or heat inactivation of the specimen may affect the absorbance values of the samples with consequent alteration of the level of the analyte. This test is suitable only for testing single samples and not pooled ones. Diagnosis of an infectious disease should not be established on the basis of a single test result. The patient's clinical history, symptomatology, as well as other diagnostic data should be considered.

REFERENCES

1. Aach R.D., Grisham J.W., Parker S.W.. Proc.Natl.Acad.Sci..USA, 68:1956, 1971.
2. Blumerg B.S., Suinick A.I., London W.T.. Hepatitis and leukemia: their relation to Australia antigen. Bull.N.Y.Acad.Med.. 44:1566, 1968.
3. Boniolo A., Dosis M., Matteja R.. J.Immunol.Meth.. 49:1, 1982.
4. Caldwell C.W., Barpet J.T.. Clin.Chim.Acta 81: 305, 1977
5. Fazekas S., De St.Groth, Scheidegger D.. J.Immunol.Meth.. 35: 1, 1980
6. Reesink H.W.. et al.. Vox.Sang.. 39:61, 1980
7. Rook G.A.W.. Lepr.Rev. 52: 281, 1981
8. Schroder J.. Med.Biol.. 58: 281, 1981
9. Almeida J.D. et al.. Lancet, ii : 1225, 1971
10. Hoofnagle J.H. et al.. Lancet, ii: 869, 1973
11. Hoofnagle J.H. et al.. N.E.J.Med., 290: 1336, 1974
12. Katchaki J.N. et al.. J.Clin.Path., 31: 837, 1978
13. Szmunness W. et al.. Am.J.Epidem., 104 : 256, 1976
14. Grebenchtchikov N. et al.. J.Immunol. Methods, 15(2) :219-231, 2002
15. Schrijver RS and Kramps JA, Rev.Sci.Tech. 17(2):550-561, 1998

All the IVD Products manufactured by the company are under the control of a certified Quality Management System approved by an EC Notified Body. Each lot is submitted to a quality control and released into the market only if conforming with the EC technical specifications and acceptance criteria.

Manufacturer:
Dia.Pro Diagnostic Bioprobes S.r.l.
Via G. Carducci n° 27 – Sesto San Giovanni (MI) – Italy

