CERTIFICATE OF ANALYSIS

Digitally signed by Bolea Petru
Date: 2026.04.03 14:16:23 EEST
Reason: MoldSign Signature
Location: Moldova
MOLDOVA EUROPEANĂ

Product : **CONFIRM' SALMONELLA KIT 50 tests**
Reference : BT01108 Expiry date : 2027.02.13
Batch N° : 0119437

Analysis	Specifications	Results
N° batch R1	117833	
N° batch R2	117834	
N° batch R3	117835	
Characters		
Appearance atex Salmonella (R1 reagent)	Milky suspension	Conform
Color latex Salmonella (R1 reagent)	White	Conform
Appearance positive control (R2 control)	Opalescente solution	Conform
Color positive control (R2 control)	White	Conform
Appearance negative control (R3 control)	Clear solution	Conform
Color négative control (R3 control)	Colorless	Conform
Microbiology		
Agglutination Salmonella Typhimurium ATCC 14028 (WDCM 00031)	Positive	Conform
Agglutination Salmonella Enteritidis CIP 82.97 (WDCM 00030)	Positive	Conform
Agglutination Escherichia coli ATCC 25922 (WDCM 00013)	Negative	Conform
Agglutination positive control (R2)	Positive	Conform
Agglutintion negative control (R3)	Negative	Conform

Beauvais August 18, 2025

CONFORM **Batch released by: Celine PILLITU (QC)**

This certificate is electronically generated and requires no signature
Biokar Diagnostics - Rue des Quarante Mines - ZAC de Ther - Allonne - BP10245 - F60002 BEAUVAIS CEDEX - FRANCE
Tel: +33(0)3 44 14 33 33 Fax: +33(0)3 44 14 33 34