



DECLARATION OF CONFORMITY

The undersigned ELMI SIA company declares, that the following instrument(s) conforms the bellow mentioned directives and standards:

Manufacturer:	ELMI SIA Krīvu street 11-136, Riga LV-1006, Latvia Phone: (+371) 67558 743 Fax: (+371) 67551 934
Equipment name:	S3, S-3M.A02, S-3M.A10, S-3M.A20, S-3L.A10, S-3L.A20, S-3.02M.A10, S-3.02M.A20, S-3.02L.A10, S-3.02L.A20, S-4
Equipment type:	Orbital Shaker
Directives:	EMC Directive 2004/108/EC Low Voltage Directive 2006/95/EC RoHS 2011/65/EC WEEE 2002/96/EC & 2012/19/EU In vitro diagnostic medical devices directive 98/79/EC

Applied standards:	<u>EN 61326-1: 2013</u> Electrical equipment for measurement, control and laboratory use EMC requirements. General requirements. <u>EN 61010-1: 2010</u> Safety requirements for electrical equipment for measurement, control and laboratory use. General requirements. <u>EN 61010-2-051: 2015</u> Particular requirements for laboratory equipment for mixing and stirring.
--------------------	--

Vitalijs Mironovs
Technical Director

Maksims Sviridovs
Manufacturing Director



30.04.2019
Date

30.04.2019
Date