

## SPECIFICATIE TEHNICA

**PRODUS:**

**Ser negativ pentru reactiile serologice-RSAR Roz Bengal/SAT/RFC, aplicate in bruceleza**

**Descriere:**

Ser bovin liber de anticorpi anti-*B.abortus* si anti- *Y.enterocolitica*.

Ambalare, etichetare, marcare conforme cu legislatia UE in vigoare la data incheierii contractului de achizitie publica.

Produsul biologic este insojit obligatoriu de certificat de calitate in valabilitate al producatorului ISO 9001/2008 pentru activitatea de fabricare kituri si reagenti de diagnostic si buletin de analiza.

Valabilitate mini de 1 an din momentul achizitionarii produsului biologic.

Ambalaj: flacon 1 ml;

Producator: IDEXX MONTPELLIER SAS



**CERTIFICATE OF ANALYSIS**  
 Certificado de Análisis / Certificat de contrôle / Chargen-Prüfprotokoll  
Pourquier\* Brucellosis Negative Control [Brucellosis Negative Control]

Product code :	P00131	Expiration date :	08-SEP-2019
Batch :	144		
Tests per kit :	1	Manufacturing date :	SEP-2014
Insert version :	0		

Reconstitute\* with 1ml distilled water

\* After reconstitution, the control serum must be stored at a temperature  $\leq -16^{\circ}\text{C}$  for up to 1 year


He can be frozen and unfrozen up to three times


Control Test	Spécifications	Results
EAT :	Negative - No agglutination	Pass
SAW :	Negative( < 30 UI/ml)	Pass
FC :	Negative( < 20 ICFTU/ml)	Pass
ELISA :	Negative	Pass

This product was performance tested and has met all quality control specifications required for release.

Title :

Signature :

Sélectionner  
 P/O 

Sélectionner  


Manufacturer: IDEXX Montpellier SAS - 326, rue de la Galéra - Parc Euromédecine - 34097 Montpellier Cedex 5 - France  
 Tél. : 33(0) 4 99 23 24 25 - Fax : 33(0) 4 67 04 20 25 - Info-montpellier@idexx.com

EN 05/01/17 vB - Date 05/08/2012

Document printed on : 30/10/2014