

# eSens *Mycoplasma hominis* QL PCR kit

**REF ES3004A**

## Instructions for Use

### 1 INTENDED USE

**eSens *Mycoplasma hominis* QL PCR kit** is an *in vitro* nucleic acid amplification test for qualitative detection of *Mycoplasma hominis* DNA in the clinical materials (urogenital swabs, urine, and prostate gland secretion) using real-time hybridization-fluorescence detection of amplified products.

NOTE: The results of PCR analysis are taken into account in complex diagnostics of disease.

### 2 PRINCIPLE OF PCR DETECTION

*Mycoplasma hominis* DNA detection by the polymerase chain reaction (PCR) is based on the amplification of the pathogen genome specific region using specific *Mycoplasma hominis* primers. In the real-time PCR, the amplified product is detected with the use of fluorescent dyes. These dyes are linked to oligonucleotide probes which bind specifically to the amplified product during thermocycling. The real-time monitoring of fluorescence intensities during the real-time PCR allows the detection of accumulating product without re-opening the reaction tubes after the PCR run.

**eSens *Mycoplasma hominis* QL PCR kit** is a qualitative test that contains the Internal Control (Internal Control-FL (IC)). It must be used in the extraction procedure in order to control the extraction process of each individual sample and to identify possible reaction inhibition.

**eSens *Mycoplasma hominis* QL PCR kit** uses “hot-start”, which greatly reduces the frequency of nonspecifically primed reactions. “Hot-start” is guaranteed by using chemically modified polymerase (TaqF). The chemically modified polymerase (TaqF) is activated by heating at 95 °C for 15 min.

The PCR kit contains the system for prevention of contamination by amplicons using the enzyme uracil-DNA-glycosylase (UDG) and dUTP. The enzyme UDG recognizes and catalyzes the destruction of the DNA containing deoxyuridine, but has no effect on DNA containing deoxythymidine. Deoxyuridine is absent in the authentic DNA, but is always present in amplicons, because dUTP is a part of dNTP mixture in the reagents for the amplification. Due to the deoxyuridine containing contaminating amplicons are sensitive to the destruction by UDG before the DNA-target amplification. So the amplicons cannot be amplified.

The enzyme UDG is thermolabile. It is inactivated by heating at temperature above 50 °C. Therefore, UDG does not destroy the target amplicons which are accumulated during PCR.

The results of amplification are registered in the following fluorescence channels:

**Table 1**

Channel for fluorophore	FAM	JOE
DNA-target	<i>Mycoplasma hominis</i> DNA	Internal Control-FL (IC) DNA
Target gene	gene 16S rRNA	Artificially synthesized sequence

### 3 CONTENT

**eSens Mycoplasma hominis QL PCR kit** (ES3004A) contains

Reagent	Description	Volume, ml	Quantity
<b>PCR-mix-1-FL Mycoplasma hominis</b>	clear liquid from colorless to light lilac colour	1.2	1 tube
<b>PCR-mix-2-FRT</b>	colorless clear liquid	0.3	2 tubes
<b>Polymerase (TaqF)</b>	colorless clear liquid	0.03	2 tubes
<b>Positive Control complex (C+)</b>	colorless clear liquid	0.2	1 tube
<b>DNA-buffer</b>	colorless clear liquid	0.5	1 tube
<b>Negative Control (C-)*</b>	colorless clear liquid	1.2	1 tube
<b>Internal Control-FL (IC)**</b>	colorless clear liquid	1.0	1 tube

\* must be used in the extraction procedure as Negative Control of Extraction.

\*\* add **10 µl** of **Internal Control-FL (IC)** during the DNA extraction directly to the sample/lysis mixture.

**eSens Mycoplasma hominis QL PCR kit** is intended for 110 reactions (including controls).

### 4 ADDITIONAL REQUIREMENTS

- Transport medium.
- DNA extraction kit.
- Disposable powder-free gloves and laboratory coat.
- Pipettes (adjustable).
- Sterile pipette tips with aerosol filters (up to 100 µl).
- Tube racks.
- Vortex mixer.
- Desktop centrifuge with a rotor for 2-ml reaction tubes.
- PCR box.
- Real-time instruments (for example, Rotor-Gene Q (QIAGEN, Germany), CFX 96 Touch, CFX 96 Opus (Bio-Rad, USA), QuantStudio 5 (Thermo Fisher Scientific), or equivalent).
- Disposable polypropylene tubes:
  - thin-walled 0.2-ml PCR tubes with domed caps if a plate-type instrument is used;

- thin-walled 0.2-ml PCR tubes with flat caps or strips of four 0.1-ml Rotor-Gene PCR tubes if a rotor-type instrument is used.
- Refrigerator for 2–8 °C.
- Deep-freezer at the temperature from minus 24 to minus 16 °C.
- Reservoir for used tips.

## 5 GENERAL PRECAUTIONS

The user should always pay attention to the following:

- Use sterile pipette tips with aerosol filters and use a new tip for every procedure.
- Store all extracted positive material (specimens, controls and amplicons) away from all other reagents and add it to the reaction mix in a distantly separated facility.
- Thaw all components thoroughly at room temperature before starting an assay.
- When thawed, mix the components and centrifuge briefly.
- Use disposable protective gloves and laboratory cloths, and protect eyes while samples and reagents handling. Thoroughly wash hands afterwards.
- Do not eat, drink, smoke, apply cosmetics, or handle contact lenses in laboratory work areas.
- Do not use a kit after its expiration date.
- Dispose of all specimens and unused reagents in accordance with local regulations.
- Samples should be considered potentially infectious and handled in biological cabinet in compliance with appropriate biosafety practices.
- Clean and disinfect all samples or reagents spills using a disinfectant, such as 0.5 % sodium hypochlorite or another suitable disinfectant.
- Avoid inhalation of vapors, samples and reagents contact with the skin, eyes, and mucous membranes. Harmful if swallowed. If these solutions come into contact, rinse the injured area immediately with water and seek medical advice if necessary.
- Safety Data Sheets (SDS) are available on request.
- Use of this product should be limited to personnel trained in DNA amplification techniques.
- Workflow in the laboratory must be one-directional, beginning in the Extraction Area and moving to the Amplification and Detection Area. Do not return samples, equipment and reagents in the area where the previous step was performed.



Some components of this kit contain sodium azide as a preservative. Do not use metal tubing for reagent transfer.

## 6 SAMPLING AND HANDLING

**eSens Mycoplasma hominis QL PCR kit** is intended for analysis of the DNA extracted with DNA extraction kits from the clinical material (urogenital swabs, urine samples (sediment of the first portion of the morning specimen), prostate gland secretion).

## 7 WORKING CONDITIONS

**eSens Mycoplasma hominis QL PCR kit** should be used at 18–25 °C.

## 8 PROTOCOL

### 8.1 DNA extraction

Any commercial nucleic acid extraction kit, if IVD-CE validated for the indicated specimen types, could be used.

**Ecoli Dx, s.r.o. recommends:**

- For the manual extraction

- **DNA-sorb-AM** (K1-12-100-CE)

- For the automatic extraction

- **ePure STD DNA Extraction Kit (E2007)**

The DNA extraction for each sample is carried out in the presence of **Internal Control-FL (IC)**.

Please carry out nucleic acid extraction according to the manufacture's instruction.

### 8.2 Preparing PCR

#### 8.2.1 Preparing tubes for PCR

The type of tubes depends on the PCR instrument used for analysis. Use disposable filter tips for adding reagents, DNA and control samples into tubes.

The total reaction volume is **25 µl**, the volume of DNA sample is **10 µl**.

1. Thaw the tube with PCR-mix-2-FRT. Vortex the tubes with PCR-mix-1-FL *Mycoplasma hominis*, PCR-mix-2-FRT, and polymerase (TaqF) then centrifuge briefly.  
Take the required number of the tubes/strips for amplification of the DNA obtained from clinical and control samples.
2. For N reactions (including 2 controls of amplification), add to a new tube:  
**10\*(N+1) µl of PCR-mix-1-FL *Mycoplasma hominis*;**  
**5.0\*(N+1) µl of PCR-mix-2-FRT;**  
**0.5\*(N+1) µl of polymerase (TaqF).**  
Vortex the tube, then centrifuge briefly. Transfer **15 µl** of the prepared mixture to each tube.
3. Add **10 µl** of **DNA** obtained at the DNA extraction stage.
4. Carry out the control reactions:

<b>NCA</b>	– Add <b>10 µl</b> of <b>DNA-buffer</b> to the tube labeled <b>NCA</b> (Negative Control of Amplification).
<b>C+</b>	– Add <b>10 µl</b> of <b>Positive Control complex (C+)</b> to the tube labeled <b>C+</b> (Positive Control of Amplification).
<b>C–</b>	– Add <b>10 µl</b> of <b>the sample extracted from the Negative Control (C–) reagent</b> to the tube labeled <b>C–</b> (Negative control of Extraction).

#### 8.2.2 Amplification

1. Create a temperature profile on your instrument as follows:

Table 2

## eSens-1 amplification program

	Rotor-type Instruments (e.g Rotor-Gene Q or equivalent)			Plate-type Instruments (e.g CFX 96 Touch, CFX 96 Opus, QuantStudio 5 or equivalent.)		
Step	Temperature, °C	Time	Cycles	Temperature, °C	Time	Cycles
1	95	15 min	1	95	15 min	1
2	95	5 s	5	95	5 s	5
	60	20 s		60	20 s	
	72	15 s		72	15 s	
3	95	5 s	40	95	5 s	40
	60	20 s Fluorescence acquiring		60	30 s Fluorescence acquiring	
	72	15 s		72	15 s	

Fluorescent signal is detected in the channels for the FAM and JOE fluorophores. Other channels are enabled if several tests are simultaneously carried out in a single run.

1. Adjust the fluorescence channel sensitivity.
2. Insert tubes into the reaction module of the device.
3. Run the amplification program with fluorescence detection.
4. Analyze results after the amplification program is completed.

## 8.3 Instrument Settings

## Test settings for rotor-type instruments

Channel	Calibrate/Gain Optimisation	Threshold	Dynamic tube	Slope Correct	More Settings/ Outlier Removal
FAM/Green	from 5 FI to 10 FI	0.1	On	Ott	0%
JOE/Yellow	from 4 FI to 8 FI	0.1	On	Ott	5%

## Test settings for plate-type instruments

Set the heating/cooling **Ramp Rate 2,5 °C/s**.

Channel	Threshold
FAM JOE/HEX	For each channel in <b>Log Scale</b> set the threshold line at the level of 10-20 % of maximum fluorescence obtained for the Positive Control of Amplification (C+) in the last amplification cycle

## 9 DATA ANALYSIS

Analysis of results is performed by the software of the real-time PCR instrument used by measuring fluorescence signal accumulation in two channels:

- The signal of the *Mycoplasma hominis* DNA amplification product is detected in the channel for the FAM fluorophore;
- The signal of the Internal Control amplification product is detected in the channel for the JOE fluorophore.

Results are interpreted by the crossing (or not-crossing) the fluorescence curve with the threshold line set at the specific level that corresponds to the presence (or absence) of a Ct value of the DNA sample in the corresponding column of the result grid.

Principle of interpretation is the following:

- *Mycoplasma hominis* DNA is **detected** in a sample if the Ct value is determined in the result grid in the channel for the FAM fluorophore. Moreover, the fluorescence curve of the sample should cross the threshold line in the area of typical exponential growth of fluorescence.
- *Mycoplasma hominis* DNA is **not detected** in a sample if the Ct value is not determined (absent) in the result grid in the channel for the FAM fluorophore (the fluorescence curve does not cross the threshold line), whereas the Ct value in the channel for JOE fluorophore is less than the specified boundary Ct value.
- The result is **invalid** if the Ct value is not determined (absent) in the channel for FAM fluorophore, whereas the Ct value in the channel for JOE fluorophore is not determined (absent) or greater than the specified boundary Ct value. In such cases, the PCR analysis should be repeated for such samples.

**The result of the analysis is considered reliable only if the results obtained for the Positive and Negative Controls of amplification as well as for the Negative Control of extraction are correct (see Table 3 and 4).**

**Table 3**

**Results for controls**

Control	Stage for control	Ct value in the channel for fluorophore	
		FAM	JOE
<b>C-</b>	DNA extraction	Absent	<boundary value
<b>NCA</b>	PCR	Absent	Absent
<b>C+</b>	PCR	<boundary value	<boundary value

**Table 4**

**Boundary Ct values**

Sample	Rotor-type instrument		Plate-type instrument	
	Channel for fluorophore			
	FAM	JOE	FAM	JOE
C+	33	30	36	33
C-	Cl is absent	30	Clis absent	33
NGA	Cl is absent		Cl is absent	
Test samples	-	30	-	33

## 10 TROUBLESHOOTING

The results of analysis are not taken into account in the following cases:

1. If the *Ct* value determined for the Positive Control of Amplification (C+) in the channel for the FAM fluorophore is greater than the boundary *Ct* value or absent, the amplification should be repeated for all samples in which *Mycoplasma hominis* DNA was not detected.
2. If the *Ct* value is determined for the Negative Control of Amplification (NCA) and/or Negative Control of Extraction (C-) in the channel for the FAM fluorophore, the PCR analysis (beginning with the DNA extraction stage) should be repeated for all samples in which *Mycoplasma hominis* DNA was detected.

## 11 TRANSPORTATION

**eSens Mycoplasma hominis QL PCR kit** should be transported at 2–8 °C for no longer than 5 days.

## 12 STABILITY AND STORAGE

All components of the **eSens Mycoplasma hominis QL PCR kit** are to be stored at 2–8 °C when not in use (except for polymerase (TaqF) and PCR-mix-2-FRT).

All components of the **eSens Mycoplasma hominis QL PCR kit** are stable until the expiry date stated on the label. PCR kit can be stored without unpacking at 2 to 8 °C for 3 months from the date of manufacture before opening. Once opened, PCR kit should be unpacked in accordance with the storage temperatures for each component. The shelf life of reagents before and after the first use is the same, unless otherwise stated.

NOTE: Polymerase (TaqF) and PCR-mix-2-FRT are to be stored at the temperature from minus 24 to minus 16 °C.

PCR-mix-1-FL *Mycoplasma hominis* is to be kept away from light.

## 13 SPECIFICATIONS

### 13.1 Sensitivity

The analytical sensitivity of **eSens Mycoplasma hominis QL PCR kit** is specified in the table below.

Clinical material	Transport medium	Nucleic acid extraction kit	Sensitivity, GE/ml*
Urogenital swabs	Transport Medium for Swabs or Transport Medium with Mucolytic Agent	DNA-sorb-AM ePure STD DNA Extraction kit	1x10 <sup>3</sup>
Urine (pretreatment is required)	–	DNA-sorb-AM ePure STD DNA Extraction kit	2x10 <sup>3</sup>

\*The quantity of genome equivalents of microorganism per 1 ml of the sample from transport medium.

## 13.2 Specificity

The analytical specificity of **eSens Mycoplasma hominis QL PCR kit** is ensured by selection of specific primers and probes as well as by selection of stringent reaction conditions. The primers and probes were checked for possible homologies to all sequences deposited in gene banks by sequence comparison analysis.














Nonspecific reactions were absent while testing human DNA samples and DNA panel of the following microorganisms: *Ureaplasma urealyticum* and *U. parvum*; *Mycoplasma genitalium*; *Gardnerella vaginalis*; *Lactobacillus* spp.; *Escherichia coli*; *Staphylococcus aureus*; *Streptococcus pyogenes*; *Streptococcus agalactiae*; *Candida albicans*; *Neisseria flava*, *N. subflava*, *N. sicca*, *N. mucosa*, and *N. gonorrhoeae*; *Chlamydia trachomatis*; *Trichomonas vaginalis*; *Treponema pallidum*; *Toxoplasma gondii*; HSV types 1 and 2; CMV; and HPV.

The clinical specificity of **eSens Mycoplasma hominis QL PCR kit** was confirmed in laboratory clinical trials.

## 14 QUALITY CONTROL

The production process, including batch release, is carried out in accordance with an established quality management system certified according to ISO 13485.

## 15 KEY TO SYMBOLS USED

	Catalogue number		Caution
	Batch code		Contains sufficient for <n> tests
	<i>In vitro</i> diagnostic medical device		Use-by Date
	Version		Consult instructions for use
	Temperature limit		Keep away from sunlight
	Manufacturer	NCA	Negative control of amplification
	Date of manufacture	C-	Negative control of extraction
	Authorized representative in the European Community	C+	Positive control of amplification
		IC	Internal control

### List of Changes Made in the Instruction Manual

VER	Location of changes	Essence of changes
01_04/2022		





Ecoli Dx, s.r.o. , Purkyňova 74/2

110 00 Praha 1, Česká republika

Tel: +420 325 209 912

Mobil: +420 739 802 523