

Product fact sheet

Blood Bank mode

Version 1.0



Manufacturer information

Sysmex Corporation www.sysmex-europe.com

Summary

Blood component safety is key in transfusion medicine and blood product manufacturing. Therefore, high standards during the whole process from donor qualification to quality monitoring of the products are required. Especially residual cell enumeration is important for process quality control. Until now, haematology analysers could not reach the lower limit of quantification required for this.

The Blood Bank mode¹ is an exclusive analysis mode of XN-Series analysers to analyse blood products such as red blood cell concentrates and platelet concentrates for blood transfusion. The advanced technology used in this mode enables counting of RBC and PLT in dense blood packs as well as the determination of reliable residual WBC and RBC counts.

The Blood Bank mode is an optional application and software licence registration is required for its activation on applicable XN-Series configurations.

¹ Note Blood Bank mode is not CE-IVD marked and thereby it is not to be used for diagnostic purposes.

Productivity

- Decreased workload by using one single analyser for multiple parameters
- Fully automated and standardised measurement
 - Automated background check analysis to avoid carryover
 - Specific profiles tailored for different blood products
- Sampler analysis capability ensures high throughput for blood banks/transfusion laboratories
- Easy mode switching function between Whole Blood² and Blood Bank mode

Ordering information

Item code	Item	Qty
ZE004088	Blood Bank Mode Licence	1

Available configurations for Blood Bank mode

Blood Bank mode is applicable to XN-1000 and XN-2000 configurations. The basic prerequisites of the XN analysis module are described in the table below. Please note that

- in case of an XN-2000, Blood Bank mode licence activation on each analyser is mandatory.
- SA-10 sampler unit is the only supported sampler for an XN-1000 configuration.



Analysis module	Default applications	Required applications
XN-10	added value XN-CBC Added value XN-DIFF	added value RET PLT-F
XN-20	added value XN-CBC XN-DIFF RET widerd value WPC	ndded yalus PLT-F

Profile overview

Profile	Description
RBC Pack	Enumerating RBC and measuring HGB and HCT in red blood cell packs
RBC Pack + Res. Cell ³	Enumerating RBC and measuring HGB and HCT, additionally counting residual WBC in red blood cell packs
PLT Pack	Enumerating PLT in platelet packs
PLT Pack + Res. Cell ³	Enumerating PLT, additionally counting residual WBC and RBC in platelet packs

³ 'Res. Cell' indicates 'residual cells' in blood packs.

² This is a dedicated mode for human whole blood analysis.

Product specifications

1 Todact opcomoducito		
Feature	Description	
Analysis mode	Blood Bank mode	
Sampling mode	Sampler analysisManual analysis	
Analysis time		
RBC Pack RBC Pack + Res. Cell PLT Pack PLT Pack + Res. Cell	 45 sec (< 1 min) 109 sec (< 2 min) 58 sec (≈ 1 min) 186 sec (≈ 3 min) 	
Technologies	 Fluorescence flow cytometry Hydrodynamically focused DC detection method Cyanide-free SLS haemoglobin method 	
Analysis parameters		

RBC Pack & RBC Pack + Res. Cell	 RBC – Red blood cell count HGB – Haemoglobin concentration HCT – Haematocrit WBC⁴ – White blood cell count
PLT Pack & PLT Pack + Res. Cell	 PLT – Platelet count WBC⁴ – White blood cell count

FLI Fack + Nes. Cell	VIDC - Write blood cell coult
Research parameters	
RBC Pack & RBC Pack + Res. Cell	 MCV – Mean corpuscular volume MCH – Mean corpuscular haemoglobin MCHC – Mean corpuscular haemoglobin concentration RDW-SD – Red blood cell distribution width (standard deviation) RDW-CV – Red blood cell distribution width (coefficient variation)
PLT Pack & PLT Pack + Res. Cell	 IPF% - Immature platelet fraction ratio IPF# - Immature platelet fraction count PDW - Platelet distribution width MPV - Mean platelet volume RBC-O⁴ - Red blood cell count from the RET channel RBC-I⁵ - Red blood cell count from the RBC/PLT channel HGB - Haemoglobin concentration

Sampler unit capacity • 10 racks (100 samples)

External output functionalities ⁶	Output to the printer (GP/LP)Output in CSV format
Quality control and calibrator materials	
XN Check Level 1 XN Check Level 2 XN Check Level 3	QC material for monitoring the Whole Blood mode in three concentration levels
XN Cal	Calibrator material
XN Cal PF	Calibrator material for PLT-F channel
⁴ Analysed when residual cell of	count analysis is performed

Analysed when residual cell count analysis is performed.

Aspiration and required sample volume

Profile	Analysis type	Tube type	Aspiration volume	Required sample volume
RBC Pack & RBC Pack + Res. Cell ⁷	Sampler analysis	Closed tube		1 mL
		Closed tube	150 µL	1 mL
	Manual	Open tube		400 µL
	analysis	Micro collection tube		260 μL
PLT Pack & PLT Pack + Res. Cell ⁷	Sampler analysis	Closed tube		1 mL
	Manual analysis	Closed tube	205 μL	1 mL
		Open tube		400 µL
		Micro collection tube		300 µL

⁷ 'Res. Cell' indicates 'residual cells' in blood packs.

 $^{^{\}rm 5}\,{\rm Reporting}$ this parameter depends on one Service Setting.

⁶ SUIT XN and *Extended* IPU integration are under development.

Performance specifications

Feature	Description		
Throughput			
RBC Pack RBC Pack + Res. Cell PLT Pack PLT Pack + Res. Cell	79 sample33 sample62 sample19 sample	es/hour es/hour	
Analysis range			
WBC	• 0.0000 –	0.3500 x10³/µ	L
RBC ⁸	• 0.00 – 8.6	60 x106/µL	
HGB	 0.0 – 26.0 0.0 – 16.1 		
НСТ	• 0.0 – 75.0)%	
PLT ⁹	 0 – 5,000 	x10³/μL	
Limit of Blank (LoB) / Limit of Detection (LoD) / Limit of Quantitation (LoQ)	LoB	LoD	LoQ
WBC (x10³/µL)	0.0000	0.0020	0.0020
RBC (x10 ⁶ /µL)	0.00	0.01	0.01
HGB (g/dL)	0.0	0.1	0.1
HCT (%)	0.0	0.1	0.1
PLT (x10³/µL)	0	1	1
Precision (repeatability)			
WBC	x10 ³ /µL) • ≤ 20.0% (x10 ³ /µL) • ≤ 10.0% (x10 ³ /µL)	0.0020 - 0.00 0.0051 - 0.01 0.0151 - 0.03 0.0301 x10 ³ /	50
RBC ⁸	• ≤ 1.5% (≥	4.00 x10 ⁶ /μL)
HGB	• ≤ 1.0%		
НСТ	• ≤ 1.5%		
PLT ⁹	•	0 – 99 x10³/μl 100 x10³/μL)	•

Precision (intermediate precision / reproducibility)	1057/ 041 01 1 1 25 1 1 1 1
WBC	 ≤ 35% (XN Check BF Level 1) ≤ 25% (XN Check BF Level 2)
RBC ⁸	 ≤ 5% (XN Check Level 1) ≤ 5% (XN Check Level 2) ≤ 5% (XN Check Level 3)
HGB	 ≤ 4% (XN Check Level 1) ≤ 3% (XN Check Level 2) ≤ 3% (XN Check Level 3)
НСТ	 ≤ 10% (XN Check Level 1) ≤ 10% (XN Check Level 2) ≤ 10% (XN Check Level 3)
PLT ⁹	 ≤ 45% (XN Check Level 1) ≤ 30% (XN Check Level 2) ≤ 15% (XN Check Level 3)
Accuracy	
WBC	• Correlation coefficient $r \ge 0.95$ Slope $0.9 - 1.1$
RBC ⁸	• within ± 2% or ± 0.03 x106/µL
HGB	• within ± 2% or ± 0.2 g/dL
НСТ	• within ± 3% or ± 1.0% HCT
PLT ⁹	• within \pm 5% or \pm 10 x10 ³ / μ L
Linearity	
WBC	• $R^2 \ge 0.95 (0 - 0.350 \times 10^3/\mu L)$
RBC ⁸	• within $\pm 2\%$ or $\pm 0.03 \times 10^6/\mu$ L (0 $- 8.00 \times 10^6/\mu$ L)
	• within ± 4% or ± 0.06 x10°/µL ¹⁰ (8.01 – 8.60 x10°/µL)
HGB	 within ± 2% or ± 0.2 g/dL (0 – 25.0 g/dL) (0 – 15.5 mmoL/L) within ± 5% or ± 0.5 g/dL¹⁰ (25.1 – 26.0 g/dL) (15.6 – 16.1 mmoL/L)
HCT	 within ± 3% or ± 1.0% HCT (0 – 75%)
PLT ⁹	• within \pm 5% or \pm 10 x10 ³ /µL (0 - 1,000 x10 ³ /µL) • within \pm 6% ¹⁰ (1,001 - 5,000 x10 ³ /µL)

Carryover	
WBC	• ≤ 0.01%
RBC ⁸	 ≤ 1.0%
HGB	• ≤ 1.0%
HCT	 ≤ 1.0%
PLT ⁹	 ≤ 1.0%

⁸ RBC count from the RBC/PLT channel.

For detailed information please always refer to 'Blood Bank mode Instructions for Use' document.

For detailed information about XN-10 and XN-20 analysis module please refer to the respective Product Fact Sheets.

⁹ PLT count from the PLT-F channel.

¹⁰ Verification using stabilized substances