



# Liquichek™ Specialty Immunoassay Control Levels LTA, 1, 2 and 3

REF

359 Level LTA 6 x 5 mL  
364 Level 1 6 x 5 mL  
365 Level 2 6 x 5 mL  
366 Level 3 6 x 5 mL  
359X MiniPak 4 x 5 mL



IVD



EXP 2024-09-30

LOT

64950

Level LTA 64951L  
Level 1 64951  
Level 2 64952  
Level 3 64953


<https://www.myinserts.com/64950>

Data reviziei 2023-10-27 → Indică informații revizuite

## METODĂ (1)

	Unități	Nivel LTA - 64951L		
		Medie	Interval	
<b>25-HIDROXI VITAMINA D (TOTAL)</b>				
Abbott Alinity i Analyzer	ng/mL	7.85	5.51	– 10.2
Abbott ARCHITECT iSystems (3L52-25/35)	ng/mL		<8.00	– 10.6
Abbott ARCHITECT iSystems (3L52-27/37)	ng/mL	<13.0		
Abbott ARCHITECT iSystems (5P02-25/35/30)	ng/mL	7.95	5.28	– 10.6
Beckman Coulter Access / 2 / 2i	ng/mL	10.9	<7.00	– 15.2
Beckman Coulter UniCel Dxl	ng/mL	9.48	<7.00	– 13.8
bioMérieux VIDAS 25-OH Vitamin D Total	ng/mL	17.4	11.9	– 23.0
Bio-Rad BioPlex 2200 (2)	ng/mL	13.0	7.80	– 18.2
Calbiotech 25-OH Vitamin D ELISA (2)	ng/mL	11.5	7.70	– 15.4
DiaSorin LIAISON 25-OH Vitamin D TOTAL (2)	ng/mL	13.8	9.65	– 17.9
DiaSorin LIAISON XL 25-OH Vitamin D TOTAL (2)	ng/mL	14.1	9.90	– 18.4
DiaSorin LIAISON XS 25-OH Vitamin D TOTAL (2)	ng/mL	12.0	8.39	– 15.6
DIAsource ImmunoAssays 25OH-Vit. D3 RIA-CT (2)	ng/mL	■		
DIAZYME (EZ Vitamin D) (2)	ng/mL	18.2	13.7	– 22.8
IDS-iSYS 25 VitDS (REF IS-2500N) (2)	ng/mL	22.4	14.3	– 30.5
Ortho Vitros MicroWell Series (tVitD) (Cat. 684 2894) (2)	ng/mL	14.6	<8.00	– 28.7
Ortho Vitros MicroWell Series (tVitD) (Cat. 684 4056) (2)	ng/mL	14.6	<12.8	– 26.8
Roche cobas e 801 (Vitamin D III, REF 09038086) (2)	ng/mL	<6.00		
Roche cobas e801 (Vitamin D II, REF 07028148190)	ng/mL	18.4	12.6	– 24.1
Roche Elecsys/E170/Cobas e Systems (Roche Vit. D Total, REF 05894913)	ng/mL	§		
Roche Elecsys/E170/cobas e Systems (Vitamin D II, REF. 07464215160/REF. 07464215190)	ng/mL	13.3	9.04	– 17.6
Siemens ADVIA Centaur Systems (VitD)	ng/mL	28.8	20.0	– 37.6
<b>Siemens Atellica CI 1900 Analyzer (VitD) (IM) →</b>	<b>ng/mL</b>	<b>21.4</b>	<b>14.7</b>	<b>– 28.1</b>
Siemens Atellica Solution (VitD) (IM)	ng/mL	21.4	14.7	– 28.1
Siemens Dimension EXL (LOCI)	ng/mL	<5.00		
SNIBE Maglumi Series (2)	ng/mL	▲		
Tosoh ST AIA-PACK 25-OH Vitamin D (2)	ng/mL	22.7	15.9	– 29.5
<b>ANTI-TIROGLOBULINĂ (ANTI-TG)</b>				
Abbott Alinity i Analyzer	IU/mL	5.80	3.06	– 8.53
Abbott ARCHITECT iSystems	IU/mL	6.96	3.35	– 10.6
Beckman Coulter Access / 2 / 2i	IU/mL	4.32	3.19	– 5.44
Beckman Coulter Anti-hTg IRMA (2)	IU/mL	<32.0		
Beckman Coulter UniCel Dxl	IU/mL	§		

**METODĂ (1)**

	Unități	Nivel LTA - 64951L			
		Medie		Interval	
<b>ANTI-TIROGLOBULINĂ (ANTI-TG) (continuare)</b>					
bioMérieux VIDAS	IU/mL	10.8	7.32	–	14.3
DiaSorin LIAISON (2)	IU/mL	20.1	13.4	–	26.8
DiaSorin LIAISON XL (2)	IU/mL	14.8	<5.00	–	25.1
DIAsource ImmunoAssays TG Ab RIA (2)	IU/mL	10.5	7.38	–	13.7
Mindray CL-Series/Wiener lab. CLIA Series (2)	IU/mL	§			
Monobind AccuBind ELISA (2)	IU/mL	29.7	19.9	–	39.5
Monobind AccuLite CLIA (2)	IU/mL	51.7	34.6	–	68.7
Roche Elecsys/E170/Cobas e Systems	IU/mL	35.9	23.3	–	48.5
Shanghai Kehua Bio-Engineering (KHB) Unico 1800/1820 (2)	IU/mL	§			
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 348) (2)	U/mL	26.8	<25.0	–	41.4
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 350) (2)	U/mL		<25.0	–	35.0
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 352) (2)	U/mL	35.4	<25.0	–	51.3
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 353) (2)	U/mL		<25.0	–	35.4
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 355) (2)	U/mL	27.9	<25.0	–	42.7
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 357) (2)	U/mL		<25.0	–	37.1
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 359) (2)	U/mL	26.5	<25.0	–	41.1
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 361) (2)	U/mL	30.0	<25.0	–	45.1
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 363) (2)	U/mL		<25.0	–	38.5
Siemens ADVIA Centaur CP (aTgII)	IU/mL	5.32	1.45	–	9.19
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 348) (2)	U/mL	41.4	24.5	–	58.3
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 350) (2)	U/mL	45.3	27.8	–	62.8
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 352) (2)	U/mL	39.6	23.0	–	56.2
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 353) (2)	U/mL	41.2	24.4	–	58.0
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 355) (2)	U/mL	52.2	33.4	–	71.0
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 357) (2)	U/mL	51.1	32.5	–	69.7
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 359) (2)	U/mL	44.0	26.7	–	61.3
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 361) (2)	U/mL	42.0	25.0	–	59.0
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 363) (2)	U/mL	26.7	<15.0	–	41.3
Siemens ADVIA Centaur Systems (aTgII)	IU/mL	4.77	2.97	–	6.57
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 351)	U/mL	36.2	23.8	–	48.7
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 354)	U/mL	33.9	22.1	–	45.7
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 356)	U/mL	35.9	23.6	–	48.3
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 358)	U/mL	37.5	24.7	–	50.3
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 360)	U/mL	28.6	18.3	–	38.9
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 362)	U/mL	28.8	<15.0	–	43.7
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 364) (2)	U/mL	29.8	<15.0	–	44.8
Siemens Atellica Solution (aTgII)	IU/mL	5.60	3.58	–	7.63
Siemens IMMULITE / IMMULITE 1000	IU/mL	▲			
Siemens IMMULITE 2000 / 2000 XPI	IU/mL	<20.0			
SNIBE Maglumi Series (2)	IU/mL	▲			
Tosoh AIA-PACK TgAb (2)	IU/mL	<12.2			
<b>ANTI-TIROPEROXIDAZĂ (ANTI-TPO)</b>					
Abbott Alinity i Analyzer	IU/mL	8.04	5.99	–	10.1
Abbott ARCHITECT iSystems	IU/mL	9.35	6.97	–	11.7
Beckman Coulter Access / 2 / 2i	IU/mL	5.68	4.17	–	7.19
Beckman Coulter Anti-TPO RIA (2)	IU/mL	12.6	7.53	–	17.6
Beckman Coulter UniCel Dxl	IU/mL	5.60	4.43	–	6.77
bioMérieux VIDAS	IU/mL	6.10	4.47	–	7.73

**METODĂ (1)**

	Unități	Nivel LTA - 64951L		
		Medie	Interval	
<b>ANTI-TIROPEROXIDAZĂ (ANTI-TPO) (continuare)</b>				
DiaSorin LIAISON (2)	IU/mL	13.9	9.49	– 18.2
DiaSorin LIAISON XL (2)	IU/mL	15.1	9.13	– 21.1
Mindray CL-Series/Wiener lab. CLIA Series (2)	IU/mL	10.5	7.16	– 13.9
Monobind AccuBind ELISA (2)	IU/mL	8.21	5.50	– 10.9
Monobind AccuLite CLIA (2)	IU/mL	11.1	7.45	– 14.8
Roche cobas e801	IU/mL	40.7	26.5	– 55.0
Roche Elecsys/E170/Cobas e Systems	IU/mL	29.8	20.5	– 39.0
Shanghai Kehua Bio-Engineering (KHB) Unico 1800/1820 (2)	IU/mL	§		
Siemens ADVIA Centaur CP (aTPO)	U/mL	<37.0		
Siemens ADVIA Centaur Systems (aTPO)	U/mL		<28.0	– 48.8
Siemens Atellica Solution (aTPO) (IM)	U/mL	<28.0		
Siemens IMMULITE / IMMULITE 1000	IU/mL	14.8	11.0	– 18.5
Siemens IMMULITE 2000 / 2000 XPi	IU/mL	<10.0		
SNIBE Maglumi Series (2)	IU/mL	▲		
Tosoh AIA-PACK TPOAb (2)	IU/mL	<8.16		
<b>ERITROPOIETINĂ (EPO)</b>				
Beckman Coulter Access / 2 / 2i	mIU/mL	15.5	12.9	– 18.1
Beckman Coulter UniCel Dxl	mIU/mL	16.0	13.3	– 18.7
Siemens ADVIA Centaur Systems	mIU/mL	26.8	21.8	– 31.9
Siemens IMMULITE 2000 / 2000 XPi (L2KEPN)	mIU/mL	14.6	8.45	– 20.8
<b>IGF-1/SOMATOMEDINĂ C</b>				
Beckman Coulter IGF-1 IRMA (2)	ng/mL	54.4	40.8	– 68.0
Cisbio Bioassays IGF1-RIACT (2)	ng/mL	59.0	47.0	– 71.0
DiaSorin LIAISON (2)	ng/mL	58.0	34.7	– 81.3
DiaSorin LIAISON XL (2)	ng/mL	67.9	54.0	– 81.9
IDS-iSYS IGF-1 (2)	ng/mL	51.4	43.2	– 59.6
Mediagnost IGF-I ELISA (IGFBP-blocked) (2)	ng/mL	71.1	49.1	– 93.1
Mediagnost IGF-I RIA (IGFBP-blocked) (2)	ng/mL	71.1	49.1	– 93.1
Roche Elecsys/E170/cobas e Systems	ng/mL	37.4	27.6	– 47.2
Siemens IMMULITE 2000 / 2000 XPi (Kit Lot 482-485, >488)	ng/mL	54.2	34.3	– 74.0
Siemens IMMULITE 2000 / 2000 XPi (Re-Standardized)	ng/mL	59.2	38.3	– 80.2
SNIBE Maglumi Series (2)	ng/mL	▲		
<b>OSTEOCALCINĂ</b>				
Cisbio Bioassays OSTEO-RIACT (2)	ng/mL	25.0	20.0	– 30.0
DiaSorin LIAISON (2)	ng/mL	▲		
DiaSorin LIAISON XL (2)	ng/mL	▲		
IDS-iSYS N-Mid Osteocalcin (2)	ng/mL	16.1	13.5	– 18.7
Roche Elecsys/E170/Cobas e Systems	ng/mL	30.3	24.2	– 36.3
Siemens IMMULITE 2000 / 2000 XPi	ng/mL	13.7	8.71	– 18.7
Snibe MAGLUMI (2)	ng/mL	§		
<b>PEPTIDĂ C</b>				
Abbott Alinity i Analyzer	ng/mL	1.04	0.782	– 1.29
Abbott ARCHITECT iSystems	ng/mL	1.03	0.773	– 1.28
Beckman Coulter Access / 2 / 2i (2)	ng/mL	0.950	0.760	– 1.14
Beckman Coulter C-Peptide IRMA (2)	ng/mL	1.24	0.990	– 1.49
Beckman Coulter UniCel Dxl Series (2)	ng/mL	0.980	0.780	– 1.17
Calbiotech ELISA (2)	ng/mL		<0.200	– 0.250
Cisbio Bioassays IRMA-CPEP (2)	ng/mL	1.05	0.841	– 1.26

**METODĂ (1)**

	Unități	Nivel LTA - 64951L		
		Medie	Interval	
<b>PEPTIDĂ C (continuare)</b>				
DIAsource ImmunoAssays C-PEP II RIA-CT (2)	ng/mL	§		
Mindray CL-Series/Wiener lab. CLIA Series (2)	ng/mL	1.05	0.770	– 1.34
Monobind AccuBind ELISA (2)	ng/mL	0.800	0.530	– 1.06
Monobind AccuLite CLIA (2)	ng/mL	0.700	0.460	– 0.940
Ortho Vitros Microwell Series (C-pep) (2)	ng/mL	1.24	0.931	– 1.55
Roche Elecsys/E170/Cobas e Systems	ng/mL	1.19	0.925	– 1.46
Siemens ADVIA Centaur CP (CpS)	ng/mL	0.905	0.652	– 1.16
Siemens ADVIA Centaur CP (CpS) (Rgt lots 206 & calibrator lot 08 & higher)	ng/mL	0.965	0.713	– 1.22
Siemens ADVIA Centaur Systems (CpS)	ng/mL	1.07	0.780	– 1.35
Siemens ADVIA Centaur Systems (CpS) (Rgt lots 206 & calibrator lot 08 & higher)	ng/mL	1.23	0.942	– 1.52
Siemens Atellica CI 1900 Analyzer (CpS) (IM)	ng/mL	1.00	0.791	– 1.22
Siemens Atellica CI Analyzer (CpS) (IM) (Rgt 207 & calibrator 09 & higher)	ng/mL	1.23	1.02	– 1.44
Siemens Atellica Solution (CpS) (IM)	ng/mL	1.00	0.791	– 1.22
Siemens Atellica Solution (CpS) (IM) (Rgt 207 & calibrator 09 & higher)	ng/mL	1.23	1.02	– 1.44
Siemens IMMULITE / IMMULITE 1000	ng/mL	0.933	0.784	– 1.08
Siemens IMMULITE 2000 / 2000 XPI	ng/mL	0.928	0.675	– 1.18
Tosoh ST AIA-PACK C-Peptide II (2)	ng/mL	1.09	0.760	– 1.42
<b>PTH (INTACT)</b>				
Abbott Alinity i Analyzer Routine Mode	pg/mL	49.7	29.2	– 70.2
Abbott Alinity i Analyzer STAT Mode	pg/mL	40.2	23.8	– 56.6
Abbott ARCHITECT iSystems Routine Mode	pg/mL	39.5	23.1	– 55.9
Abbott ARCHITECT iSystems STAT Mode	pg/mL	39.8	27.4	– 52.1
Beckman Coulter Access / 2 / 2i Intraoperative Mode (2)	pg/mL	29.7	23.8	– 35.6
Beckman Coulter Access / 2 / 2i Routine Mode (2)	pg/mL	34.7	27.8	– 41.7
Beckman Coulter PTH IRMA (2)	pg/mL	30.0	19.5	– 40.5
Beckman Coulter UniCel Dxl Intraoperative Mode (2)	pg/mL	28.4	22.7	– 34.1
Beckman Coulter UniCel Dxl Routine Mode (2)	pg/mL	33.5	26.8	– 40.2
Biomerica ELISA (2)	pg/mL	39.0	23.5	– 54.5
bioMérieux VIDAS	pg/mL	21.4	12.9	– 29.9
Calbiotech Bio-intact PTH ELISA (2)	pg/mL	§		
Calbiotech intact PTH ELISA (2)	pg/mL	36.0	23.9	– 48.2
DiaSorin LIAISON 1-84 PTH (310630) (2)	pg/mL	15.8	10.1	– 21.5
DiaSorin LIAISON N-TACT PTH II (317910) (2)	pg/mL	37.4	26.2	– 48.6
DiaSorin LIAISON XL 1-84 PTH (310630) (2)	pg/mL	15.9	10.2	– 21.6
DiaSorin LIAISON XL N-TACT PTH II (317910) (2)	pg/mL	39.1	27.4	– 50.8
DIAsource ImmunoAssays hPTH-120 min-IRMA (2)	pg/mL	17.6	14.1	– 21.2
IDS-iSYS Intact PTH (2)	pg/mL	28.9	23.1	– 34.7
Mindray CL-Series/Wiener lab. CLIA Series (2)	pg/mL	22.8	16.6	– 28.9
Ortho Vitros MicroWell Series (iPTH) (2)	pg/mL	23.0	13.1	– 32.9
Roche cobas e 801	pg/mL	26.7	19.9	– 33.5
Roche Elecsys / Cobas e411	pg/mL	26.7	19.8	– 33.5
Roche Elecsys / Cobas e411 (STAT)	pg/mL	26.1	19.4	– 32.7
Roche Elecsys / E170 / cobas e Systems	pg/mL	21.9	13.6	– 30.1
Roche Elecsys / E170 / cobas e Systems (STAT)	pg/mL	20.5	12.3	– 28.7
Siemens ADVIA Centaur CP (PTH) (REF 10699154/10699155)	pg/mL	37.2	25.3	– 49.1
Siemens ADVIA Centaur Systems (PTH) (REF 10699154/10699155)	pg/mL	37.1	26.8	– 47.3
<b>Siemens Atellica CI 1900 Analyzer (PTH) (IM) →</b>	<b>pg/mL</b>	<b>36.6</b>	<b>23.0</b>	<b>– 50.2</b>

**METODĂ (1)**

	Unități	Nivel LTA - 64951L			
		Medie		Interval	
<b>PTH (INTACT) (continuare)</b>					
Siemens Atellica Solution (PTH) (IM)	pg/mL	36.6	23.0	–	50.2
Siemens IMMULITE / IMMULITE 1000	pg/mL	25.8	17.0	–	34.7
Siemens IMMULITE / IMMULITE 1000 (Turbo)	pg/mL	22.5	13.3	–	31.8
Siemens IMMULITE 2000 / 2000 XPi	pg/mL	29.9	19.9	–	39.9
SNIBE Maglumi Series (2)	pg/mL	▲			
Tosoh ST AIA-PACK Intact PTH (2)	pg/mL	14.9	10.4	–	19.4

**METODĂ (1)**

	Unități	Nivel 1 - 64951			
		Medie		Interval	
<b>25-HIDROXI VITAMINA D (TOTAL)</b>					
Abbott Alinity i Analyzer	ng/mL	9.23	6.54	–	11.9
Abbott ARCHITECT iSystems (3L52-25/35)	ng/mL	9.35	<8.00	–	12.4
Abbott ARCHITECT iSystems (3L52-27/37)	ng/mL	<13.0			
Abbott ARCHITECT iSystems (5P02-25/35/30)	ng/mL	9.47	6.44	–	12.5
Beckman Coulter Access / 2 / 2i	ng/mL	18.5	12.6	–	24.4
Beckman Coulter UniCel Dxl	ng/mL	22.2	15.3	–	29.1
bioMérieux VIDAS 25-OH Vitamin D Total	ng/mL	19.9	13.7	–	26.1
Bio-Rad BioPlex 2200 (2)	ng/mL	17.9	11.7	–	24.1
Calbiotech 25-OH Vitamin D ELISA (2)	ng/mL	17.8	11.9	–	23.7
DiaSorin LIAISON 25-OH Vitamin D TOTAL (2)	ng/mL	19.0	13.3	–	24.7
DiaSorin LIAISON XL 25-OH Vitamin D TOTAL (2)	ng/mL	19.3	13.5	–	25.1
DiaSorin LIAISON XS 25-OH Vitamin D TOTAL (2)	ng/mL	16.2	11.3	–	21.1
DIAsource ImmunoAssays 25OH-Vit. D3 RIA-CT (2)	ng/mL	6.46	4.52	–	8.40
DIAZYME (EZ Vitamin D) (2)	ng/mL	27.0	20.2	–	33.7
IDS-iSYS 25 VitDS (REF IS-2500N) (2)	ng/mL	30.6	19.6	–	41.6
Ortho Vitros MicroWell Series (tVitD) (Cat. 684 2894) (2)	ng/mL	25.7	9.74	–	41.7
Ortho Vitros MicroWell Series (tVitD) (Cat. 684 4056) (2)	ng/mL	25.7	<12.8	–	41.7
Roche cobas e 801 (Vitamin D III, REF 09038086) (2)	ng/mL	<6.00			
Roche cobas e801 (Vitamin D II, REF 07028148190)	ng/mL	24.5	17.1	–	31.8
Roche Elecsys/E170/Cobas e Systems (Roche Vit. D Total, REF 05894913)	ng/mL	§			
Roche Elecsys/E170/cobas e Systems (Vitamin D II, REF. 07464215160/REF. 07464215190)	ng/mL	19.4	13.5	–	25.3
Siemens ADVIA Centaur Systems (VitD)	ng/mL	41.6	29.3	–	54.0
<b>Siemens Atellica CI 1900 Analyzer (VitD) (IM) →</b>	<b>ng/mL</b>	<b>32.5</b>	<b>22.7</b>	<b>–</b>	<b>42.3</b>
Siemens Atellica Solution (VitD) (IM)	ng/mL	32.5	22.7	–	42.3
Siemens Dimension EXL (LOCI)	ng/mL	5.93	<5.00	–	10.5
SNIBE Maglumi Series (2)	ng/mL	▲			
Tosoh ST AIA-PACK 25-OH Vitamin D (2)	ng/mL	31.2	21.8	–	40.6
<b>ANTI-TIROGLOBULINĂ (ANTI-TG)</b>					
Abbott Alinity i Analyzer	IU/mL	36.7	28.3	–	45.2
Abbott ARCHITECT iSystems	IU/mL	37.1	28.4	–	45.7
Beckman Coulter Access / 2 / 2i	IU/mL	63.2	46.6	–	79.8
Beckman Coulter Anti-hTg IRMA (2)	IU/mL	69.4	45.1	–	93.7
Beckman Coulter UniCel Dxl	IU/mL	§			
bioMérieux VIDAS	IU/mL	66.2	47.3	–	85.2

**METODĂ (1)**

	Unități	Nivel 1 - 64951		
		Medie	Interval	
<b>ANTI-TIROGLOBULINĂ (ANTI-TG) (continuare)</b>				
DiaSorin LIAISON (2)	IU/mL	84.2	55.3	– 113
DiaSorin LIAISON XL (2)	IU/mL	74.9	53.8	– 96.1
DIASource ImmunoAssays TG Ab RIA (2)	IU/mL	61.9	46.4	– 77.3
Mindray CL-Series/Wiener lab. CLIA Series (2)	IU/mL	§		
Monobind AccuBind ELISA (2)	IU/mL	106	71.3	– 142
Monobind AccuLite CLIA (2)	IU/mL	120	80.2	– 159
Roche Elecsys/E170/Cobas e Systems	IU/mL	138	96.2	– 180
Shanghai Kehua Bio-Engineering (KHB) Unico 1800/1820 (2)	IU/mL	§		
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 348) (2)	U/mL	80.9	55.6	– 105
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 350) (2)	U/mL	74.5	51.5	– 97.5
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 352) (2)	U/mL	91.7	65.3	– 118
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 353) (2)	U/mL	75.9	52.6	– 99.2
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 355) (2)	U/mL	83.9	59.0	– 109
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 357) (2)	U/mL	83.6	58.8	– 108
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 359) (2)	U/mL	80.1	56.0	– 104
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 361) (2)	U/mL	85.8	60.5	– 111
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 363) (2)	U/mL	81.8	57.3	– 106
Siemens ADVIA Centaur CP (aTgII)	IU/mL	52.7	35.5	– 69.8
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 348) (2)	U/mL	99.5	71.5	– 127
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 350) (2)	U/mL	95.9	68.6	– 123
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 352) (2)	U/mL	105	75.9	– 134
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 353) (2)	U/mL	98.3	70.6	– 126
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 355) (2)	U/mL	111	80.7	– 141
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 357) (2)	U/mL	102	73.5	– 130
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 359) (2)	U/mL	103	74.3	– 132
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 361) (2)	U/mL	104	75.1	– 133
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 363) (2)	U/mL	78.5	54.7	– 102
Siemens ADVIA Centaur Systems (aTgII)	IU/mL	53.8	38.5	– 69.1
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 351)	U/mL	90.3	62.7	– 118
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 354)	U/mL	92.6	64.3	– 121
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 356)	U/mL	96.6	67.3	– 126
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 358)	U/mL	92.0	64.0	– 120
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 360)	U/mL	84.5	58.5	– 110
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 362)	U/mL	83.2	58.5	– 108
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 364) (2)	U/mL	82.4	57.8	– 107
Siemens Atellica Solution (aTgII)	IU/mL	56.6	40.5	– 72.7
Siemens IMMULITE / IMMULITE 1000	IU/mL	▲		
Siemens IMMULITE 2000 / 2000 XPI	IU/mL	34.8	24.6	– 44.9
SNIBE Maglumi Series (2)	IU/mL	▲		
Tosoh AIA-PACK TgAb (2)	IU/mL	65.1	40.4	– 89.8
<b>ANTI-TIROPEROXIDAZĂ (ANTI-TPO)</b>				
Abbott Alinity i Analyzer	IU/mL	24.1	18.0	– 30.2
Abbott ARCHITECT iSystems	IU/mL	26.9	20.1	– 33.7
Beckman Coulter Access / 2 / 2i	IU/mL	18.4	13.7	– 23.2
Beckman Coulter Anti-TPO RIA (2)	IU/mL	42.2	25.3	– 59.1
Beckman Coulter UniCel Dxl	IU/mL	17.3	13.4	– 21.1
bioMérieux VIDAS	IU/mL	19.5	14.4	– 24.7
DiaSorin LIAISON (2)	IU/mL	34.2	22.8	– 45.5

**METODĂ (1)**

	Unități	Nivel 1 - 64951		
		Medie	Interval	
<b>ANTI-TIROPEROXIDAZĂ (ANTI-TPO) (continuare)</b>				
DiaSorin LIAISON XL (2)	IU/mL	36.9	20.9	– 52.9
Mindray CL-Series/Wiener lab. CLIA Series (2)	IU/mL	24.4	16.6	– 32.3
Monobind AccuBind ELISA (2)	IU/mL	23.0	15.4	– 30.6
Monobind AccuLite CLIA (2)	IU/mL	21.8	14.6	– 29.0
Roche cobas e801	IU/mL	59.8	39.3	– 80.2
Roche Elecsys/E170/Cobas e Systems	IU/mL	46.6	32.9	– 60.3
Shanghai Kehua Bio-Engineering (KHB) Unico 1800/1820 (2)	IU/mL	§		
Siemens ADVIA Centaur CP (aTPO)	U/mL	71.3	54.7	– 87.8
Siemens ADVIA Centaur Systems (aTPO)	U/mL	74.8	45.9	– 104
Siemens Atellica Solution (aTPO) (IM)	U/mL	64.4	48.7	– 80.1
Siemens IMMULITE / IMMULITE 1000	IU/mL	36.3	27.0	– 45.6
Siemens IMMULITE 2000 / 2000 Xpi	IU/mL	28.1	21.0	– 35.2
SNIBE Maglumi Series (2)	IU/mL	▲		
Tosoh AIA-PACK TPOAb (2)	IU/mL	16.9	11.9	– 22.0
<b>ERITROPOIETINĂ (EPO)</b>				
Beckman Coulter Access / 2 / 2i	mIU/mL	15.2	12.6	– 17.7
Beckman Coulter UniCel DxI	mIU/mL	15.8	13.2	– 18.5
Siemens ADVIA Centaur Systems	mIU/mL	26.0	21.0	– 30.9
Siemens IMMULITE 2000 / 2000 Xpi (L2KEPN)	mIU/mL	14.5	8.34	– 20.6
<b>IGF-1/SOMATOMEDINĂ C</b>				
Beckman Coulter IGF-1 IRMA (2)	ng/mL	59.8	44.9	– 74.8
Cisbio Bioassays IGF1-RIACT (2)	ng/mL	63.0	50.0	– 76.0
DiaSorin LIAISON (2)	ng/mL	59.0	37.1	– 80.9
DiaSorin LIAISON XL (2)	ng/mL	70.0	53.7	– 86.4
IDS-iSYS IGF-1 (2)	ng/mL	52.2	43.8	– 60.6
Mediagnost IGF-I ELISA (IGFBP-blocked) (2)	ng/mL	71.3	50.9	– 91.8
Mediagnost IGF-I RIA (IGFBP-blocked) (2)	ng/mL	71.3	50.9	– 91.8
Roche Elecsys/E170/cobas e Systems	ng/mL	38.1	28.3	– 47.9
Siemens IMMULITE 2000 / 2000 Xpi (Kit Lot 482-485, >488)	ng/mL	56.0	35.7	– 76.2
Siemens IMMULITE 2000 / 2000 Xpi (Re-Standardized)	ng/mL	59.8	38.7	– 81.0
SNIBE Maglumi Series (2)	ng/mL	▲		
<b>OSTEOCALCINĂ</b>				
Cisbio Bioassays OSTEO-RIACT (2)	ng/mL	25.9	19.8	– 32.0
DiaSorin LIAISON (2)	ng/mL	▲		
DiaSorin LIAISON XL (2)	ng/mL	▲		
IDS-iSYS N-Mid Osteocalcin (2)	ng/mL	16.4	13.8	– 19.0
Roche Elecsys/E170/Cobas e Systems	ng/mL	30.4	24.3	– 36.4
Siemens IMMULITE 2000 / 2000 Xpi	ng/mL	14.3	9.04	– 19.5
Snibe MAGLUMI (2)	ng/mL	§		
<b>PEPTIDĂ C</b>				
Abbott Alinity i Analyzer	ng/mL	1.03	0.780	– 1.29
Abbott ARCHITECT iSystems	ng/mL	1.00	0.749	– 1.25
Beckman Coulter Access / 2 / 2i (2)	ng/mL	0.950	0.760	– 1.14
Beckman Coulter C-Peptide IRMA (2)	ng/mL	1.25	1.00	– 1.50
Beckman Coulter UniCel DxI Series (2)	ng/mL	0.990	0.790	– 1.18
Calbiotech ELISA (2)	ng/mL	0.210	<0.200	– 0.280
Cisbio Bioassays IRMA-CPEP (2)	ng/mL	1.05	0.841	– 1.26
DIAsource ImmunoAssays C-PEP II RIA-CT (2)	ng/mL	§		

**METODĂ (1)**

	Unități	Nivel 1 - 64951		
		Medie	Interval	
<b>PEPTIDĂ C (continuare)</b>				
Mindray CL-Series/Wiener lab. CLIA Series (2)	ng/mL	1.06	0.770	– 1.35
Monobind AccuBind ELISA (2)	ng/mL	0.890	0.600	– 1.18
Monobind AccuLite CLIA (2)	ng/mL	0.730	0.490	– 0.970
Ortho Vitros Microwell Series (C-pep) (2)	ng/mL	1.26	0.945	– 1.58
Roche Elecsys/E170/Cobas e Systems	ng/mL	1.20	0.932	– 1.47
Siemens ADVIA Centaur CP (CpS)	ng/mL	0.924	0.671	– 1.18
Siemens ADVIA Centaur CP (CpS) (Rgt lots 206 & calibrator lot 08 & higher)	ng/mL	0.985	0.733	– 1.24
Siemens ADVIA Centaur Systems (CpS)	ng/mL	1.04	0.756	– 1.32
Siemens ADVIA Centaur Systems (CpS) (Rgt lots 206 & calibrator lot 08 & higher)	ng/mL	1.23	0.948	– 1.51
Siemens Atellica CI 1900 Analyzer (CpS) (IM)	ng/mL	0.995	0.783	– 1.21
Siemens Atellica CI Analyzer (CpS) (IM) (Rgt 207 & calibrator 09 & higher)	ng/mL	1.23	1.02	– 1.44
Siemens Atellica Solution (CpS) (IM)	ng/mL	0.995	0.783	– 1.21
Siemens Atellica Solution (CpS) (IM) (Rgt 207 & calibrator 09 & higher)	ng/mL	1.23	1.02	– 1.44
Siemens IMMULITE / IMMULITE 1000	ng/mL	0.911	0.763	– 1.06
Siemens IMMULITE 2000 / 2000 XPI	ng/mL	0.949	0.697	– 1.20
Tosoh ST AIA-PACK C-Peptide II (2)	ng/mL	1.00	0.700	– 1.30
<b>PTH (INTACT)</b>				
Abbott Alinity i Analyzer Routine Mode	pg/mL	52.0	30.5	– 73.4
Abbott Alinity i Analyzer STAT Mode	pg/mL	41.6	24.6	– 58.6
Abbott ARCHITECT iSystems Routine Mode	pg/mL	40.3	23.6	– 57.0
Abbott ARCHITECT iSystems STAT Mode	pg/mL	41.0	28.4	– 53.7
Beckman Coulter Access / 2 / 2i Intraoperative Mode (2)	pg/mL	30.9	24.7	– 37.1
Beckman Coulter Access / 2 / 2i Routine Mode (2)	pg/mL	35.9	28.7	– 43.0
Beckman Coulter PTH IRMA (2)	pg/mL	32.7	22.9	– 42.5
Beckman Coulter UniCel DxI Intraoperative Mode (2)	pg/mL	30.6	24.5	– 36.7
Beckman Coulter UniCel DxI Routine Mode (2)	pg/mL	34.7	27.8	– 41.6
Biomerica ELISA (2)	pg/mL	41.3	28.5	– 54.2
bioMérieux VIDAS	pg/mL	21.2	12.7	– 29.7
Calbiotech Bio-intact PTH ELISA (2)	pg/mL	§		
Calbiotech intact PTH ELISA (2)	pg/mL	35.7	23.7	– 47.7
DiaSorin LIAISON 1-84 PTH (310630) (2)	pg/mL	16.9	10.8	– 23.0
DiaSorin LIAISON N-TACT PTH II (317910) (2)	pg/mL	38.9	27.2	– 50.6
DiaSorin LIAISON XL 1-84 PTH (310630) (2)	pg/mL	17.0	10.9	– 23.2
DiaSorin LIAISON XL N-TACT PTH II (317910) (2)	pg/mL	41.2	28.8	– 53.5
DIAsource ImmunoAssays hPTH-120 min-IRMA (2)	pg/mL	18.3	14.6	– 21.9
IDS-iSYS Intact PTH (2)	pg/mL	29.3	23.4	– 35.2
Mindray CL-Series/Wiener lab. CLIA Series (2)	pg/mL	23.2	17.0	– 29.5
Ortho Vitros MicroWell Series (iPTH) (2)	pg/mL	23.3	13.2	– 33.4
Roche cobas e 801	pg/mL	28.9	21.6	– 36.2
Roche Elecsys / Cobas e411	pg/mL	28.4	21.2	– 35.7
Roche Elecsys / Cobas e411 (STAT)	pg/mL	27.6	20.6	– 34.5
Roche Elecsys / E170 / cobas e Systems	pg/mL	24.0	15.7	– 32.4
Roche Elecsys / E170 / cobas e Systems (STAT)	pg/mL	22.8	14.7	– 31.0
Siemens ADVIA Centaur CP (PTH) (REF 10699154/10699155)	pg/mL	38.7	26.5	– 50.8
Siemens ADVIA Centaur Systems (PTH) (REF 10699154/10699155)	pg/mL	39.1	28.5	– 49.7
<b>Siemens Atellica CI 1900 Analyzer (PTH) (IM) →</b>	<b>pg/mL</b>	<b>37.0</b>	<b>23.3</b>	<b>– 50.7</b>
Siemens Atellica Solution (PTH) (IM)	pg/mL	37.0	23.3	– 50.7



**METODĂ (1)**

	Unități	Nivel 1 - 64951			
		Medie		Interval	
<b>PTH (INTACT) (continuare)</b>					
Siemens IMMULITE / IMMULITE 1000	pg/mL	26.4	17.4	–	35.4
Siemens IMMULITE / IMMULITE 1000 (Turbo)	pg/mL	20.4	11.2	–	29.7
Siemens IMMULITE 2000 / 2000 XPI	pg/mL	30.2	20.1	–	40.2
SNIBE Maglumi Series (2)	pg/mL	▲			
Tosoh ST AIA-PACK Intact PTH (2)	pg/mL	15.0	10.5	–	19.5

**METODĂ (1)**

	Unități	Nivel 2 - 64952			
		Medie		Interval	
<b>25-HIDROXI VITAMINA D (TOTAL)</b>					
Abbott Alinity i Analyzer	ng/mL	18.3	13.3	–	23.4
Abbott ARCHITECT iSystems (3L52-25/35)	ng/mL	19.0	13.6	–	24.3
Abbott ARCHITECT iSystems (3L52-27/37)	ng/mL	19.0	13.6	–	24.3
Abbott ARCHITECT iSystems (5P02-25/35/30)	ng/mL	19.1	13.7	–	24.4
Beckman Coulter Access / 2 / 2i	ng/mL	46.0	32.4	–	59.6
Beckman Coulter UniCel Dxl	ng/mL	51.9	36.7	–	67.1
bioMérieux VIDAS 25-OH Vitamin D Total	ng/mL	28.5	20.0	–	37.1
Bio-Rad BioPlex 2200 (2)	ng/mL	36.8	30.7	–	43.0
Calbiotech 25-OH Vitamin D ELISA (2)	ng/mL	28.7	19.1	–	38.3
DiaSorin LIAISON 25-OH Vitamin D TOTAL (2)	ng/mL	43.4	30.4	–	56.5
DiaSorin LIAISON XL 25-OH Vitamin D TOTAL (2)	ng/mL	42.1	29.5	–	54.7
DiaSorin LIAISON XS 25-OH Vitamin D TOTAL (2)	ng/mL	36.1	25.3	–	46.9
DIAsource ImmunoAssays 25OH-Vit. D3 RIA-CT (2)	ng/mL	25.1	20.1	–	30.2
DIAZYME (EZ Vitamin D) (2)	ng/mL	43.2	32.4	–	54.0
IDS-iSYS 25 VitDS (REF IS-2500N) (2)	ng/mL	61.8	47.0	–	76.6
Ortho Vitros MicroWell Series (tVitD) (Cat. 684 2894) (2)	ng/mL	58.1	35.1	–	81.1
Ortho Vitros MicroWell Series (tVitD) (Cat. 684 4056) (2)	ng/mL	58.1	35.1	–	81.1
Roche cobas e 801 (Vitamin D III, REF 09038086) (2)	ng/mL	25.6	17.2	–	34.0
Roche cobas e801 (Vitamin D II, REF 07028148190)	ng/mL	47.0	33.5	–	60.5
Roche Elecsys/E170/Cobas e Systems (Roche Vit. D Total, REF 05894913)	ng/mL	§			
Roche Elecsys/E170/cobas e Systems (Vitamin D II, REF. 07464215160/REF. 07464215190)	ng/mL	41.5	29.6	–	53.3
Siemens ADVIA Centaur Systems (VitD)	ng/mL	94.4	67.3	–	121
<b>Siemens Atellica CI 1900 Analyzer (VitD) (IM) →</b>	<b>ng/mL</b>	<b>82.1</b>	<b>58.4</b>	–	<b>106</b>
Siemens Atellica Solution (VitD) (IM)	ng/mL	82.1	58.4	–	106
Siemens Dimension EXL (LOCI)	ng/mL	17.5	10.9	–	24.2
SNIBE Maglumi Series (2)	ng/mL	▲			
Tosoh ST AIA-PACK 25-OH Vitamin D (2)	ng/mL	56.0	39.2	–	72.8
<b>ANTI-TIROGLOBULINĂ (ANTI-TG)</b>					
Abbott Alinity i Analyzer	IU/mL	70.9	56.0	–	85.7
Abbott ARCHITECT iSystems	IU/mL	70.1	55.9	–	84.2
Beckman Coulter Access / 2 / 2i	IU/mL	119	86.9	–	151
Beckman Coulter Anti-hTg IRMA (2)	IU/mL	142	92.5	–	192
Beckman Coulter UniCel Dxl	IU/mL	§			
bioMérieux VIDAS	IU/mL	125	89.2	–	160
DiaSorin LIAISON (2)	IU/mL	146	96.0	–	196

**METODĂ (1)**

	Unități	Nivel 2 - 64952		
		Medie	Interval	
<b>ANTI-TIROGLOBULINĂ (ANTI-TG) (continuare)</b>				
DiaSorin LIAISON XL (2)	IU/mL	128	102	– 153
DIAsource ImmunoAssays TG Ab RIA (2)	IU/mL	113	84.6	– 141
Mindray CL-Series/Wiener lab. CLIA Series (2)	IU/mL	§		
Monobind AccuBind ELISA (2)	IU/mL	181	121	– 240
Monobind AccuLite CLIA (2)	IU/mL	176	118	– 234
Roche Elecsys/E170/Cobas e Systems	IU/mL	301	212	– 389
Shanghai Kehua Bio-Engineering (KHB) Unico 1800/1820 (2)	IU/mL	§		
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 348) (2)	U/mL	134	99.1	– 169
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 350) (2)	U/mL	126	92.7	– 159
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 352) (2)	U/mL	150	112	– 188
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 353) (2)	U/mL	124	91.1	– 157
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 355) (2)	U/mL	138	102	– 174
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 357) (2)	U/mL	133	98.3	– 168
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 359) (2)	U/mL	133	98.3	– 168
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 361) (2)	U/mL	139	103	– 175
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 363) (2)	U/mL	136	101	– 171
Siemens ADVIA Centaur CP (aTgII)	IU/mL	98.0	68.0	– 128
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 348) (2)	U/mL	156	117	– 195
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 350) (2)	U/mL	157	117	– 197
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 352) (2)	U/mL	164	123	– 205
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 353) (2)	U/mL	149	111	– 187
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 355) (2)	U/mL	179	135	– 223
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 357) (2)	U/mL	170	128	– 212
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 359) (2)	U/mL	170	128	– 212
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 361) (2)	U/mL	161	121	– 201
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 363) (2)	U/mL	129	95.1	– 163
Siemens ADVIA Centaur Systems (aTgII)	IU/mL	103	74.4	– 132
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 351)	U/mL	150	105	– 194
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 354)	U/mL	151	106	– 196
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 356)	U/mL	157	111	– 203
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 358)	U/mL	150	106	– 195
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 360)	U/mL	141	98.8	– 182
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 362)	U/mL	142	105	– 179
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 364) (2)	U/mL	136	101	– 171
Siemens Atellica Solution (aTgII)	IU/mL	107	77.3	– 137
Siemens IMMULITE / IMMULITE 1000	IU/mL	▲		
Siemens IMMULITE 2000 / 2000 XPI	IU/mL	66.4	47.4	– 85.4
SNIBE Maglumi Series (2)	IU/mL	▲		
Tosoh AIA-PACK TgAb (2)	IU/mL	125	79.2	– 171
<b>ANTI-TIROPEROXIDAZĂ (ANTI-TPO)</b>				
Abbott Alinity i Analyzer	IU/mL	64.4	48.1	– 80.7
Abbott ARCHITECT iSystems	IU/mL	73.0	54.5	– 91.5
Beckman Coulter Access / 2 / 2i	IU/mL	48.8	36.3	– 61.3
Beckman Coulter Anti-TPO RIA (2)	IU/mL	167	100	– 233
Beckman Coulter UniCel DxI	IU/mL	47.8	34.7	– 60.8
bioMérieux VIDAS	IU/mL	49.9	36.7	– 63.1
DiaSorin LIAISON (2)	IU/mL	78.0	49.4	– 107
DiaSorin LIAISON XL (2)	IU/mL	88.3	50.8	– 126

**METODĂ (1)**

	Unități	Nivel 2 - 64952			
		Medie		Interval	
<b>ANTI-TIROPEROXIDAZĂ (ANTI-TPO) (continuare)</b>					
Mindray CL-Series/Wiener lab. CLIA Series (2)	IU/mL	64.0	43.5	–	84.5
Monobind AccuBind ELISA (2)	IU/mL	56.9	38.1	–	75.7
Monobind AccuLite CLIA (2)	IU/mL	58.7	39.4	–	78.1
Roche cobas e801	IU/mL	122	81.8	–	163
Roche Elecsys/E170/Cobas e Systems	IU/mL	90.0	64.6	–	116
Shanghai Kehua Bio-Engineering (KHB) Unico 1800/1820 (2)	IU/mL	§			
Siemens ADVIA Centaur CP (aTPO)	U/mL	191	152	–	230
Siemens ADVIA Centaur Systems (aTPO)	U/mL	196	147	–	246
Siemens Atellica Solution (aTPO) (IM)	U/mL	184	145	–	223
Siemens IMMULITE / IMMULITE 1000	IU/mL	97.2	72.2	–	122
Siemens IMMULITE 2000 / 2000 XPI	IU/mL	73.6	54.9	–	92.2
SNIBE Maglumi Series (2)	IU/mL	▲			
Tosoh AIA-PACK TPOAb (2)	IU/mL	43.0	29.4	–	56.7
<b>ERITROPOIETINĂ (EPO)</b>					
Beckman Coulter Access / 2 / 2i	mIU/mL	63.7	52.8	–	74.7
Beckman Coulter UniCel DxI	mIU/mL	66.6	55.1	–	78.0
Siemens ADVIA Centaur Systems	mIU/mL	113	94.9	–	130
Siemens IMMULITE 2000 / 2000 XPI (L2KEPN)	mIU/mL	65.6	42.7	–	88.4
<b>IGF-1/SOMATOMEDINĂ C</b>					
Beckman Coulter IGF-1 IRMA (2)	ng/mL	252	189	–	315
Cisbio Bioassays IGF1-RIACT (2)	ng/mL	268	206	–	330
DiaSorin LIAISON (2)	ng/mL	236	148	–	324
DiaSorin LIAISON XL (2)	ng/mL	272	217	–	326
IDS-iSYS IGF-1 (2)	ng/mL	195	164	–	226
Mediagnost IGF-I ELISA (IGFBP-blocked) (2)	ng/mL	250	205	–	295
Mediagnost IGF-I RIA (IGFBP-blocked) (2)	ng/mL	250	205	–	295
Roche Elecsys/E170/cobas e Systems	ng/mL	138	115	–	161
Siemens IMMULITE 2000 / 2000 XPI (Kit Lot 482-485, >488)	ng/mL	242	179	–	304
Siemens IMMULITE 2000 / 2000 XPI (Re-Standardized)	ng/mL	216	159	–	273
SNIBE Maglumi Series (2)	ng/mL	▲			
<b>OSTEOCALCINĂ</b>					
Cisbio Bioassays OSTEO-RIACT (2)	ng/mL	86.0	68.0	–	104
DiaSorin LIAISON (2)	ng/mL	▲			
DiaSorin LIAISON XL (2)	ng/mL	▲			
IDS-iSYS N-Mid Osteocalcin (2)	ng/mL	67.3	56.5	–	78.1
Roche Elecsys/E170/Cobas e Systems	ng/mL	102	71.2	–	132
Siemens IMMULITE 2000 / 2000 XPI	ng/mL	52.3	31.6	–	73.1
Snibe MAGLUMI (2)	ng/mL	§			
<b>PEPTIDĂ C</b>					
Abbott Alinity i Analyzer	ng/mL	4.79	3.62	–	5.95
Abbott ARCHITECT iSystems	ng/mL	4.49	3.42	–	5.57
Beckman Coulter Access / 2 / 2i (2)	ng/mL	4.56	3.64	–	5.47
Beckman Coulter C-Peptide IRMA (2)	ng/mL	5.60	4.48	–	6.72
Beckman Coulter UniCel DxI Series (2)	ng/mL	4.72	3.78	–	5.67
Calbiotech ELISA (2)	ng/mL	1.64	1.09	–	2.19
Cisbio Bioassays IRMA-CPEP (2)	ng/mL	4.68	3.72	–	5.65
DIAsource ImmunoAssays C-PEP II RIA-CT (2)	ng/mL	§			
Mindray CL-Series/Wiener lab. CLIA Series (2)	ng/mL	5.16	3.77	–	6.56

**METODĂ (1)**

	Unități	Nivel 2 - 64952			
		Medie		Interval	
<b>PEPTIDĂ C (continuare)</b>					
Monobind AccuBind ELISA (2)	ng/mL	3.52	2.36	–	4.68
Monobind AccuLite CLIA (2)	ng/mL	2.90	1.94	–	3.86
Ortho Vitros Microwell Series (C-pep) (2)	ng/mL	5.58	4.17	–	6.99
Roche Elecsys/E170/Cobas e Systems	ng/mL	4.82	3.69	–	5.95
Siemens ADVIA Centaur CP (CpS)	ng/mL	3.82	2.89	–	4.74
Siemens ADVIA Centaur CP (CpS) (Rgt lots 206 & calibrator lot 08 & higher)	ng/mL	4.20	3.27	–	5.13
Siemens ADVIA Centaur Systems (CpS)	ng/mL	4.46	3.47	–	5.45
Siemens ADVIA Centaur Systems (CpS) (Rgt lots 206 & calibrator lot 08 & higher)	ng/mL	4.68	3.69	–	5.67
Siemens Atellica CI 1900 Analyzer (CpS) (IM)	ng/mL	4.02	3.33	–	4.71
Siemens Atellica CI Analyzer (CpS) (IM) (Rgt 207 & calibrator 09 & higher)	ng/mL	4.83	4.14	–	5.52
Siemens Atellica Solution (CpS) (IM)	ng/mL	4.02	3.33	–	4.71
Siemens Atellica Solution (CpS) (IM) (Rgt 207 & calibrator 09 & higher)	ng/mL	4.83	4.14	–	5.52
Siemens IMMULITE / IMMULITE 1000	ng/mL	4.99	3.87	–	6.10
Siemens IMMULITE 2000 / 2000 XPi	ng/mL	5.01	3.77	–	6.24
Tosoh ST AIA-PACK C-Peptide II (2)	ng/mL	4.84	3.39	–	6.29
<b>PTH (INTACT)</b>					
Abbott Alinity i Analyzer Routine Mode	pg/mL	380	222	–	537
Abbott Alinity i Analyzer STAT Mode	pg/mL	311	183	–	439
Abbott ARCHITECT iSystems Routine Mode	pg/mL	291	169	–	413
Abbott ARCHITECT iSystems STAT Mode	pg/mL	303	225	–	382
Beckman Coulter Access / 2 / 2i Intraoperative Mode (2)	pg/mL	224	179	–	268
Beckman Coulter Access / 2 / 2i Routine Mode (2)	pg/mL	267	213	–	320
Beckman Coulter PTH IRMA (2)	pg/mL	268	201	–	335
Beckman Coulter UniCel DxI Intraoperative Mode (2)	pg/mL	232	185	–	278
Beckman Coulter UniCel DxI Routine Mode (2)	pg/mL	268	214	–	321
Biomerica ELISA (2)	pg/mL	292	254	–	331
bioMérieux VIDAS	pg/mL	160	117	–	204
Calbiotech Bio-intact PTH ELISA (2)	pg/mL	§			
Calbiotech intact PTH ELISA (2)	pg/mL	280	186	–	373
DiaSorin LIAISON 1-84 PTH (310630) (2)	pg/mL	132	96.2	–	167
DiaSorin LIAISON N-TACT PTH II (317910) (2)	pg/mL	419	293	–	544
DiaSorin LIAISON XL 1-84 PTH (310630) (2)	pg/mL	138	101	–	176
DiaSorin LIAISON XL N-TACT PTH II (317910) (2)	pg/mL	446	312	–	579
DIAsource ImmunoAssays hPTH-120 min-IRMA (2)	pg/mL	121	96.9	–	145
IDS-iSYS Intact PTH (2)	pg/mL	216	185	–	246
Mindray CL-Series/Wiener lab. CLIA Series (2)	pg/mL	178	130	–	226
Ortho Vitros MicroWell Series (iPTH) (2)	pg/mL	183	110	–	255
Roche cobas e 801	pg/mL	178	136	–	220
Roche Elecsys / Cobas e411	pg/mL	170	129	–	210
Roche Elecsys / Cobas e411 (STAT)	pg/mL	156	119	–	193
Roche Elecsys / E170 / cobas e Systems	pg/mL	147	108	–	186
Roche Elecsys / E170 / cobas e Systems (STAT)	pg/mL	131	96.2	–	165
Siemens ADVIA Centaur CP (PTH) (REF 10699154/10699155)	pg/mL	310	251	–	368
Siemens ADVIA Centaur Systems (PTH) (REF 10699154/10699155)	pg/mL	309	245	–	372
<b>Siemens Atellica CI 1900 Analyzer (PTH) (IM) →</b>	<b>pg/mL</b>	<b>306</b>	<b>235</b>	<b>–</b>	<b>376</b>
Siemens Atellica Solution (PTH) (IM)	pg/mL	306	235	–	376
Siemens IMMULITE / IMMULITE 1000	pg/mL	226	167	–	285

**METODĂ (1)**

	Unități	Nivel 2 - 64952			
		Medie		Interval	
<b>PTH (INTACT) (continuare)</b>					
Siemens IMMULITE / IMMULITE 1000 (Turbo)	pg/mL	186	108	–	264
Siemens IMMULITE 2000 / 2000 XPI	pg/mL	266	195	–	336
SNIBE Maglumi Series (2)	pg/mL	▲			
Tosoh ST AIA-PACK Intact PTH (2)	pg/mL	114	79.9	–	148

**METODĂ (1)**

	Unități	Nivel 3 - 64953			
		Medie		Interval	
<b>25-HIDROXI VITAMINA D (TOTAL)</b>					
Abbott Alinity i Analyzer	ng/mL	42.0	31.6	–	52.4
Abbott ARCHITECT iSystems (3L52-25/35)	ng/mL	42.6	31.4	–	53.7
Abbott ARCHITECT iSystems (3L52-27/37)	ng/mL	42.6	31.4	–	53.7
Abbott ARCHITECT iSystems (5P02-25/35/30)	ng/mL	43.5	32.1	–	54.9
Beckman Coulter Access / 2 / 2i	ng/mL	110	78.3	–	>120
Beckman Coulter UniCel Dxl	ng/mL	>120			
bioMérieux VIDAS 25-OH Vitamin D Total	ng/mL	56.8	40.4	–	73.1
Bio-Rad BioPlex 2200 (2)	ng/mL	88.9	76.2	–	102
Calbiotech 25-OH Vitamin D ELISA (2)	ng/mL	57.9	38.6	–	77.1
DiaSorin LIAISON 25-OH Vitamin D TOTAL (2)	ng/mL	111	70.7	–	150
DiaSorin LIAISON XL 25-OH Vitamin D TOTAL (2)	ng/mL	107	75.1	–	140
DiaSorin LIAISON XS 25-OH Vitamin D TOTAL (2)	ng/mL	86.6	60.6	–	113
DIAsource ImmunoAssays 25OH-Vit. D3 RIA-CT (2)	ng/mL	85.7	64.3	–	107
DIAZYME (EZ Vitamin D) (2)	ng/mL	74.1	55.6	–	92.6
IDS-iSYS 25 VitDS (REF IS-2500N) (2)	ng/mL	>110			
Ortho Vitros MicroWell Series (tVitD) (Cat. 684 2894) (2)	ng/mL	139	101	–	>150
Ortho Vitros MicroWell Series (tVitD) (Cat. 684 4056) (2)	ng/mL	>126			
Roche cobas e 801 (Vitamin D III, REF 09038086) (2)	ng/mL	77.9	54.5	–	101
Roche cobas e801 (Vitamin D II, REF 07028148190)	ng/mL		75.7	–	>100
Roche Elecsys/E170/Cobas e Systems (Roche Vit. D Total, REF 05894913)	ng/mL	§			
Roche Elecsys/E170/cobas e Systems (Vitamin D II, REF. 07464215160/REF. 07464215190)	ng/mL	>100			
Siemens ADVIA Centaur Systems (VitD)	ng/mL	>150			
<b>Siemens Atellica CI 1900 Analyzer (VitD) (IM) →</b>	<b>ng/mL</b>	<b>&gt;150</b>			
Siemens Atellica Solution (VitD) (IM)	ng/mL	>150			
Siemens Dimension EXL (LOCI)	ng/mL	52.9	40.4	–	65.4
SNIBE Maglumi Series (2)	ng/mL	▲			
Tosoh ST AIA-PACK 25-OH Vitamin D (2)	ng/mL	119	83.0	–	>120
<b>ANTI-TIROGLOBULINĂ (ANTI-TG)</b>					
Abbott Alinity i Analyzer	IU/mL	163	130	–	195
Abbott ARCHITECT iSystems	IU/mL	165	135	–	195
Beckman Coulter Access / 2 / 2i	IU/mL	258	188	–	328
Beckman Coulter Anti-hTg IRMA (2)	IU/mL	335	218	–	452
Beckman Coulter UniCel Dxl	IU/mL	§			
bioMérieux VIDAS	IU/mL	283	203	–	362
DiaSorin LIAISON (2)	IU/mL	310	195	–	424
DiaSorin LIAISON XL (2)	IU/mL	280	203	–	357

**METODĂ (1)**

	Unități	Nivel 3 - 64953		
		Medie	Interval	
<b>ANTI-TIROGLOBULINĂ (ANTI-TG) (continuare)</b>				
DIAsource ImmunoAssays TG Ab RIA (2)	IU/mL	273	191	– 355
Mindray CL-Series/Wiener lab. CLIA Series (2)	IU/mL	§		
Monobind AccuBind ELISA (2)	IU/mL	366	245	– 487
Monobind AccuLite CLIA (2)	IU/mL	342	214	– 469
Roche Elecsys/E170/Cobas e Systems	IU/mL	996	708	– 1284
Shanghai Kehua Bio-Engineering (KHB) Unico 1800/1820 (2)	IU/mL	§		
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 348) (2)	U/mL	324	250	– 398
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 350) (2)	U/mL	304	234	– 374
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 352) (2)	U/mL	357	277	– 437
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 353) (2)	U/mL	301	232	– 370
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 355) (2)	U/mL	323	250	– 396
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 357) (2)	U/mL	326	252	– 400
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 359) (2)	U/mL	324	250	– 398
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 361) (2)	U/mL	311	240	– 382
Siemens ADVIA Centaur CP (aTG) (Rgt. Lot 363) (2)	U/mL	324	250	– 398
Siemens ADVIA Centaur CP (aTgII)	IU/mL	214	151	– 276
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 348) (2)	U/mL	384	298	– 470
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 350) (2)	U/mL	380	295	– 465
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 352) (2)	U/mL	389	302	– 476
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 353) (2)	U/mL	344	266	– 422
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 355) (2)	U/mL	383	297	– 469
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 357) (2)	U/mL	371	288	– 454
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 359) (2)	U/mL	359	278	– 440
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 361) (2)	U/mL	361	280	– 442
Siemens ADVIA Centaur Systems (aTG) (Rgt. Lot 363) (2)	U/mL	324	250	– 398
Siemens ADVIA Centaur Systems (aTgII)	IU/mL	223	161	– 285
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 351)	U/mL	329	235	– 424
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 354)	U/mL	349	249	– 449
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 356)	U/mL	354	252	– 455
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 358)	U/mL	348	248	– 448
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 360)	U/mL	337	240	– 434
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 362)	U/mL	329	254	– 404
Siemens Atellica Solution (aTG) (IM) (Rgt. Lot 364) (2)	U/mL	321	248	– 394
Siemens Atellica Solution (aTgII)	IU/mL	230	166	– 294
Siemens IMMULITE / IMMULITE 1000	IU/mL	▲		
Siemens IMMULITE 2000 / 2000 XPI	IU/mL	173	124	– 222
SNIBE Maglumi Series (2)	IU/mL	▲		
Tosoh AIA-PACK TgAb (2)	IU/mL	300	198	– 402
<b>ANTI-TIROPEROXIDAZĂ (ANTI-TPO)</b>				
Abbott Alinity i Analyzer	IU/mL	165	123	– 207
Abbott ARCHITECT iSystems	IU/mL	182	136	– 229
Beckman Coulter Access / 2 / 2i	IU/mL	124	92.3	– 156
Beckman Coulter Anti-TPO RIA (2)	IU/mL	516	336	– 697
Beckman Coulter UniCel Dxl	IU/mL	119	84.4	– 154
bioMérieux VIDAS	IU/mL	129	94.5	– 163
DiaSorin LIAISON (2)	IU/mL	196	109	– 283
DiaSorin LIAISON XL (2)	IU/mL	218	146	– 290
Mindray CL-Series/Wiener lab. CLIA Series (2)	IU/mL	164	111	– 216

**METODĂ (1)**

	Unități	Nivel 3 - 64953		
		Medie	Interval	
<b>ANTI-TIROPEROXIDAZĂ (ANTI-TPO) (continuare)</b>				
Monobind AccuBind ELISA (2)	IU/mL	141	94.2	– 187
Monobind AccuLite CLIA (2)	IU/mL	135	90.6	– 180
Roche cobas e801	IU/mL	279	188	– 370
Roche Elecsys/E170/Cobas e Systems	IU/mL	209	161	– 257
Shanghai Kehua Bio-Engineering (KHB) Unico 1800/1820 (2)	IU/mL	§		
Siemens ADVIA Centaur CP (aTPO)	U/mL	444	358	– 530
Siemens ADVIA Centaur Systems (aTPO)	U/mL	454	361	– 546
Siemens Atellica Solution (aTPO) (IM)	U/mL	439	351	– 527
Siemens IMMULITE / IMMULITE 1000	IU/mL	252	187	– 317
Siemens IMMULITE 2000 / 2000 XPI	IU/mL	207	155	– 260
SNIBE Maglumi Series (2)	IU/mL	▲		
Tosoh AIA-PACK TPOAb (2)	IU/mL	108	70.2	– 145
<b>ERITROPOIETINĂ (EPO)</b>				
Beckman Coulter Access / 2 / 2i	mIU/mL	131	108	– 154
Beckman Coulter UniCel DxI	mIU/mL	136	113	– 160
Siemens ADVIA Centaur Systems	mIU/mL	224	197	– 251
Siemens IMMULITE 2000 / 2000 XPI (L2KEPN)	mIU/mL	136	89.8	– 181
<b>IGF-1/SOMATOMEDINĂ C</b>				
Beckman Coulter IGF-1 IRMA (2)	ng/mL	590	442	– 737
Cisbio Bioassays IGF1-RIACT (2)	ng/mL	670	530	– 810
DiaSorin LIAISON (2)	ng/mL	482	300	– 664
DiaSorin LIAISON XL (2)	ng/mL	586	466	– 706
IDS-iSYS IGF-1 (2)	ng/mL	425	357	– 493
Mediagnost IGF-I ELISA (IGFBP-blocked) (2)	ng/mL	557	391	– 723
Mediagnost IGF-I RIA (IGFBP-blocked) (2)	ng/mL	557	391	– 723
Roche Elecsys/E170/cobas e Systems	ng/mL	314	267	– 360
Siemens IMMULITE 2000 / 2000 XPI (Kit Lot 482-485, >488)	ng/mL	580	439	– 720
Siemens IMMULITE 2000 / 2000 XPI (Re-Standardized)	ng/mL	459	346	– 572
SNIBE Maglumi Series (2)	ng/mL	▲		
<b>OSTEOCALCINĂ</b>				
Cisbio Bioassays OSTEO-RIACT (2)	ng/mL	180	144	– 216
DiaSorin LIAISON (2)	ng/mL	▲		
DiaSorin LIAISON XL (2)	ng/mL	▲		
IDS-iSYS N-Mid Osteocalcin (2)	ng/mL	161	136	– 187
Roche Elecsys/E170/Cobas e Systems	ng/mL	215	147	– 282
Siemens IMMULITE 2000 / 2000 XPI	ng/mL		67.9	– >100
Snibe MAGLUMI (2)	ng/mL	§		
<b>PEPTIDĂ C</b>				
Abbott Alinity i Analyzer	ng/mL	12.7	9.51	– 15.8
Abbott ARCHITECT iSystems	ng/mL	12.7	9.59	– 15.8
Beckman Coulter Access / 2 / 2i (2)	ng/mL	12.6	10.1	– 15.1
Beckman Coulter C-Peptide IRMA (2)	ng/mL	13.8	10.4	– 17.3
Beckman Coulter UniCel DxI Series (2)	ng/mL	12.6	10.1	– 15.2
Calbiotech ELISA (2)	ng/mL	5.25	3.50	– 7.00
Cisbio Bioassays IRMA-CPEP (2)	ng/mL	12.9	10.2	– 15.6
DIAsource ImmunoAssays C-PEP II RIA-CT (2)	ng/mL	§		
Mindray CL-Series/Wiener lab. CLIA Series (2)	ng/mL	14.2	10.4	– 18.0
Monobind AccuBind ELISA (2)	ng/mL	9.15	6.13	– >10.0

**METODĂ (1)**

	Unități	Nivel 3 - 64953			
		Medie		Interval	
<b>PEPTIDĂ C (continuare)</b>					
Monobind AccuLite CLIA (2)	ng/mL	8.01	5.37	–	>10.0
Ortho Vitros Microwell Series (C-pep) (2)	ng/mL	13.2	9.69	–	16.7
Roche Elecsys/E170/Cobas e Systems	ng/mL	12.7	9.68	–	15.8
Siemens ADVIA Centaur CP (CpS)	ng/mL	10.9	8.14	–	13.6
Siemens ADVIA Centaur CP (CpS) (Rgt lots 206 & calibrator lot 08 & higher)	ng/mL	10.6	7.87	–	13.3
Siemens ADVIA Centaur Systems (CpS)	ng/mL	11.2	8.83	–	13.6
Siemens ADVIA Centaur Systems (CpS) (Rgt lots 206 & calibrator lot 08 & higher)	ng/mL	12.1	9.71	–	14.5
Siemens Atellica CI 1900 Analyzer (CpS) (IM)	ng/mL	11.1	9.26	–	12.9
Siemens Atellica CI Analyzer (CpS) (IM) (Rgt 207 & calibrator 09 & higher)	ng/mL	12.2	10.4	–	14.0
Siemens Atellica Solution (CpS) (IM)	ng/mL	11.1	9.26	–	12.9
Siemens Atellica Solution (CpS) (IM) (Rgt 207 & calibrator 09 & higher)	ng/mL	12.2	10.4	–	14.0
Siemens IMMULITE / IMMULITE 1000	ng/mL	17.6	13.4	–	>20.0
Siemens IMMULITE 2000 / 2000 XPI	ng/mL	14.9	11.2	–	18.7
Tosoh ST AIA-PACK C-Peptide II (2)	ng/mL	13.3	9.30	–	17.3
<b>PTH (INTACT)</b>					
Abbott Alinity i Analyzer Routine Mode	pg/mL	993	671	–	1316
Abbott Alinity i Analyzer STAT Mode	pg/mL	810	489	–	1132
Abbott ARCHITECT iSystems Routine Mode	pg/mL	715	415	–	1015
Abbott ARCHITECT iSystems STAT Mode	pg/mL	782	583	–	980
Beckman Coulter Access / 2 / 2i Intraoperative Mode (2)	pg/mL	543	434	–	652
Beckman Coulter Access / 2 / 2i Routine Mode (2)	pg/mL	653	522	–	783
Beckman Coulter PTH IRMA (2)	pg/mL	668	468	–	869
Beckman Coulter UniCel DxI Intraoperative Mode (2)	pg/mL	564	451	–	677
Beckman Coulter UniCel DxI Routine Mode (2)	pg/mL	661	529	–	793
Biomerica ELISA (2)	pg/mL	701	604	–	798
bioMérieux VIDAS	pg/mL	405	298	–	512
Calbiotech Bio-intact PTH ELISA (2)	pg/mL	§			
Calbiotech intact PTH ELISA (2)	pg/mL	848	564	–	>1000
DiaSorin LIAISON 1-84 PTH (310630) (2)	pg/mL	354	258	–	449
DiaSorin LIAISON N-TACT PTH II (317910) (2)	pg/mL	1154	808	–	1501
DiaSorin LIAISON XL 1-84 PTH (310630) (2)	pg/mL	385	281	–	489
DiaSorin LIAISON XL N-TACT PTH II (317910) (2)	pg/mL	1196	837	–	1554
DIAsource ImmunoAssays hPTH-120 min-IRMA (2)	pg/mL	346	277	–	416
IDS-iSYS Intact PTH (2)	pg/mL	581	500	–	662
Mindray CL-Series/Wiener lab. CLIA Series (2)	pg/mL	473	345	–	601
Ortho Vitros MicroWell Series (iPTH) (2)	pg/mL	468	281	–	655
Roche cobas e 801	pg/mL	464	356	–	572
Roche Elecsys / Cobas e411	pg/mL	442	338	–	547
Roche Elecsys / Cobas e411 (STAT)	pg/mL	401	308	–	494
Roche Elecsys / E170 / cobas e Systems	pg/mL	387	289	–	485
Roche Elecsys / E170 / cobas e Systems (STAT)	pg/mL	304	227	–	382
Siemens ADVIA Centaur CP (PTH) (REF 10699154/10699155)	pg/mL	821	673	–	968
Siemens ADVIA Centaur Systems (PTH) (REF 10699154/10699155)	pg/mL	797	638	–	955
<b>Siemens Atellica CI 1900 Analyzer (PTH) (IM) →</b>	<b>pg/mL</b>	<b>813</b>	<b>634</b>	<b>–</b>	<b>991</b>
Siemens Atellica Solution (PTH) (IM)	pg/mL	813	634	–	991
Siemens IMMULITE / IMMULITE 1000	pg/mL	585	435	–	734
Siemens IMMULITE / IMMULITE 1000 (Turbo)	pg/mL	469	273	–	666



## METODĂ (1)

	Unități	Nivel 3 - 64953		
		Medie	Interval	
<b>PTH (INTACT) (continuare)</b>				
Siemens IMMULITE 2000 / 2000 XPi	pg/mL	696	515	– 878
SNIBE Maglumi Series (2)	pg/mL	▲		
Tosoh ST AIA-PACK Intact PTH (2)	pg/mL	390	271	– 509

## NOTE DE SUBSOL

(1) Este posibil să nu se aplice toate notele de subsol fișei dvs. de date personalizată selectată.

(2) Valorile atribuite au fost determinate utilizând protocolul furnizat de producătorul reactivului și/sau al instrumentului și este posibil să nu corespundă intervalelor de +/-3 SD.

▲ Datele nu sunt disponibile momentan. Vă rugăm să cereți mai multe informații.

§ Datele necesare stabilirii valorilor medii și intervalelor acceptabile pentru acest test nu au fost obținute din cauza participării limitate la activitate. În cazul în care instituția dvs. este interesată să participe la Programul de Atribuire a Valorilor (Value Assignment Program) pentru acest test, vă rugăm să contactați biroul dvs. local Bio-Rad.

■ Nu este cazul