

Declaration of Conformity

We, Katronic AG & Co. KG, declare under our sole responsibility that the product listed below to which this declaration relates are in conformity with the following directives and standards:

- EMC Directive 2014/30/EU for Electromagnetic Compatibility
- Low Voltage Directive 2014/35/EU for Electrical Safety
- Directive 2011/65/EU on the Restriction of Hazardous Substances (RoHS)
- BS 8452:2010 - Use of clamp-on ultrasonic flow-metering techniques for fluid applications
- ASME MFC-5.1:2011 - Measurement of Liquid Flow in Closed Conduits Using Transit-Time Ultrasonic Flowmeters

Name of Products	Description
KATflow 100, 150, 200, 210 and 230	Ultrasonic flowmeter with associated Katronic transducers

Class	Standard	Description
EMC Directive	BS EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use – EMC requirements
Immunity	BS EN 61326-1:2013 BS EN 61000-4-2:2009 BS EN 61000-4-3+A2:2010 BS EN 61000-4-4:2012 BS EN 61000-4-5:2014+A1:2017 BS EN 61000-4-6:2014 BS EN 61000-4-11:2014 +A1:2017	Electrical equipment for continuous unattended use Electrostatic discharge RF field Electric fast transient/burst Surge RF conducted AC mains voltage dips and interruption
Emission	BS EN 61326-1:2013	Electrical equipment Class B
Low Voltage Directive	BS EN 61010-1:2010 +A1 2019	Safety requirements for electrical equipment for measurement, control and laboratory use

Wernigerode, 11 May 2020
For and behalf on Katronic AG & Co. KG



Karsten Frahn
CEO, Katronic AG & Co. KG