



201819000873

TEST REPORT



中国认可
国际互认
检测
TESTING
CNAS L7673

Applicant : CARAVAN EXPRESS SRL
Address : Republic of MOLDOVA, Chisinau city, Decebal Boulevard 76, postal code MD-2038

The following merchandise was (were) submitted and identified by the client as:

Name of Sample : Bactericidal air recirculator
Test Type : Commission
Sample Quantity : 1 PC
Model : RBI 100-1200
Production date or batch No. : 2024-6-3
Brand : ALICOM
Manufacturer: CARAVAN EXPRESS SRL
Sample Received : 2025/08/11
Test Period : 2025/08/19 – 2025/08/28
Test Items : Please refer to next page(s).
Test Method : Please refer to next page(s).
Test Result : Please refer to next page(s).
Sample Description : Machine
Note: /

Edited by: 李 晓 岩

Approved by: 王 磊

Checked by: 王 磊

Official Seal: 检验检测专用章

Report No.: JKK25080020B(E)

Report Date: 2025/09/03

Page No.: 2/4

TEST RESULTS:

Table 1 Test data of virus removal								
Virus strain	Test time (min)	Test No.	Control group		Test group		Removal rate K_t (%)	Average (%)
			0 min V_0 (TCID ₅₀ /m ³)	60 min V_t (TCID ₅₀ /m ³)	0 min V_1 (TCID ₅₀ /m ³)	60 min V_2 (TCID ₅₀ /m ³)		
H1N1	60	1	1.69×10^7	5.71×10^6	2.91×10^7	2.80×10^5	97.14	97.60
		2	1.14×10^7	3.52×10^6	2.22×10^7	1.19×10^5	98.26	
		3	1.03×10^7	2.91×10^6	2.80×10^7	2.06×10^5	97.39	

*****TO BE CONTINUED *****

检测

Table 2 Inspection instructions

1. Test method	GB/T 18801-2022 Air cleaner (Annex H)
2. Test item	Virus strain: H1N1 (VR-1469) Cell: MDCK cell
3. Test equipment	Test chamber (30 m ³), Aerosol generator, Liquid impingement sampler etc.
4. Test condition	1) Environment temperature: (20~25) °C 2) Environment humidity: (50~70)%RH
5. Operation conditions of the machine	Set the switch to position "Maximum Wind Speed".
6. Test procedure	1) To test the virus titer of the returned sample, follow the following steps: <ol style="list-style-type: none"> 10 times dilute the recovered liquid; Dilute solution was added to a 96-well cell culture plate containing cells grown into monolayers, and a negative control group was set up and an equal amount of culture solution was added; After incubation at 35 °C and 5%CO₂ for 120 min, the supernatant was discarded, and the maintenance culture medium containing double antibody (Final concentration:100U/mL Penicillin and 100 µg/mL Streptomycin) was added for 3~5 days, and the cell growth was observed every day; When the cells inoculated with the virus appear to become round or shrink, the occurrence of cytopathic changes was recorded; Viral titers were calculated by the Reed-Muench method and expressed as TCID₅₀; 2) Calculate the virus titer and the removal rate, and this experiment repeated 3 times.
7. Computational formula	<p>Natural decay rate $N_t(\%) = \frac{V_0 - V_t}{V_0} \times 100$</p> <p>$N_t$ - Natural decay rate(%) Keep two decimal places</p> <p>V_0 - the original virus titer of control group(TCID₅₀/m³)</p> <p>V_t - the final virus titer of control group(TCID₅₀/m³)</p> <p>Removal rate $K_t(\%) = \frac{V_1 \times (1 - N_t) - V_2}{V_1 \times (1 - N_t)} \times 100$</p> <p>$K_t$ - Removal rate(%) Keep two decimal places</p> <p>V_1 - the original virus titer of test group (TCID₅₀/m³)</p> <p>V_2 - the final virus titer of test group (TCID₅₀/m³)</p>

***** END OF REPORT *****

Statement

1. This report is issued by The CAS Testing Technical Services (GuangZhou) Co.,Ltd. (hereinafter referred to as "Our Company").
2. This report is invalid if not affixed with authorized stamp of test and paging seal.
3. This report is invalid without signature of editor, verifier and approver.
4. This report is invalid if being supplemented, deleted or altered.
5. Without written permission of our Company, this report can not be reproduced in part (except in whole).
6. The result(s) shown in this report refer only to the sample(s) tested.
7. Objections to this report must be submitted to our Company within 15 days. Otherwise, it will automatically deem to have accepted this report.
8. The Client shall be responsible for the accuracy, authenticity and completeness of the samples and information submitted for inspection, and the disputes arising therefrom shall be borne by the Client.
9. As any reports is issued as a result of this application for testing services, our Company will strictly keep confidentiality to the Clients. Except where disclosure is required on the basis of laws, regulations, judgments, and rulings (including in accordance with summons, court, or government proceedings).
10. The result(s) or conclusion(s) shown in this report about the description of the characteristics, composition, properties or quality are based on the specific time, methods and applicable criteria. Using different methods and criteria or under different environmental conditions for testing may come to different conclusions.
11. Since our Company's causes lead to modify the contents of this report, our Company shall reissue this report and bear the modification cost. The Client shall return the original report. Since the Client's causes lead to modify the contents of this report, the Client need to submit an application form for the change of report to our Company. The Client shall bear the modification cost and return the original report if our Company approves to reissue this report.