

SELF DECLARATION OF CONFORMITY

products: We declare under our sole responsibility in accordance with MHRA Registration Number IVD 000100 that the following CE marked

_		
EDMA code(s) EDI	EDMA description	TCS product code and description
14.50.01.90 Oth	Other Controls/Standards/Calibrators,	Selectrol - All MM codes
Mic	Vlicrobiology	

diagnostic medical devices. Regulations 2002 (SI 2002 No.618) and The Medical Devices (Amendment) Regulations 2003 (SI 2003 No.1697) for in-vitro conform to the relevant provisions of the In-vitro Diagnostic Medical Devices Directive 98/79/EC and The Medical Devices

as certified by BSi, certificate number FS 28907. Directive 98/79/EC and continued maintenance of an approved Quality Management System meeting the requirements of ISO 9001, This declaration is made on the basis of meeting the requirements of Annexes I and III of the In-Vitro Diagnostic Medical Devices

Name: Position: Signed by: Signed by: Name: Position: Managing Director **Regulatory Affairs Manager** Sue Brown Lynda Preston bon Date: Date: 30.04.2016 30.04.2016

lss 2/30.04.16

ANNEX TO SELF DECLARATION OF CONFORMITY - SELECTROL (Dated 30.04.2016)

List of product codes and descriptions – August 2018

Product	Description
MM01-10	Enterobacter cloacae 10 Disc
MM02-10	Escherichia coli 10 Disc
MM02-25	Escherichia coli 25 Disc
MM04-10	Klebsiella pneumoniae 10 Disc
MM04-25	Klebsiella pneumoniae 25 Disc
MM05-5	Neisseria gonorrhoeae 5 Disc
MM07-10	Shigella sonnei 10 Disc
MM08-10	Legionella pneumophila serogroup 1 10d
MM09-10	Proteus vulgaris 10 Disc
MM09-25	Proteus vulgaris 25 Disc
MM10-10	Pseudomonas aeruginosa 10 Disc
MM10-10	Pseudomonas aeruginosa 10 Disc
MM10-25	Pseudomonas aeruginosa 25 Disc
MM100-10	Haemophilus influenzae 10 Disc
MM11-10	Salmonella Typhimurium 10 Disc
MM11-25	Salmonella Typhimurium 25 Disc
MM12-10	Serratia marcescens 10 Disc
MM13-10	Staphylococcus aureus 10 Disc
MM13-10	Staphylococcus aureus 10 Disc
MM13-25	Staphylococcus aureus 25 Disc
MM14-10	Staphylococcus aureus 10 Disc
MM14-25	Staphylococcus aureus 25 Disc
MM15-10	Staphylococcus epidermidis 10 Disc
MM15-10	Staphylococcus epidermidis 10 Disc
MM15-25	Staphylococcus epidermidis 25 Disc
MM16-10	Streptococcus agalactiae 10 Disc
MM17-10	Enterococcus faecalis 10 Disc
MM17-25	Enterococcus faecalis 25 Disc
MM18-10	Enterococcus faecalis 10 Disc
MM18-10	Enterococcus faecalis 10 Disc
MM18-25	Enterococcus faecalis 25 Disc
MM19-10	Streptococcus pneumoniae 10 Disc
MM20-10	Streptococcus pyogenes 10 Disc
MM20-25	Streptococcus pyogenes 25 Disc
MM21-10	Bacillus cereus 10 Disc
MM24-10	Escherichia coli 10 Disc
MM24-25	Escherichia coli 25 Disc
MM26-10	Enterobacter aerogenes 10 Disc
MM26-25	Enterobacter aerogenes 25 Disc

Product	Description
MM27-10	Citrobacter freundii 10 Disc
MM28-10	Candida albicans 10 Disc
MM28-25	Candida albicans 25 Disc
MM29-10	Bacillus subtilis 10 Disc
MM29-25	Bacillus subtilis 25 Disc
MM30-10	Staphylococcus aureus 10 Disc
MM30-25	Staphylococcus aureus 25 Disc
MM31-10	Clostridium sporogenes 10 Disc
MM33-10	Escherichia coli 10 Disc
MM33-25	Escherichia coli 25 Disc
MM34-10	Escherichia coli (mcr-1) 10 Disc
MM35-10	Enterococcus hirae 10 Disc
MM36-10	Campylobacter jejuni 10 Disc
MM37-10	Haemophilus influenzae 10 Disc
MM38-10	Escherichia coli 10 Disc
MM38-25	Escherichia coli 25 Disc
MM40-25	Pseudomonas aeruginosa 25 Disc
MM41-10	Pseudomonas aeruginosa 10 Disc
MM42-10	Candida albicans 10 Disc
MM42-25	Candida albicans 25 Disc
MM43-10	Proteus mirabilis 10 Disc
MM44-10	Bacteroides fragilis 10 Disc
MM44-10	Bacteroides fragilis 10 Disc
MM45-10	Clostridium perfringens 10 Disc
MM45-10	Clostridium perfringens 10 Disc
MM46-10	Staphylococcus aureus 10 Disc
MM46-10	Staphylococcus aureus 10 Disc
MM46-25	Staphylococcus aureus 25 Disc
MM47-10	Listeria monocytogenes 10 Disc
MM48-10	Listeria monocytogenes 10 Disc
MM48-25	Listeria monocytogenes 25 Disc
MM50-10	Saccharomyces cerevisiae 10 Disc
MM51-10	Enterobacter cloacae 10 Disc
MM52-10	Enterococcus faecalis 10 Disc
MM53-10	Candida krusei 10 Disc
MM54-10	Candida parapsilosis 10 Disc
MM55-10	Klebsiella pneumoniae 10 disc
MM56-10	Klebsiella pneumoniae 10 Disc
MM57-10	Escherichia coli 10 Disc
MM58-10	Klebsiella pneumoniae 10 Disc
MM59-10	Klebsiella pneumoniae 10 Disc
MM60-10	Streptococcus dysgalactiae 10 Disc
MM62-10	Salmonella Enteritidis 10 Disc

Product	Description
MM63-10	Escherichia coli 10 Disc
MM64-10	Staphylococcus aureus 10 Disc
MM65-10	Pseudomonas aeruginosa 10 Disc
MM68-10	Proteus mirabilis 10 Disc
MM70-10	Burkholderia cepacia 10 Disc
MM73-10	Saccharomyces cerevisiae 10 Disc
MM73-25	Saccharomyces cerevisiae 25 Disc
MM75-10	Escherichia coli 10 Disc
MM76-10	Lactobacillus brevis 10 Disc
MM77-10	Listeria monocytogenes 10 Disc
MM80-10	Yersinia enterocolitica 10 Disc
MM80-25	Yersinia enterocolitica 25 Disc
MM81-10	Haemophilus influenzae 10 Disc
MM82-10	Campylobacter jejuni 10 Disc
MM83-10	Klebsiella pneumoniae 10 Disc
MM84-10	Salmonella Nottingham 10 Disc
MM85-10	Staphylococcus aureus 10 Disc
MM86-10	Bacillus cereus 10 Disc
MM86-25	Bacillus cereus 25 Disc
MM87-10	Listeria monocytogenes 10 Disc
MM88-10	Klebsiella aerogenes 10 Dis
MM89-10	Salmonella Poona 10 Disc
MM89-25	Salmonella Poona 25 Disc
MM91-10	Staphylococcus aureus 10 Disc
MM91-10	Staphylococcus aureus 10 Disc
MM92-10	Listeria innocua 10 Disc
MM93-10	Escherichia coli 0157 non-toxigenic 10d
MM93-25	Escherichia coli 0157 non-toxigenic 25D
MM94-10	Aspergillus brasiliensis 10 Disc
MM95-10	Streptococcus pneumoniae 10 Disc
MM97-10	Rhodococcus equi 10 Disc
MM99-10	Legionella anisa 10 Disc
MM99-10	Legionella anisa 10 Disc

Signed by: Signed by:

Date: 21.08.2018

Name: Position: Sue Brown Regulatory Affairs Manager

Signed by:_ $\mathbb{H}(\mathcal{O})$

Name: Position:

Lynda Preston Managing Director

Date: 21 - 08 - 2018

tan Santa Sant Santa Sant

MALSI DUT

Alban Barat S

/	-		_			
(bioscie	CC	-		5	
	iences	è) n		1	1
	-	Č	0	-		
				-	/	/

SELF DECLARATION OF CONFORMITY

products: We declare under our sole responsibility in accordance with MHRA account number 0000009546 that the following CE marked

GMDN Term	GMDN Code	TCS product code and description
General microbial isolate identification control IVD	63319	Selectrol – All MM codes

diagnostic medical devices. conform to the relevant provisions of the In-vitro Diagnostic Medical Devices Directive 98/79/EC and The Medical Devices Regulations 2002 (SI 2002 No.618) and The Medical Devices (Amendment) Regulations 2003 (SI 2003 No.1697) for in-vitro

as certified by BSi, certificate number FS 28907. Directive 98/79/EC and continued maintenance of an approved Quality Management System meeting the requirements of ISO 9001, This declaration is made on the basis of meeting the requirements of Annexes I and III of the In-Vitro Diagnostic Medical Devices

Signed by: Sue Bow	Date:_	Date: <u>September 2020</u>
Vame: Sue Brown Position: Quality Assurance & Regulatory Affairs Manager	inager	
Signed by: While (Calon	Date:_	September 2020
Vame: ^D osition: Managing Director		
<u>Manufacturer</u> TCS Biosciences Ltd		EC Authorised Representative TCS Biosciences Europe B.V.
Botolph Claydon Buckingham MK18 2LR UK		Provincialeweg 6 9864 PD Kornhorn The Netherlands







Certificate of Registration

QUALITY MANAGEMENT SYSTEM - ISO 9001:2015

This is to certify that:

TCS Biosciences Limited Botolph Claydon Buckingham MK18 2LR United Kingdom

Holds Certificate Number:

FS 28907

and operates a Quality Management System which complies with the requirements of ISO 9001:2015 for the following scope:

The procurement, manufacture, design, development and sale of a range of diagnostic products for clinical, pharmaceutical, food and environmental laboratory testing.

For and on behalf of BSI:

Original Registration Date: 1994-08-11 Latest Revision Date: 2022-01-14



Matt Page, Managing Director Assurance - UK & Ireland

Effective Date: 2022-01-27 Expiry Date: 2025-01-26

Page: 1 of 1

...making excellence a habit."

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract. An electronic certificate can be authenticated <u>online</u>. Printed copies can be validated at www.bsigroup.com/ClientDirectory

Information and Contact: BSI, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP. Tel: + 44 345 080 9000 BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK. A Member of the BSI Group of Companies.







Certificate of Registration

ENVIRONMENTAL MANAGEMENT SYSTEM - ISO 14001:2015

This is to certify that:

TCS Biosciences Ltd Botolph Claydon Buckingham MK18 2LR United Kingdom

Holds Certificate Number:

EMS 590359

and operates an Environmental Management System which complies with the requirements of ISO 14001:2015 for the following scope:

The procurement, manufacture, design, development and sale of a range of diagnostic products for clinical, pharmaceutical, food and environmental laboratory testing.

For and on behalf of BSI:

Original Registration Date: 2013-06-10 Latest Revision Date: 2022-01-14



Matt Page, Managing Director Assurance - UK & Ireland

Effective Date: 2022-01-27 Expiry Date: 2025-01-26


Page: 1 of 1

...making excellence a habit."

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract. An electronic certificate can be authenticated <u>online</u>. Printed copies can be validated at www.bsigroup.com/ClientDirectory

Information and Contact: BSI, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP. Tel: + 44 345 080 9000 BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK. A Member of the BSI Group of Companies.

accuracy and quality as a science









Selectrol®: Manufactured under licence from Public Health England Culture Collections

SELECTROL® - FREEZE-DRIED ORGANISMS IN A DISC

Quality control of microbial characterisation tests, culture media and antimicrobial susceptibility determinations is best accomplished by the use of microorganisms with well-documented and stable phenotypic and genotypic characteristics.

Bacterial and fungal strains have been selected and recommended by expert bodies, such as EUCAST, CLSI and the European Pharmacopoeia, on the basis of their suitability for monitoring test performance and ensuring the validity of results for testing used in clinical, food, pharmaceutical, water and veterinary laboratories.

Products derived from the cultures in the collections should be manufactured using the minimum number of sub-cultures, to minimise the possibility of alterations to the phenotype due to mutations. See also page 14.

Selectrol strains are manufactured exclusively from Public Health England Culture Collections (NCTC[®] and NCPF[®]) and are first generation subcultures, unlike many products on the market which are 2nd, 3rd or 4th generation subcultures. They are preserved by long-term storage as freeze-dried cells in order to minimise any alterations to the phenotype caused by mutations.

Passages

A Selectrol[®] disc is a first generation subculture from a **master culture** sourced from Public Health England Culture Collections, and is designed to be used to obtain **working stock** cultures for use in testing. It is generally accepted that no more than a total of five passages should be made from the **master culture**, in order to avoid genetic drift and mutant selection. Therefore, no more than four passages (fresh cultures) from the **working stock** should be made.

Shelf life

For most strains, Selectrol[®] discs are guaranteed to contain at least 10⁶ organisms at the time of purchase; this number is sufficient to ensure that when the discs are used and stored as directed there will be viable organisms cultivable up to the stated end of the shelf life, which is usually 9 months from the time the vial is first opened.

Quality Control

Selectrol[®] batches are tested in our UKAS accredited testing laboratory number 2496. A test report for each batch of Selectrol[®] can be accessed via our website. The reporting of Selectrol[®] test results via the website comes under our UKAS accreditation.

Selectrol[®] cultures are rigorously tested to confirm identity, to confirm the possession of essential phenotypic characteristics and to exclude contamination with other organisms. Photographic evidence of the test results is retained for each batch, along with retained appropriately stored samples.



Glossary

AMRHAI: Antimicrobial Resistance and Healthcare Associated Infections reference unit

ATCC®: American Type Culture Collection. ATCC[®] strains are listed for reference only. ATCC[®] is a registered trademark of the American Type Culture Collection.

BSAC: British Society for Antimicrobial Chemotherapy - Now superseded by EUCAST

CLSI: Clinical Laboratory Standards Institute. (USA)

CPE: Carbapenemase Producing Enterobacteriaceae

CRE: Carbapenem Resistant Enterobacteriaceae

Culture collection: Cultures of fully characterised organisms maintained in such a way as to minimise sub-culturing. See page 14.

ESBL: Extended Spectrum Beta-Lactamase-producing organism.

EUCAST: European Committee on Antimicrobial Susceptibility Testing.

First generation derivative: A single passage from a master culture, for example a Selectrol® disc.

Master culture: Culture derived from a reference culture vial.

NCPF[®]: National Collection of Pathogenic Fungi. NCPF[®] is a registered trademark of Public Health England.

NCTC[®]: National Collection of Type Cultures. NCTC[®] is a registered trademark of Public Health England.

Passage: An equivalent term for a subculture.

PHE: Public Health England.

Reference cultures: Quality control strains selected on the basis of their phenotypic biochemical and antimicrobial susceptibility characteristics to be used as controls in microbiological testing. These are obtained as freeze-dried vials from culture collections.

Stock culture: Cultures derived from a Selectrol® disc, which can be stored for up to a week, usually on agar slants.

Working cultures: Stock cultures further sub-cultured to provide 18-24 hour growth for use in testing.

WDCM: World Data Centre for Microorganisms

WFCC: World Federation for Culture Collections



SIGNIFICANT PROPERTIES AND USES OF SELECTROL® ORGANISMS

Aspergillus brasiliensis (formerly Aspergillus niger):

MM94 – NCPF[®] 2275 / ATCC[®] 16404 / WDCM 00053 – used in pharmaceutical industry for testing media and preservatives. Colonies are initially white or yellowish and on the reverse greyish or greenish-yellow. Sporing heads on the colony surface are initially pale, becoming dark brown to black. Sporulation may be inhibited in sealed plates.

Bacillus cereus:

MM21 – NCTC[®] 10320 / ATCC[®] 9634 / WDCM 00001 (recently renamed *Bacillus toyonensis*) – ISO 11133 recommended media and ID test control organism.

MM86 - NCTC[®] 7464 / ATCC[®] 10876 - PHE recommended media and ID test control organism.

Bacillus subtilis (Bacillus subtilis subsp. spizizenii):

MM29 - NCTC[®] 10400 / ATCC[®] 6633 / WDCM 00003 - used in antibiotic assays (fully sensitive), PHE recommended media and ID test control organism.



Bacteroides fragilis:

MM44 – NCTC[®] 9343 / ATCC[®] 25285 – type strain, PHE recommended strain for media and sensitivity test control.

Campylobacter jejuni (Campylobacter jejuni subsp. jejuni):

MM82 - NCTC[®] 11322 / ATCC[®] 29428 / WDCM 00156 - PHE recommended strain for media control.

MM36 - NCTC® 11351 / ATCC® 33560 - EUCAST recommended strain for susceptibility testing.

Candida albicans:

MM28 - NCPF[®] 3255 / ATCC[®] 2091 / WDCM 00055 - sensitivity control / industrial use.

MM42 – NCPF[®] 3179 / ATCC[®] 10231 / WDCM 00054 – pharmaceutical / media testing / PHE recommended strain for media control.

CRE = 'Carbapenem Resistant Enterobacteriaceae' / CPE = 'Carbapenemase Producing Enterobacteriaceae'

There are 5 carbapenemases which are currently a significant problem in the UK – KPC, OXA-48, IMP, NDM and VIM – and PHE recommend that all clinically-significant Gram-negative bacteria should be routinely screened for carbapenemase production, using a recommended carbapenem² such as ertapenem or meropenem. Resistant isolates may be investigated further to determine which resistance mechanism is involved using the Modified Hodge Test, MALDI-TOF, PCR or a reference laboratory.

MM55 Klebsiella pneumoniae - NCTC® 13440 - produces a Class B VIM-1 Carbapenemase.

MM56 Klebsiella pneumoniae - NCTC® 13443 - produces a Class B NDM-1 Carbapenemase.

MM58 *Klebsiella pneumoniae* – NCTC[®] 13438 – produces a Class A KPC-3 Carbapenemase.

MM59 Klebsiella pneumoniae - NCTC® 13442 - produces a Class D OXA-48 Carbapenemase.

MM57 Escherichia coli - NCTC® 13476 - produces a Class B IMP Carbapenemase.

MM33 Escherichia coli - NCTC® 10418 / ATCC® 10536 - recommended by PHE as a negative control for CRE testing.



Citrobacter freundii:

MM27 - NCTC® 9750 / ATCC® 8090 - type strain.

Clostridium perfringens:

MM45 – NCTC[®] 8237 / ATCC[®] 13124 / WDCM 00007 – type strain. PHE recommended strain for food testing (Tryptose Sulphite Cycloserine agar – lactose and gelatin positive) and sensitivity test control. *Clostridium perfringens* is listed in Schedule 5 of the Anti-terrorism, Crime and Security Act 2001, and should be securely stored in accordance with the guidelines of the Act. However, MM45 is a type A strain, which <u>does not</u> produce the lethal epsilon toxin of potential interest to bioterrorists.

Clostridium sporogenes:

MM31 – NCTC[®] 532 / ATCC[®] 19404 / WDCM 00008 – used for media control. PHE recommended strain for media QC (lactose gelatin medium for ID of *C. perfringens* lactose negative and gelatin positive).

Enterobacter aerogenes:

MM26 - NCTC® 10006 / ATCC® 13048 / WDCM 00175 - type strain; used in water, paint and adhesive testing.

Enterobacter cloacae:

MM01 - NCTC® 13380 / ATCC® 23355 / WDCM 00082 - disinfectant control, media testing.

MM51- NCTC[®] 13406 - PHE recommended strain for QC of AmpC (de-repressed) detection.

Enterococcus faecalis:

MM52 – NCTC[®] 13379 / ATCC[®] 51299 / WDCM 00085 – is vancomycin resistant (low-level VanB mediated) and also shows highlevel resistance to aminoglycosides. It is used to confirm methodologies used to detect these resistances are working correctly. Lancefield group D.

MM17 – NCTC[®] 775 / ATCC[®] 19433 / WDCM 00009 – used in water industry and QC. PHE recommended strain for media control. Fully sensitive. Lancefield group D.

MM18 – NCTC[®] 12697 / ATCC[®] 29212 / WDCM 00087 – is fully sensitive to vancomycin and gentamicin. PHE recommended positive control strain for aesculin test. CLSI, EUCAST recommended media control for sulpha / trimethoprim testing and general susceptibility testing control. Lancefield group D.





Enterococcus hirae:

MM35 – NCTC[®] 13383 / ATCC[®] 10541 / WDCM 00011 – disinfectant control. Used in microbiological assays. Colonies are alphahaemolytic on sheep blood agar.

Escherichia coli strains:

MM02 – NCTC[®] 12241 / ATCC[®] 25922 / WDCM 00013 – EUCAST, CLSI, PHE recommended control strain for susceptibility testing (fully sensitive). Exhibits 2 colony types – the most prevalent type is slightly irregular, smooth and translucent. The secondary type appears more opaque. It is preferable to maintain cultures on agar as passage in broth can result in a change in MIC levels.



MM57 - NCTC[®] 13476 - CRE testing control; produces a Class B IMP Carbapenemase.

MM33 – NCTC[®] 10418 / ATCC[®] 10536 – (PHE recommended alternative to NCTC 12241) fully sensitive control strain. PHE recommended positive control for indole test, ONPG test, negative control for oxidase test, PHE recommended negative control for CRE and ESBL testing.

MM24 – NCTC[®] 11954 / ATCC[®] 35218 – beta-lactamase positive strain. CLSI recommended strain for susceptibility testing ONLY for penicillin / beta-lactamase inhibitor combinations. Sensitive to amoxicillin / clavulanic acid.

MM75 – NCTC[®] 9001 / ATCC[®] 11775 / WDCM 00090 – used in water / chemical industry. PHE recommended strain for media QC.

MM93 – NCTC[®] 12900 / ATCC[®] 700728 / WDCM 00014 – O157 strain (non-toxigenic). PHE recommended strain for media QC.

MM63 - NCTC® 11560 - beta-lactamase positive strain.

MM38 – NCTC[®] 12923 / ATCC[®] 8739 / WDCM 00012 – used in pharmaceutical / water industry. Three colony types: A) Entire, glistening, smooth and translucent. B) Entire, glistening smooth and opaque. C) Irregular, rough and translucent. The rough colonies appear after 48 hours incubation.

MM34 – NCTC[®] 13846 – Possesses the plasmid-mediated mcr-1 colistin resistance mechanism gene and is recommended by PHE and EUCAST as a control for tests to detect this increasingly prevalent resistance, in conjunction with NCTC® 12241 / ATCC® 25922 (Selectrol strain MM02) as a negative control.



Haemophilus influenzae strains:

MM81 - NCTC[®] 12699 / ATCC[®] 49247 – is a 'BLNAR' strain – (beta-lactamase non-producing ampicillin / amoxycillin resistant). These strains are important clinically because the susceptibility results obtained using conventional testing procedures maybe misleading in the case cephalosporins. PHE, CLSI recommended QC strain for susceptibility testing media.

MM98 – NCTC[®] 11931 – a fully sensitive strain. PHE recommended strain for porphyrin synthesis test, chocolate agar control.

MM100 – NCTC[®] 8468 / ATCC[®] 9334 / CCUG 23946 – another fully sensitive strain, which reportedly gives results which are easier to interpret when Mueller-Hinton medium is used in preference to Iso-Sensitest medium. MIC for amoxycillin is 0.5 mg/l.

MM37 - NCTC® 12975 / ATCC® 49766 - recommended by EUCAST.




Klebsiella strains:

MM04 *Klebsiella pneumoniae* – NCTC[®] 9633 / ATCC[®] 13883 / WDCM 00097 – type strain. Two colony types may be seen. The predominant type is entire and opaque. The secondary type is slightly smaller and translucent.

MM83 *Klebsiella pneumoniae* – NCTC[®] 13368 / ATCC[®] 700603 – ESBL-producing strain used as control for ESBL testing. There are two colony types.

MM55 Klebsiella pneumoniae – NCTC[®] 13440 – CRE testing control; produces a Class B VIM-1 Carbapenemase.



MM56 Klebsiella pneumoniae – NCTC[®] 13443 – CRE testing control; produces a Class B NDM-1 Carbapenemase.

MM58 Klebsiella pneumoniae – NCTC[®] 13438 – CRE testing control; produces a Class A KPC-3 Carbapenemase.

MM59 Klebsiella pneumoniae - NCTC® 13442 - CRE testing control; produces a Class D OXA-48 Carbapenemase.

MM88 *Klebsiella aerogenes (Raoultella planticola)* – NCTC[®] 9528 – used in water / pharmaceutical industry. PHE recommended negative control for Tryptone Bile X-Glucuronide agar and Yeast Extract agar.



Lactobacillus brevis:

MM76 - NCTC[®] 13386 / ATCC[®] 8287 - used in food industry.

Legionella pneumophila serogroup 1:

MM08 – NCTC[®] 11192 / ATCC[®] 33152 / WDCM 00107 – derived from strain isolated from first recognised outbreak of legionellosis in Philadelphia at the Legionnaires' Convention 1976

Listeria innocua:

MM92 - NCTC[®] 11288 / ATCC[®] 33090 / WDCM 00017 - type strain. Non-pathogenic.

Listeria monocytogenes:

MM87 – NCTC[®] 11994 / WDCM 00019 – type strain, PHE recommended positive control strain for Listeria detection in food. Serotype 4b, most common serovar isolated from human infections.

MM48 – NCTC[®] 7973 / ATCC[®] 35152 / WDCM 00109 – produces 2 phenotypes, one is beta-haemolytic and virulent, the other non-haemolytic and non-virulent. Serovar 1/2a.

MM77 – NCTC[®] 13372 / ATCC[®] 7644 – used in food microbiology Q.C. Colonies exhibit beta-haemolysis on sheep blood agar.

Neisseria gonorrhoeae:

 $MM96 - NCTC^{\circ}$ 12700 / $ATCC^{\circ}$ 49226 - has low-level, but clinically relevant, resistance to penicillin – MIC of penicillin is 0.5 mg/l. PHE recommended control for susceptibility testing – methodology assesses the ability of testing to detect resistance rather than sensitivity; this strain has low-level, but clinically relevant, resistance to penicillin – MIC of penicillin is 0.5 mg/l. Some variation in size and texture of colonies may be observed. Increased CO₂ is helpful in growth.

MM05 – NCTC[®] 8375 / ATCC[®] 19424 – is fully sensitive – MIC of penicillin is 0.06 mg/l. PHE recommended strain for media QC.

Proteus mirabilis:

MM43 – NCTC[®] 13376 / ATCC[®] 14153 – pharmaceutical / disinfectant / media control. MM68 – NCTC[®] 10975 – media control. PHE recommended control for motility test.



Proteus vulgaris:

MM09 – NCTC[®] 4175 / ATCC[®] 13315 – was the type strain, but is atypical and has been recognised as a separate species – *Proteus hauseri* – it is used for media control. Colonies are glistening with spreading edges.

Pseudomonas aeruginosa strains:

MM10 – NCTC[®] 12903 / ATCC[®] 27853 / WDCM 00025 – is fully sensitive to anti-pseudomonal antibiotics (EUCAST susceptibility test control). 2 colony types may be observed: A) predominantly flat, spreading edges and rough surface; B) small and compact. Produces both fluorescein and pyocyanin pigments.



MM65 - NCTC® 10662 / ATCC® 25668 / WDCM 00114 - is fully sensitive. PHE recommended control strain for media control

MM40 – NCTC[®] 12924 / ATCC[®] 9027 / WDCM 00026 – used in water industry / disinfectant testing. Colonies on agar plates are entire, glistening and mucoid with a grainy surface. This strain also produces both fluorescein and pyocyanin pigments.

MM41 – NCTC[®] 13359 / ATCC[®] 15442 – used in water industry / disinfectant testing. May produce up to 3 different colony types. Pyocyanin is not produced.

Rhodococcus equi:

MM97 - NCTC[®] 1621 / ATCC[®] 6939 / WDCM 00028 - type strain.

Saccharomyces cerevisiae:

MM73 – NCPF® 3178 – PHE recommended strain for food testing and enumeration of yeasts and moulds.

MM50 — NCTC® 10716 / WDCM 00058 - used for QC of culture media and for antifungal susceptibility testing.

Salmonella serotypes:

MM11 Salmonella Typhimurium – NCTC[®] 12023 / ATCC[®] 14028 / WDCM 00031 – (1,4,5,12: i: 1,2) Used for media/test QC. This is a common serotype from animals and from human infections.

The strains listed below are unusual serotypes, used to avoid any chance of confusion with strains commonly found in animals, food, etc, and are used to control media and detection methods in the food industry:

MM89 Salmonella Poona - NCTC[®] 4840 - (13,22: z: 1,6) PHE recommended control strain for food testing.

MM84 Salmonella Nottingham - NCTC® 7832 - (16: d: e,n,z15) PHE recommended control for water testing.

Serratia marcescens:

MM12 – NCTC[®] 13382 / ATCC[®] 8100 – used for disinfectant testing. PHE recommended negative control for indole test. Colonies are entire, glistening, smooth and translucent. Non-pigmented.



Staphylococcus aureus:

(A) Fully sensitive:

MM85 – NCTC[®] 6571 / ATCC[®] 9144 / WDCM 00035 – historically used for susceptibility testing ('Oxford staph'), but largely superseded by MM13 as it has unusually low MIC's and so is unrepresentative of normal range of Staph aureus strains. Sensitive to penicillin and cefoxitin / methicillin / oxacillin. PHE recommended coagulase, DNAse and catalase positive control.

MM13 – NCTC[®] 12981 / ATCC[®] 25923 / WDCM 00034 – used in susceptibility and media testing/QC. Fully sensitive to all antistaphylococcal antibiotics (including penicillin and methicillin / oxacillin). It is preferable to maintain cultures on agar as passage in broth can result in a change in MIC levels. Colonies are circular white to cream, convex to flat in elevation. After 48 hours incubation a few grey/translucent variants may be noted. Beta-haemolytic on sheep blood agar.

B) Penicillin resistant:

MM14 – NCTC[®] 12973 / ATCC[®] 29213 / WDCM 00131 – used for susceptibility testing, especially for automated methodology. EUCAST, CLSI strain. Sensitive to cefoxitin / methicillin / oxacillin. Penicillin resistant – weak beta-lactamase producer. Colonies are beta-haemolytic, and a golden-orange colour.

MM30 – NCTC[®] 7447 / ATCC[®] 6538P / WDCM 00033 – used for susceptibility testing/antibiotic assay, disinfectant testing. Cefoxitin / methicillin / oxacillin sensitive. Penicillin resistant. Colonies are weakly beta-haemolytic, coagulase positive and betalactamase negative.

(C) MRSA (cefoxitin / methicillin / oxacillin resistant):

MM91 – NCTC[®] 13373 / ATCC[®] 43300 / WDCM 00211 (MRSA) – Possesses mecA gene but is hetero-resistant, (so as few as one per thousand cells demonstrate the resistance) and consequently has low-level cefoxitin /oxacillin/methicillin resistance (4.0 mg/l MIC of oxacillin, 8.0 mg/l MIC of cefoxitin – methicillin sensitive strains have MIC of 0.12-0.5 for oxacillin and 1-4 for cefoxitin.); it is used to confirm testing procedures for methicillin resistance are working and provides a more stringent test than testing with an MRSA which shows homogeneous resistance and has a much higher MIC. This organism will have a zone of inhibition reduced in size compared to a fully cefoxitin / oxacillin / methicillin sensitive strain (such as MM13). CLSI recommended strain for MRSA testing. There are two colony types: 1) Beta-haemolytic with a slight yellow tint. 2) Non-haemolytic and white.

MM64 – NCTC[®] 12493 / WDCM 00212 (MRSA) – possesses mecA gene and shows homogeneous resistance with MIC of >64 for methicillin, which produces high-level cefoxitin / methicillin / oxacillin resistance. EUCAST recommended strain. Instances have been reported where loss of the mecA gene has occurred during storage.

D) Other:

MM46 – NCTC[®] 10788 / ATCC[®] 6538 / WDCM 00032 – used in pharmaceutical industry for testing disinfectants etc. Usually yellow pigmented colonies, or can produce a white colonial variant. Beta-haemolytic.





Staphylococcus epidermidis:

MM15 - NCTC® 13360 / ATCC® 12228 / WDCM 00036 - used for media control / antibiotic assay. Colonies are small and betahaemolytic.

Streptococcus agalactiae: (Beta-haemolytic Streptococcus group B)

MM16 - NCTC® 8181 / ATCC® 13813 - type strain, used for QC. PHE recommended negative control for aesculin test.

Streptococcus pneumoniae strains:

MM95 – NCTC[®] 12977 / ATCC[®] 49619 – has low-level, but clinically relevant, resistance to penicillin – this organism is used to assess detection of resistance rather than sensitivity. PHE recommended positive control for bile solubility test. CLSI, EUCAST recommended control strain for susceptibility testing. Serotype 19F.

MM19 – NCTC[®] 12695 / ATCC[®] 6303 – is fully sensitive. Colonies are mucoid and alpha-haemolytic. A few colonies may have an irregular edge. Serotype 3.



Streptococcus pyogenes:

MM20 – NCTC[®] 12696 / ATCC[®] 19615 – used for QC and media testing. Lancefield group A, beta-haemolytic. PHE recommended blood agar control.

Vibrio parahaemolyticus:

MM06 – NCTC[®] 10885 / WDCM 00185 – used for QC of media and ID testing. PHE recommended strain used mainly in the food industry.

Yersinia enterocolitica:

MM80 - NCTC[®] 12982 / ATCC[®] 9610 / WDCM 00038 - type strain, used for media control. Serotype O:8, which is a pathogenic serotype, commonest in USA.

References:

- 1 European Committee on Antimicrobial Susceptibility Testing (EUCAST). Routine and Extended Internal Quality Control for MIC Determination and Disc Diffusion. Version 7.0 01.01.2017.
- 2 UK Standards for Microbiology Investigations. Example Reference Strains for Microbiology Investigations Test Procedures: Bacteriology—Test Procedures | TP 1 | Issue No. 2 | 05.01.2015. Public Health England (PHE).
- 3 Performance Standards for Antimicrobial Disc Susceptibility Tests: Approved Standard—11th Edition. Clinical and Laboratory Standards Institute (CLSI).



How to use Selectrol®


Always warm the vial to ambient temperature before opening.

Be sure to use non-selective culture media to revive the organisms.

For the more fastidious organisms, such as anaerobes, it is generally better to use agar rather than broth for revival.



Place disc on suitable growth medium such as blood agar



Leave disc for a few minutes to liquefy, then spread plate and incubate to produce isolated colonies





Place disc in a small volume of a suitable broth medium such as brain-heart infusion



Allow disc a few minutes to dissolve, then spread aliquot onto a plate of suitable growth medium

Obtain a stock culture which can be used to prepare an inoculum for biochemical and antibiotic susceptibility tests



Out-of-specification results

Laboratories use Selectrol[®] for Quality Control of culture media, biochemical identification tests and antimicrobial susceptibility testing. When a laboratory test result, an MIC or biochemical reaction, is unexpected or out-of-specification, the test should first be repeated to confirm it; an out-of-specification result is an indication that the testing procedure should be reviewed; it is not, in the first instance, a sign of a problem with the control organism.

If incorrect results are obtained on retesting, the explanation could be:

- The test procedure was not followed correctly check standard operating procedures
- There is an instrumentation error check calibration, mechanical functioning, etc
- There is a problem with the consumables out of date, incorrect storage, etc
- The culture of the control organism has become contaminated

Technical Support

If no explanation for out-of-spec results can be found, but repeated tests still give unacceptable results, please contact TCS and / or your relevant reference laboratory or instrument manufacturer for advice. For example, contact AMRHAI at Colindale, London if MIC results are consistently outside the acceptable range. Please retain any remaining discs of organisms about which you have concerns so they can be returned to TCS and investigated alongside retained samples.





Preparing QC and Validation Spikes from Selectrol®

Preparing the spike

- Place a Selectrol[®] disc in Brain Heart Infusion (BHI) broth* or equivalent, and culture (typically for 18 hours) at the appropriate temperature for the organism (typically 37°C)

- Assume the count in the broth to be 10⁸ organisms per ml ------ (A)
- Mix and transfer 100 μl of (A) to 100 ml of saline or 1/4 strength Ringer's solution -- (B)
- Mix and transfer 100 µl of (B) to 10 ml of saline or ¼ strength Ringer's solution --- (C)
- Mix and transfer 100 µl of (C) to your homogenised food sample.

Verifying the inoculum

- Pipette 5 x 10 µl drops from (C) onto each of two agar plates for Miles and Misra counts.

Using the assumptions and dilutions above:

- (A) contains 10⁸ organisms per ml
- (B) contains 10⁵ organisms per ml
- (C) contains 10³ organisms per ml

If the Miles and Misra counts indicate that the required count was not achieved:

- If the count was too high by a factor of 10, reduce the volume transferred from (A) to (B) from 100 µl to 10 µl
- If the count was too low by a factor of 10, increase the volume transferred from (A) to (B) from 100 µl to 1 ml.

Keep a record of the correct dilutions for each organism type for future use. You will find that this method is very repeatable.

*Note: BHI broth will work for most of the Selectrol[®] organisms; however, for fastidious organisms an appropriate culture broth must be selected, e.g. Fastidious Anaerobe Broth for strictly anaerobic organisms.





Culture Collections

Cultures of microorganisms have been deposited and subsequently maintained in 589 collections in 68 countries, and many of the cultures are derived from the same original isolate; the history of each organism, its properties and names of the culture collections which hold it are detailed in the relevant catalogues and websites.

Some of the organisms have been selected and recommended by expert organisations to be supplied as controls for microbiological tests, and when the identical cultures are present in more than one collection they will have a specific designation for each, incorporating the abbreviation for the collection and a reference number.

For example:- *Staphylococcus aureus* NCTC 7447, widely recommended as a control for antimicrobial susceptibility testing, is held in 30 collections, and consequently the phenotypically and genotypically identical organism has 30 different references, such as ATCC 6538P, CIP 53.156, DSM 346 and so on.

In an effort to minimise potential confusion and help users find local sources of reference strains, the WFCC and the WDCM initiated a system that ascribes each recommended QC strain a reference number (WDCM 00001 onwards), cites all collections that contain it and provides contact details and each collection's unique reference. For example, the strain of *Staphylococcus aureus* NCTC 7447 (Selectrol[®] strain MM33) mentioned above is designated WDCM 00033.

Staphylococcus aureus WDCM 00033

AHU 1142; ATCC[™] 6538P; BCRC 10451; BTCC 209P; BU 395; CCM 2022; CCTM 596; CCUG 1828; CECT 240; CIP 53.156; CN 3784; CNCTC Mau 28/58; DSM 346; FIRDI 451; IAM 1011; IAM 12082; IEM Mau 28/58; IFO 12732; IFO 3061; IID 671; IMET 10904; JCM 2151; LMG 8195; NCIMB 8625; NCTC 7447; NRRL B-313; OUT 8232; PCI 1209; PZH 8/54; RIMD 3109007; VNIIA 209P;

Products derived from the cultures in the collections should be manufactured using the minimum number of sub-cultures, to minimise the possibility of alterations to the phenotype due to mutations. Ideally, as in the case of **Selectrol®**, a single sub-culture only is used, so the **Selectrol®** product is a 'first generation derivative' of a culture supplied by NCTC, and will be identical with regard to its properties and suitability for use in QC applications to a culture of the particular organism obtained from any of the other WDCM listed culture collections.

Every effort has been made to ensure the accuracy of the information in this document, however TCS makes no warranties, expressed or implied, regarding errors or omissions and assumes no legal liability or responsibility for loss or damage resulting from the use of information contained within.

Selectrol Strain Index

Strain Name	Designation	Code	WDCM
Aspergillus brasiliensis	NCPF [®] 2275 / ATCC [®] 16404	MM94	00053
Bacillus cereus	NCTC [®] 10320 / ATCC [®] 9634	MM21	00001
Bacillus cereus	NCTC [®] 7464 / ATCC [®] 10876	MM86	
Bacillus subtilis	NCTC [®] 10400 / ATCC [®] 6633	MM29	00003
Bacteroides fragilis	NCTC [®] 9343 / ATCC [®] 25285	MM44	
Campylobacter jejuni	NCTC [®] 11351 / ATCC [®] 33560	MM36	
Campylobacter jejuni	NCTC [®] 11322 / ATCC [®] 29428	MM82	00156
Candida albicans	NCPF [®] 3255 / ATCC [®] 2091	MM28	00055
Candida albicans	NCPF [®] 3179 / ATCC [®] 10231	MM42	00054
Citrobacter freundii	NCTC [®] 9750 / ATCC [®] 8090	MM27	
Clostridium perfringens	NCTC [®] 8237 / ATCC [®] 13124	MM45	00007
Clostridium sporogenes	NCTC [®] 532 / ATCC [®] 19404	MM31	00008
Enterobacter aerogenes	NCTC [®] 10006 / ATCC [®] 13048	MM26	00175
Enterobacter cloacae	NCTC [®] 13380 / ATCC [®] 23355	MM01	00082
Enterobacter cloacae	NCTC [®] 13406	MM51	
Enterococcus faecalis	NCTC [®] 775 / ATCC [®] 19433	MM17	00009
Enterococcus faecalis	NCTC [®] 12697 / ATCC [®] 29212	MM18	00087
Enterococcus faecalis	NCTC [®] 13379 / ATCC [®] 51299	MM52	00085
Enterococcus hirae	NCTC [®] 13383 /ATCC [®] 10541	MM35	00011
Escherichia coli	NCTC [®] 12241 / ATCC [®] 25922	MM02	00013
Escherichia coli	NCTC [®] 11954 / ATCC [®] 35218	MM24	
Escherichia coli	NCTC [°] 10418 / ATCC [°] 10536	MM33	
Escherichia coli	NCTC [®] 12923 / ATCC [®] 8739	MM38	00012
Escherichia coli	NCTC [®] 11560	MM63	
Escherichia coli	NCTC [®] 9001 / ATCC [®] 11775	MM75	00090
Escherichia coli CRE	NCTC [®] 13476	MM57	
Escherichia coli (mcr-1)	NCTC [®] 13846	MM34	
Escherichia coli O157 (non-toxigenic)	NCTC [®] 12900 / ATCC [®] 700728	MM93	00014
Haemophilus influenzae	NCTC [®] 8468 / ATCC [®] 9334	MM100	
Haemophilus influenzae	NCTC [®] 12975 / ATCC [®] 49766	MM37	
Haemophilus influenzae	NCTC [®] 12699 / ATCC [®] 49247	MM81	
Haemophilus influenzae	NCTC [®] 11931	MM98	
Klebsiella aerogenes	NCTC [®] 9528	MM88	
Klebsiella pneumoniae	NCTC [®] 9633 / ATCC [®] 13883	MM04	00097
Klebsiella pneumoniae	NCTC [®] 13368 / ATCC [®] 700603	MM83	
Klebsiella pneumoniae CRE	NCTC [®] 13440	MM55	
Klebsiella pneumoniae CRE	NCTC [®] 13443	MM56	
Klebsiella pneumoniae CRE	NCTC [®] 13438	MM58	

Selectrol Strain Index

Strain Name	Designation	Code	WDCM
Klebsiella pneumoniae CRE	NCTC [®] 13442	MM59	
Lactobacillus brevis	NCTC [®] 13386 / ATCC [®] 8287	MM76	
Legionella pneumophila serogroup 1	NCTC [®] 11192 / ATCC [®] 33152	MM08	00107
Listeria innocua	NCTC [®] 11288 / ATCC [®] 33090	MM92	00017
Listeria monocytogenes	NCTC [°] 7973 / ATCC [°] 35152	MM48	00109
Listeria monocytogenes	NCTC [®] 13372 ATCC [®] 7644	MM77	
Listeria monocytogenes	NCTC [®] 11994	MM87	00019
Neisseria gonorrhoeae	NCTC [®] 8375 / ATCC [®] 19424	MM05	-
Neisseria gonorrhoeae	NCTC [®] 12700 / ATCC [®] 49226	MM96	
Proteus mirabilis	NCTC [®] 13376 / ATCC [®] 14153	MM43	
Proteus mirabilis	NCTC [®] 10975	MM68	
P <mark>roteus</mark> vulgaris	NCTC [°] 4175 / ATCC [°] 13315	MM09	
P <mark>seudomon</mark> as aeruginosa	NCTC [®] 12903 / ATCC [®] 27853	MM10	00025
Pseudomonas aeruginosa	NCTC [®] 12924 / ATCC [®] 9027	MM40	00026
Pseudomonas aeruginosa	NCTC [®] 13359 / ATCC [®] 15442	MM41	
Pseudomonas aeruginosa	NCTC [®] 10662 / ATCC [®] 25668	MM65	00114
Rhodococcus equi	NCTC [®] 1621 / ATCC [®] 6939	MM97	00028
Saccharomyces cerevisiae	NCTC [®] 10716/ ATCC [®] 9763	MM50	00058
Saccharomyces cerevisiae	NCPF [®] 3178	MM73	1
Salmonella Nottingham	NCTC [®] 7832	MM84	
Salmonella Poona	NCTC [®] 4840	MM89	
Salmonella Typhimurium	NCTC [®] 12023/ ATCC [®] 14028	MM11	00031
Serratia marcescens	NCTC [®] 13382 / ATCC [®] 8100	MM12	
Staphylococcus aureus	NCTC [°] 12981 / ATCC [°] 25923	MM13	00034
Staphylococcus aureus	NCTC [®] 12973 / ATCC [®] 29213	MM14	00131
Staphylococcus aureus	NCTC [®] 7447 / ATCC [®] 6538P	MM30	00033
Staphylococcus aureus	NCTC [®] 10788 / ATCC [®] 6538	MM46	00032
Staphylococcus aureus	NCTC [®] 6571 / ATCC [®] 9144	MM85	00035
Staphylococcus aureus (MRSA)	NCTC [®] 12493	MM64	00212
Staphylococcus aureus (MRSA)	NCTC [®] 13373 / ATCC [®] 43300	MM91	00211
Staphylococcus epidermidis	NCTC [®] 13360 / ATCC [®] 12228	MM15	00036
Streptococcus agalactiae	NCTC [®] 8181 / ATCC [®] 13813	MM16	4/1
Streptococcus pneumoniae	NCTC [®] 12695 /ATCC [®] 6303	MM19	
Streptococcus pneumoniae	NCTC [®] 12977 /ATCC [®] 49619	MM95	
Streptococcus pyogenes	NCTC [®] 12696 /ATCC [®] 19615	MM20	
Vibrio parahaemolyticus	NCTC [®] 10885	MM06	00185
Yersinia enterocolitica	NCTC [®] 12982 / ATCC [®] 9610	MM80	00038

Selectrol Strains Listed by WDCM Number

WDCM	Strain Name	Designation	Code
00001	Bacillus cereus	NCTC [®] 10320 / ATCC [®] 9634	MM21
00003	Bacillus subtilis	NCTC [®] 10400 / ATCC [®] 6633	MM29
00007	Clostridium perfringens	NCTC [®] 8237 / ATCC [®] 13124	MM45
00008	Clostridium sporogenes	NCTC [®] 532 / ATCC [®] 19404	MM31
00009	Enterococcus faecalis	NCTC [®] 775 / ATCC [®] 19433	MM17
00011	Enterococcus hirae	NCTC [®] 13383 /ATCC [®] 10541	MM35
00012	Escherichia coli	NCTC [®] 12923 / ATCC [®] 8739	MM38
00013	Escherichia coli	NCTC [®] 12241 / ATCC [®] 25922	MM02
00014	Escherichia coli O157 (non-toxigenic)	NCTC [®] 12900 / ATCC [®] 700728	MM93
00017	Listeria innocua	NCTC [®] 11288 / ATCC [®] 33090	MM92
00019	Listeria monocytogenes	NCTC [®] 11994	MM87
00025	Pseudomonas aeruginosa	NCTC [®] 12903 / ATCC [®] 27853	MM10
00026	Pseudomonas aeruginosa	NCTC [®] 12924 / ATCC [®] 9027	MM40
00028	Rhodococcus equi	NCTC [®] 1621 / ATCC [®] 6939	MM97
00031	Salmonella Typhimurium	NCTC [®] 12023/ ATCC [®] 14028	MM11
00032	Staphylococcus aureus	NCTC [®] 10788 / ATCC [®] 6538	MM46
00033	Staphylococcus aureus	NCTC [®] 7447 / ATCC [®] 6538P	MM30
00034	Staphylococcus aureus	NCTC [®] 12981 / ATCC [®] 25923	MM13
00035	Staphylococcus aureus	NCTC [®] 6571 / ATCC [®] 9144	MM85
00036	Staphylococcus epidermidis	NCTC [®] 13360 / ATCC [®] 12228	MM15
00038	Yersinia enterocolitica	NCTC [®] 12982 / ATCC [®] 9610	MM80
00053	Aspergillus brasiliensis	NCPF [®] 2275 / ATCC [®] 16404	MM94
00054	Candida albicans	NCPF [®] 3179 / ATCC [®] 10231	MM42
00055	Candida albicans	NCPF [®] 3255 / ATCC [®] 2091	MM28
00058	Saccharomyces cerevisiae	NCTC [®] 10716/ ATCC [®] 9763	MM50
00082	Enterobacter cloacae	NCTC [®] 13380 / ATCC [®] 23355	MM01
00085	Enterococcus faecalis	NCTC [®] 13379 / ATCC [®] 51299	MM52
00087	Enterococcus faecalis	NCTC [®] 12697 / ATCC [®] 29212	MM18
00090	Escherichia coli	NCTC [®] 9001 / ATCC [®] 11775	MM75
00097	Klebsiella pneumoniae	NCTC [®] 9633 / ATCC [®] 13883	MM04
00107	Legionella pneumophila serogroup 1	NCTC [®] 11192 / ATCC [®] 33152	MM08
00109	Listeria monocytogenes	NCTC [®] 7973 / ATCC [®] 35152	MM48
00114	Pseudomonas aeruginosa	NCTC [®] 10662 / ATCC [®] 25668	MM65
00131	Staphylococcus aureus	NCTC [®] 12973 / ATCC [®] 29213	MM14
00156	Campylobacter jejuni	NCTC [®] 11322 / ATCC [®] 29428	MM82
00175	Enterobacter aerogenes	NCTC [®] 10006 / ATCC [®] 13048	MM26
00185	Vibrio parahaemolyticus	NCTC [®] 10885	MM06
00211	Staphylococcus aureus (MRSA)	NCTC [®] 13373 / ATCC [®] 43300	MM91
00212	Staphylococcus aureus (MRSA)	NCTC [®] 12493	MM64

Notes





www.tcsbiosciences.co.uk



TCS Biosciences Ltd Botolph Claydon, Buckingham, MK18 2LR, United Kingdom t: +44 (0) 1296 714222, f: +44 (0) 1296 714806, e: sales@tcsgroup.co.uk