



Ref. Certif. No.

DE 3 - 41071

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE**
CERTIFICAT D'ESSAI OC

Product

Name and address of the applicant

Name and address of the manufacturer

Name and address of the factory

Ratings and principal characteristics

Trade mark (if any)

Model/type Ref.

Additional information (if necessary)

A sample of the product was tested and found
to be in conformity withas shown in the Test Report Ref. No.
which forms part of this certificate

Humidifiers for medical use

Hamilton Medical AGVia Crusch 8
7402 Bonaduz
SWITZERLANDHamilton Medical AG
Via Crusch 8, 7402 Bonaduz, SWITZERLANDHamilton Medical AG
Parc Industrial Vial 10, 7013 Domat/Ems, SWITZERLANDRated Voltage: Ref: 950008 100 VAC (Japan)
Ref: 950004, 110-127 VAC (USA)
Ref: 950001, 220-240 VAC (Europe)

Rated Frequency: 50/60 Hz

Rated Power Input: 260 VA (Japan)
293 VA (USA)
283 VA (Europe)Protection Class: I
Type: BFProtection against
Ingress of water: IP21

HAMILTON MEDICAL

HAMILTON-H900, 950008, 950004, 950001

Certificate DE 3 – 40622 issued on 2016-10-05 is replaced by this
certificate due to technical changes. Cl. 11.7, Biocompatibility and
clause 17 EMC are excluded. IEC 62304:2006, IEC 62366:2007 and
IEC 62366:2007/AMD1:2014 are also applied.IEC 60601-1:2005
IEC 60601-1:2005/AMD1:2012
IEC 60601-1-8:2006
IEC 60601-1-8:2006/AMD1:2012
IEC 60601-1-6:2010
IEC 60601-1-6:2010/AMD1:2013

713145152

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

CB 058067 0078 Rev. 00

Date, 2019-01-11

(Paul Bielawski)



Product Service

Page 1 of 1

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany