

SUZHOU COLOUR-WAY NEW MATERIAL CO., LTD.
NO.20, ANMIN ROAD, HUANGDAI TOWN, XIANGCHENG DISTRICT,
215152 SUZHOU CITY, JIANGSU PROVINCE, PEOPLE'S REPUBLIC OF CHINA

The following sample(s) was/were submitted and identified by the client as:

Sample Description : EXAMINATION GLOVES
Material : NATURE LATEX
Lot No. : 20060120
Lot Size : 50000PAIRS
Sample Quantity : 300PCS
Manufacture : SUZHOU COLOUR-WAY NEW MATERIAL CO., LTD.
Manufacturer Date : 2020.06
Storage Condition : AVOID LIGHT
Sample Receiving Date : AUG.12,2020
Testing Period : AUG.12,2020 TO AUG.27,2020

Test Requested

Result


EN 455-2:2015 Medical Gloves for Single Use – Part 2: Requirements and
Testing for Physical Properties (Clause 4.2, 4.3, 5.2, 5.3)

Pass

Remark: Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.

SGS-CSTC Standards
Technical Services (Qingdao)
Co., Ltd.

Jessica Gao



Jessica Gao
Approved Signatory

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.



Test Conducted:

EN 455-2:2015 Medical gloves for single use – Part 2: Requirements and testing for physical properties

Number of test sample	:	26 Pieces
Type	:	examination/procedure gloves: b)
Size	:	Examination/procedure gloves: M

Clause	Test Items	Result
4	Dimensions	---
4.2	Length	Pass (See result 1)
4.3	Width	Pass (See result 1)
5	Strength	---
5.2	Force at break	Pass (See result 2)
5.3	Force at break after challenge testing	Pass (See result 2)

Result 1: Dimensions

Size No.	M	
	Length (mm)	Width (mm)
1	245	94
2	240	94
3	241	93
4	249	94
5	246	95
6	241	94
7	248	94
8	246	94
9	249	95
10	244	94
11	243	95
12	240	94
13	244	94
Standard requirement	≥240	95±10
Median value	244	94



Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)93071443, or email: CH.Doccheck@sgs.com



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Result 2: Strength

Size: M			
Force at break (N)			
Before aging		After aging	
No.	/	No.	/
1	9.8	1	8.2
2	9.5	2	8.1
3	10.2	3	8.8
4	10.4	4	10.3
5	11.0	5	10.1
6	10.2	6	9.9
7	10.1	7	8.8
8	10.3	8	9.3
9	10.4	9	9.7
10	10.7	10	9.6
11	10.8	11	10.2
12	10.7	12	10.4
13	10.6	13	10.1
Standard requirement	≥6.0	Standard requirement	≥6.0
Median value	10.4	Median value	9.7

Remark: The declaration of conformity is only based on the actual value of laboratory activity, measurement uncertainty of the results not take into account.

Sample Photo:

Received sample

Received sample



SGS authenticate the photo on original report only

End of Report

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

