

## **EC Declaration of Conformity**

We, TaiDoc Technology Corporation

B1-7F, No.127, Wugong 2nd Road, 24888 Wugu Dist, New Taipei City, TAIWAN declare under our sole responsibility that the product

**Product Name** 

: Blood Glucose Monitoring System

**Product Model** 

:TD-4227

Classification

: 98/79/EC (IVDD), Annex II, List B

**Conformity Assessment Route** 

98/79/EC (IVD), Annex IV excluding section 4&6

**EC Certificate Number** 

: V1 052126 0069 Rev.00

**European Representative** 

: McdNet EC-REP GmbH

Borkstraße 10, 48163 Münster, Germany

Notified Body (CE0123)

: TÜV SÜD Product Service GmbH

Ridlerstraße 65, 80339 München, Germany

**GMDN** code

: 62537

to which this declaration relates is in conformity with the following standard(s) or other normative document(s):

ISO 13485:2016	Medical devices - Quality management systems - Requirements for regulatory purposes
EN ISO 14971:2012	Medical devices - Application of risk management to medical devices
EN 61010-1:2010	Safety requirements for electrical equipment for measurement, control and laboratory usc. General requirements.
EN 62304:2006+A1:2015	Medical device software Software life cycle processes
EN ISO 15197:2015	In vitro diagnostic test systems —Requirements for blood-glucose monitoring systems forself-testing in managing diabetes mellitus
EN ISO 23640:2015	In vitro diagnostic medical devices. Evaluation of stability of in vitro diagnostic reagents
EN 15223-1:2016	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied. Part 1: General requirements



fet 1866-2-6625-8188 Fac: (886-2-6625-0288

www.taldoc.com	
EN ISO 18113-1:2011	In vitro diagnostic medical devices. Information supplied by the manufacture (labelling). In vitro diagnostic instruments for professional use
EN ISO 18113-3:2011	In vitro diagnostic medical devices. Information supplied by the manufacture (labelling). In vitro diagnostic instruments for professional use
EN ISO 18113-2:2011	In vitro diagnostic medical Devices. Information supplied by the manufacture (labelling). In vitro diagnostic reagents for professional use
EN ISO 18113-4:2011	In vitro diagnostic medical Devices. Information supplied by the manufacture (labelling). In vitro diagnostic reagents for self-testing
EN ISO 18113-5:2011	In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 5: In vitro diagnostic instruments for self-testing
EN 13532:2002	General requirements for in vitro diagnostic medical devices for self-testing
EN 13612:2002	Performance evaluation of in vitro diagnostic medical devices
EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements
EN 61326-2-6 :2013	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-6: Particular requirements - In vitro diagnostic (IVD) medical equipment
EN 61010-2-101:2015	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment
EN 62366-1:2015	Medical devices Application of usability engineering to medical
EN 60601-1-6:2010+A1:2015	Medical electrical equipment. General requirements for basic safety and essential performance. Collateral standard. Usability
EN ISO 17511:2003	In vitro diagnostic medical devices. Measurement of quantities in biological samples. Metrological traceability of values assigned to calibrators and control materials
EN 50581:2012	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances
	84.



泰博科技股份有限公司 | 55年出7488日股區日工 1817月 61-7日 TaiDoc Technology Corp. 81-71. No.177. Wagong 2nd Rd., Wago Dist. Rew Japel City 24888, Jaiwan

fel: 1886-7-6675-8188 Tax: 1886-7-6675-0788

The object of the declaration described above is in conformity with Directive 98/79/EC,2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment.

Date of Issue

Management Representative