

DAC-SpectroMed S.R.L. Chisinau	Declaration of EC-Conformity	File No.	F-PS-10-05
		Rev. No.	2
	Product: CRP-Latex	Rev. Date	07-16-2019
		Doc.No	STD-DCE-116
		Page	1 / 1

DAC-SpectroMed SRL

Nicolae Testemitanu str. 37, MD-2025, Chisinau, Republic of Moldova

hereby declares under its own responsibility that the *in-vitro diagnostics* medical devices:

CRP-Latex

REF	1033C50	50 tests
REF	1033C100	100 tests
REF	1033C200	200 tests
REF	1033C250	250 tests
REF	1033C500	500 tests

- are classified as not listed in Annex II of the EC Council Directive 98/79/EC from 27th October 1998 on in-vitro diagnostic medical devices;
- conform to the relevant provisions of the EC Council Directive 98/79/EC from 27th October 1998 on in-vitro diagnostic medical devices.

Harmonized standards applied:

EN ISO 9001:2015

EN ISO 14971:2012

EN ISO 18113-1:2011

EN ISO 18113-2:2011

EN ISO 23640:2011

EN ISO 13612:2002

EN ISO 15223-1:2012

EN ISO 13485:2016

DAC-SpectroMed SRL has a Quality System in place based on EN ISO 9001:2015 and EN ISO 13485:2016, issued by the IQNet&SRAC.

Our Authorized Representative in EU is Qarad B.V., with address at Flight Forum 40, 5657 DB Eindhoven, The Netherlands.





General Manager

Virschi Roman

16.07.2019

Head of Normative-Technical Documentation
and Certification Department



Emet Natalia

16.07.2019