



Technical Data

Grams Stain-Kit

K001

Intended Use

Grams Stain Kit is used for differentiation of bacteria on the basis of their gram nature.

Composition**

Ingredients

Gram's Crystal Violet (S012)(Solution A)

Crystal Violet	-
Ethyl alcohol, 95%	2.000 gm
	20.000 ml

Gram's Crystal Violet (S012)(Solution B)

Ammonium oxalate	-
Distilled Water	0.800 gm
	80.000 ml

Solution A and B are mixed and stored for 24 hours before use. The resulting stain is stable.

Gram's Decolourizer(S032)

Ethyl alcohol, 95%	-
Acetone	50.0 ml
	50.0 ml

Gram's Iodine(S013)

Iodine	-
Potassium iodide	1.000 gm
Distilled water	2.000 gm
	300.000 ml

Safranin, 0.5% w/v(S027)

Safranin O	-
Ethyl alcohol, 95%	0.500 gm
	100.000 ml

**Formula adjusted, standardized to suit performance parameters

Directions

1. Prepare a thin smear on clear, dry glass slide.
2. Allow it to air dry and fix by gentle heat.
3. Flood with Gram's Crystal Violet (S012) for 1 minute. (If over staining results in improper decolourization of known gram-negative organisms, use less crystal violet).
4. Drain the stain.
5. Flood the smear with Gram's Iodine (S013). Allow it to remain for 1 minute.
6. Decolourize with Gram's Decolourizer (S032) until the blue dye no longer flows from the smear.
7. Wash with tap water.
8. Counter stain with 0.5% w/v Safranin (S027). Allow it to remain for 1 minute.
9. Wash with water.
10. Allow the slide to air dry or blot dry between sheets of clean bibulous paper and examine under oil immersion objective.

Principle And Interpretation

The Gram stain is a differential staining technique most widely applied in all microbiology disciplines laboratories. It is one of the most important criteria in any identification scheme for all types of bacterial isolates. Different mechanisms have been proposed to explain the gram reaction. There are many physiological differences between gram-positive and gram-negative cell walls. Ever since Christian Gram has discovered Gram staining, this process has been extensively investigated and redefined. In practice, a thin smear of bacterial cells is stained with crystal violet, then treated with an iodine containing mordant to increase the binding of primary stain. A decolourizing solution of alcohol or acetone is used to remove the crystal violet from cells which bind it weakly and then the counterstain (like safranin) is used to provide a colour contrast in those cells that are decolourized. Gram-positive bacteria have a thick mesh-like cell wall made of peptidoglycan (50–90% of cell envelope), and as a result are stained purple by crystal violet, whereas gram-negative bacteria have a thinner layer (10% of cell envelope), so do not retain the purple stain and are counter-stained pink by safranin. In a properly stained smear by gram staining procedure, the gram-positive bacteria appear blue to purple and gram negative cells appear pink to red.

Type of specimen

Any isolated colony on primary or subculture plates can be isolated from following specimens. Clinical specimen: Blood, urine, CSF, pus, wounds, lesions, body tissues, sputum etc. From environment: Air, water, soil, sludge, waste water, food, dairy samples etc.

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines. For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines. For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards. Generally the smear is made in laboratory; however, when there is a concern that transport will be delayed or that the preservation for culture will alter the specimen, prepare smear and submit slides to the laboratory.

Warning and Precautions :

In Vitro diagnostic Use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Use results of Gram stains in conjunction with other clinical and laboratory findings. Use additional procedures (e.g., special stains, inclusion of selective media, etc.) to confirm findings suggested by gram-stained smears.
2. Proper smear preparation is key to obtaining good gram staining results. Avoid excessive material or thick smears which may interfere with the passage of light and lead to distortion of images.
3. Overheating slides during heat fixation can distort the appearance of the organisms.
4. Only fresh cultures and specimens should be gram stained since cell wall integrity of older cells may give improper gram-staining characteristics. Gram positive organisms that have lost cell wall integrity because of old age or antibiotic treatment may appear pink.
5. The decolorization step is the most important step in the gram-staining process. Over decolorization results in a abundance of bacteria that appear gram negative, while under decolorization results in too many bacteria that appear to be gram-positive.
6. The procedure given is based on an ideal thin smear of cells. Staining and decolorization times may vary depending on the sample and its thickness.
7. False Gram stain results may be related to inadequately collected specimens or delay in transit.

Performance and Evaluation

Performance of the product is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Microscopic examination

Gram staining is carried out using gram's Stain Kit and observed under oil immersion lens.

Results

Gram-positive organisms : Violet coloured

Gram-negative organisms : Pinkish red coloured

Storage and Shelf Life

Store between 10- 30°C in tightly closed container and away from bright light. Use before expiry date on label. On opening, product should be properly stored in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques.

Reference

1. Downes F. P. and Ito K. (Ed.), 2001, Compendium of Methods for the Microbiological Examination of Foods, 4th ed., APHA, Washington, D.C.
2. Rice E.W., Baird, R.B., Eaton A. D., Clesceri L. S. (Eds.), 2012, Standard Methods for the Examination of Water and Wastewater, 22nd ed., APHA, Washington, D.C.
3. Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.
4. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
5. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock, D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
6. Shanholtzer, C.J., P. Schaper, and L.R. Peterson. 1982. Concentrated Gram stain smear prepared with a cytospin centrifuge. J.clin. Microbiol. 16:1052-1056
7. Thorpe, J.E., R.P. Banghman, P.T. Frame, T.A. Wessler, and J.L. Staneck. 1987. Bronchoalveolar lavage for diagnosing acute bacterial pneumoniae. J.Infect.Dis. 155:855-861
8. Brown, M.S., and T.C. Wu. 1986. The Gram stain morphology of fungi, mycobacteria, and Pneumocystis carinii. J.Med. Technol. 3:495-499
9. George Clark et al, 1981, 4th ed., Staining procedures: 17(375-379)
10. Godkar B. P., 1996, Textbook of medical laboratory technology: 23(309-313)


Revision : 01 / 2019



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
B /4-6 , MIDC, Dindori, Nashik MH
www.himedialabs.com



CE Partner 4U, Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.