



# JianErKang Medical Co.,Ltd

No 1 Jianerkang Road Zhixi Town Industrial Zone  
Jintan District Changzhou City Jiangsu China

## CERTIFICATE OF ANALYSIS

REPORT NO: F2024090603

QA6.0-011

|                              |                                       |   |              |   |            |
|------------------------------|---------------------------------------|---|--------------|---|------------|
| Absorbent Gauze              | 40S/30×26                             |   | Product Name | 203282 Gauze bandage 14cmx7m<br>non sterile 100% cotton |            |
| Contract No.                 | Samples                               |   | Lot No.      |   | 20240903   |
| Mfg date                     | 2024-09-03                            |   | Expiry Date  |   | 2029-09-02 |
| Quantity                     | 4 rolls                               |   | Test date    |   | 2024.09.05 |
| Customer:                    |                                       |   |              |   |            |
| Item                         | Unit                                  | Standard（EN14079）                                     |              | Test result   | Result     |
| Warp/ Weft                   | Pc/10cm                               | 121±5/101±5   |              | 118/100   | pass       |
| Mass per square              | g/㎡                                   | ≥32   |              | 32.5  | pass       |
| Minimum breaking load        | N                                     | warp≥60 N/ weft≥40 N                                  |              | 65.1/42.3   | pass       |
| whiteness                    | %                                     | Not be less than 80                                   |              | 81  | pass       |
| Loss on drying               | %                                     | Not be more than8.0                                   |              | 4.2   | pass       |
| Sinking time                 | s                                     | Not be more than10                                    |              | 2.7   | pass       |
| Surface active substance     | mm                                    | shall not exceed 2 mm.                                |              | no  | pass       |
| Fluorescence                 | /                                     | Purple、 small yellow particles、 no blue               |              | conformance   | pass       |
| Acidity-alkalinity           | /                                     | May not show pink（phenolphthalein、 methyl orange）     |              | conformance   | pass       |
| Water-soluble substances     | %                                     | Not be more than0.50                                  |              | 0.27  | pass       |
| Ether-soluble substances     | %                                     | Not be more than0.50                                  |              | 0.20  | pass       |
| Starch and dextrin           | /                                     | not show any blue, violet, reddish or brownish colour |              | conformance   | pass       |
| Sulphated ash                | %                                     | □ ≤0. 40； □ ≤0. 75； □ ≤1. 20； □ ≤0. 45                |              | 0.14  | pass       |
| Extractable colouring matter | /                                     | Not more than the purified water                      |              | conformance   | pass       |
| Foreign fibres               | /                                     | Only allow some little isolated fiber                 |              | conformance   | pass       |
| Fiber identification         | A                                     | um  | 10-40        | 24  | pass       |
|                              | B                                     |   | Purple       | Purple  | pass       |
|                              | C                                     |   | No soluble   | No soluble  | pass       |
| Conclusions                  | This product conforms to the Standard |   |              |   |            |

Inspector: Feng Weifen 健尔康医疗科技股份有限公司  
JIANERKANG MEDICAL CO.,LTD

Reviewed by: Ma Xiao jia

陳國平