

Certificate of Analysis

ThermoFisher
SCIENTIFIC

POP-7 POLYMER DX (384 SAMPLES)

REF 4393709
LOT 2601234
 2026-02-13
 2026-08-18

TEST	RESULT
------	--------

VISUAL INSPECTION

NOTE The polymer is particulate free at room temperature, but upon refrigeration, urea may precipitate out of solution.

Clear, particulate free, viscous solution at room temperature	Pass
---	------

FLUORESCENCE TEST

The ratio of the background fluorescence intensity of the polymer to that of water is measured for each bulk lot to ensure conformance to manufacturing specifications

Fluorescence Ratio	Pass
--------------------	------

RESOLUTION TEST

Resolution is evaluated by measuring separations and widths of peaks in a multi-fragment DNA standard on a multi-capillary Genetic Analyzer

Resolution	Pass
------------	------

Manufactured in compliance with our ISO 13485 certified quality management system and 21 CFR 820 quality system regulations.

Made in USA

Life Technologies Holdings Pte Ltd
Block 33, Marsiling Industrial Estate Road 3 #07-06
Singapore 739256
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com



**Quality Assurance
Issued 17FEB2026**