



## DECLARATION OF CONFORMITY



DLAB Scientific Co., Ltd  
Add: Yu An Road 31, Airport  
Economic core zone, Shun yi District,  
Beijing, 101318 P.R.China

Product: **Clinical Centrifuges**

Model (s): DM0408 DM0424

The designated products are conform to the guideline:

**2014/35/EU** Low Voltage Directive  
**2014/30/EU** Electromagnetic Compatibility Directive  
**2015/863/EU** RoHS Directive

The conformity of the products with the requisites as set forth in the guideline is evidenced by the full compliance with the following standards:

|                |                                                                                                                                                            |
|----------------|------------------------------------------------------------------------------------------------------------------------------------------------------------|
| EN 61010-1     | Safety Requirements for Electrical Equipment for Measurement Control, and Laboratory Use – Part 1: General Requirements                                    |
| EN 61010-2-020 | Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-020: Particular requirements for laboratory centrifuges |
| EN 61326-1     | Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements                                         |
| EN 61000-3-2   | Electromagnetic compatibility (EMC) - Part 3-2: Limits - Limits for harmonic current emissions                                                             |
| EN 61000-3-3   | Electromagnetic compatibility (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker                                   |

Beijing

2025-01-01

  
DLAB Scientific Co., Ltd  
Quality Manager  
