

General Product Information

Product Name: FLEX Mono MxH CD15, Clone Carb-3 RTU, Link China

Code Number: IR06261-2CN

Storage data: See Product label

Lot Specific Data

Lot Number: 11775395C Expiration date: 2026SEP30

Specifications

Test	Specification	Result
Appearance	Clear, colorless liquid	Pass
Immunohistochemistry Specific Stain	HE: 3.0 to 4.0 specific staining LE: 0.5 to 3.0 specific staining	Pass
Immunohistochemistry Background	≤1.0 background staining	Pass
Immunohistochemistry Equivalent	0.25 grade	Pass
Immunohistochemistry Comparable	Staining pattern of test lot must be consistent with the antibody used and comparable to control slide	Pass

 QC approval
 Eric Smith
 Date
 2025MAR04

 QA approval
 Jeff Taylor
 Date
 2025MAR05



General Product Information

Product Name: FLEX Monoclonal Mouse X-H CD1a Clone 010, RTU, (Link), CN

Code Number: IR06961-2CN

Storage data: See Product label

Lot Specific Data

Lot Number: 11794330A Expiration date: 2026JUN30

Test	Specification	Result
Appearance	Clear, colorless liquid	Pass
Immunohistochemistry Specific Stain	HE: 2.5 to 4.0 specific staining LE: 1.0 to 2.5 specific staining	Pass
Immunohistochemistry Background	≤1.0 background staining	Pass
Immunohistochemistry Equivalent	0.25 grade	Pass
Immunohistochemistry Comparable	Staining pattern of test lot must be consistent with the antibody used and comparable to control slide	Pass

QC approval	Maria Jeronimo	Date	2025FEB21
QA approval	Jeff Taylor	_ Date	2025FEB21



General Product Information

Product Name: FLEX Polyclonal Rabbit Anti-Human CD3, Ready-to-Use (Link)

Code Number: IR503

Storage data: See Product label

Lot Specific Data

Lot Number: 41793148 Expiration date: 2026AUG31

Test	Specification	Result
Immunohistochemistry	Human Tissue section. Location of staining and staining intensity of the lot must comply to reference lot. This applies for both specific and unspecific	Pass
	staining.	

QC approval	Merete Johnsoen	_ Date	2024OCT02
QA approval	Janne Bayer	_ Date	2024DEC04



General Product Information

Product Name: FLEX Monoclonal Mouse Anti-Human CD21, Clone 1F8, Ready-to-Use

(Link)

Code Number: IR608

Storage data: See Product label

Lot Specific Data

Lot Number: 41774434 Expiration date: 2026SEP30

Test	Specification	Result
Immunohistochemistry	Human Tissue section. Location of staining and staining intensity of the lot must comply to reference lot. This applies for both specific and unspecific	Pass
	staining.	

QC approval	Merete Johnsoen	Date	2024OCT18
QA approval	Janni Buch	Date	2024NOV04



General Product Information

Product Name: FLEX Monoclonal Mouse Anti-Human CD8, Clone C8/144B, Ready-to-

Use (Link)

Code Number: IR623

Storage data: See Product label

Lot Specific Data

Lot Number: 41718521 Expiration date: 2026MAR31

Test	Specification	Result
Immunohistochemistry	Human Tissue section. Location of staining and staining intensity of the lot must comply to reference lot. This applies for both specific and unspecific	Pass
	staining.	

QC approval	Heidi Joergensen	_ Date	2024MAY02
QA approval	Maria Nielsen	_ Date	2024MAY24



General Product Information

Product Name: FLEX Monoclonal Mouse Anti-Human CD138, Clone MI15, Ready-to-Use

(Link)

Code Number: IR642

Storage data: See Product label

Lot Specific Data

Lot Number: 41752866 Expiration date: 2026AUG31

Test	Specification	Result
Immunohistochemistry	Human Tissue section. Location of staining and staining intensity of the lot must comply to reference lot. This applies for both specific and unspecific	Pass
	staining.	

QC approval	Heidi Joergensen	_ Date	2024SEP19
QA approval	Maria Nielsen	_ Date	2024OCT01



General Product Information

Product Name: FLEX Monoclonal Mouse Anti-Human CD4, Clone 4B12, Ready-to-Use

(Link)

Code Number: IR649

Storage data: See Product label

Lot Specific Data

Lot Number: 41773478A Expiration date: 2026OCT31

Test	Specification	Result
immunonistocnemistry	Human Tissue section. Location of staining and staining intensity of the lot must comply to reference lot. This applies for both specific and unspecific	Pass
	staining.	

QC approval	Merete Johnsoen	_ Date	2024NOV14
QA approval	Janni Buch	_ Date	2025JAN29



General Product Information

Product Name: FLEX Monoclonal Mouse Anti-Human B-Cell-Specific Activator Protein,

Clone DAK-Pax5, Ready-to-Use (Link)

Code Number: IR650

Storage data: See Product label

Lot Specific Data

Lot Number: 41745073A Expiration date: 2026JUL31

Test	Specification	Result
Immunohistochemistry	Human Tissue section. Location of staining and staining intensity of the lot must comply to reference lot. This applies for both specific and unspecific	Pass
	staining.	

QC approval	Merete Johnsoen	_ Date	2024AUG07
QA approval	Lisbet Drewsen	_ Date	2025JAN02



General Product Information

Product Name: FLEX Monoclonal Mouse Anti-Human Beta-Catenin, Clone ß-Catenin-1,

Ready-to-Use (Link)

Code Number: IR702

Storage data: See Product label

Lot Specific Data

Lot Number: 41749786 Expiration date: 2026JUN30

Test	Specification	Result
Immunonistochemistry	Human Tissue section. Location of staining and staining intensity of the lot must comply to reference lot. This applies for both specific and unspecific	Pass
	staining.	

QC approval	Merete Johnsoen	Date	2024AUG23
QA approval	Janni Buch	_ Date	2024SEP12



General Product Information

Product Name: FLEX Monoclonal Mouse Anti-Human CD10, Clone DAK-CD10, Ready-

to-Use (Link)

Code Number: IR786

Storage data: See Product label

Lot Specific Data

Lot Number: 41790703 Expiration date: 2025SEP30

Test	Specification	Result
immunonistocnemistry	Human Tissue section. Location of staining and staining intensity of the lot must comply to reference lot. This applies for both specific and unspecific	Pass
	staining.	

QC approval	Merete Johnsoen	Date	2024OCT18
QA approval	Janni Buch	_ Date	2024DEC20



General Product Information

Product Name: EnVision FLEX+, Mouse, High pH (Link)

Code Number: K8002

Storage data: See Product label

Lot Specific Data

Lot Number: 41803698 Expiration date: 2025NOV12

Test	Specification	Result
Immunohistochemistry	Human Tissue section. Location of staining and staining intensity of the lot must comply to reference lot. This applies for both specific and unspecific	Pass
	staining.	

QC approval	Line Hansen	_ Date	_2025JAN17
QA approval	Maria Nielsen	_ Date	2025FEB27



General Product Information

Product Name: EnVision FLEX Target Retrieval Solution Low pH (50x)

Code Number: K8005

Storage data: See Product label

Lot Specific Data

Lot Number: 41793185 Expiration date: 2025NOV19

Test	Specification	Result
pH measurement	pH: 5,50 – 5,60	Pass

QC approval	Annette Wisler	_ Date	2024NOV20
QA approval	Janni Buch	Date	2025JAN27



General Product Information

Product Name: EnVision FLEX Antibody Diluent

Code Number: K8006

Storage data: See Product label

Lot Specific Data

Lot Number: 41788283 Expiration date: 2026MAR27

Test	Specification	Result
Immunohistochemistry	Human Tissue section. Location of staining and	Pass
Inimunoriistochemistry	staining intensity of the lot must comply to reference	ass
	lot. This applies for both specific and unspecific	
	staining.	

QC approval	MICHAEL KONGSGAARD	_ Date	2020DEC07
QA approval	Janne Bayer	_ Date	2024DEC18



General Product Information

Product Name: EnVision FLEX Wash Buffer (20x)

Code Number: K8007

Storage data: See Product label

Lot Specific Data

Lot Number: 11806788 Expiration date: 2026FEB28

Specifications

Test	Specification	Result
Appearance	The solution is clear with no undissolved raw materials by visual inspection	Pass
рН	7.7 (±0.1) @ 25°C	Pass
Conductivity	150 mS/cm (±3.0) @ 25°C	Pass

 QC approval
 Jon St.Onge
 Date
 2025MAR04

 QA approval
 Jeff Taylor
 Date
 2025MAR05



General Product Information

Product FLEX IHC Microscopes Slides

Code K8020

Presentation Material provided with K8020:

Coated microscope slides, 75 mm (W) x 25 mm (D) x 1 mm (H) with a white-painted label area. Provided as 5

boxes each containing 100 slides.

Storage Store at 15-30 °C.

Lot Specific Data

Lot Number: 20240706 Expiration date: 2026JUL06

Specifications

Test	Test Material	Specification
QA control	-	Certificate of Compliance is related to DMRI and IFU

Date: 2025FEB27

f. QC Approval



General Product Information

Product Name: Monoclonal Mouse Anti-Human Epithelial Membrane Antigen, Clone

E29

Code Number: M0613

Storage data: See Product label

Lot Specific Data

Lot Number: 41790451 Expiration date: 2027NOV07

Test	Specification	Result
immunonistocnemistry	Human Tissue section. Location of staining and staining intensity of the lot must comply to reference lot. This applies for both specific and unspecific	Pass
	staining.	

QC approval	Charlotte Kroell	Date	2023MAY25
QA approval	Janne Bayer	_ Date	2024NOV28



General Product Information

Product Name: Mo a Hu CD45, LCA, cls 2B11 + PD7/26, CN

Code Number: M0701

Storage data: See Product label

Lot Specific Data

Lot Number: 41700680A Expiration date: 2027MAR08

Test	Specification	Result
Immunohistochemistry	Human Tissue section. Location of staining and staining intensity of the lot must comply to reference lot. This applies for both specific and unspecific	Pass
	staining.	

QC approval	Merete Johnsoen	_ Date	2024MAR15
QA approval	Janni Buch	_ Date	2024NOV22



General Product Information

Product Name: Monoclonal Mouse Anti-Human CD30, Clone Ber-H2

Code Number: M0751

Storage data: See Product label

Lot Specific Data

Lot Number: 41775232 Expiration date: 2027OCT15

Test	Specification	Result
Immunohistochemistry	Human Tissue section. Location of staining and staining intensity of the lot must comply to reference lot. This applies for both specific and unspecific	Pass
	staining.	

QC approval	Charlotte Kroell	_ Date	2024FEB07
QA approval	Janni Buch	_ Date	2024NOV06



General Product Information

Product Name: Monoclonal Mouse Anti-Human CD20cy, Clone L26

Code Number: M0755

Storage data: See Product label

Lot Specific Data

Lot Number: 41769962 Expiration date: 2027SEP20

Test	Specification	Result
Immunohistochemistry	Human Tissue section. Location of staining and staining intensity of the lot must comply to reference lot. This applies for both specific and unspecific	Pass
	staining.	

QC approval	Perihan Turan	_ Date	2021OCT28
QA approval	Janni Buch	_ Date	2024OCT22