
	BREATHING SYSTEMS TECHNICAL DATA SHEET	Document No	TDS.BS
		Release Date	09.08.2023
		Rev. No	01
		Rev. Date	05.09.2023
		Page No	1 / 3


BREATHING CIRCUIT NAME/ REFERENCE NUMBER	Anesthesia Circuit, Adult, Extendible Tubing, Latex-free Bag/31305020-15																						
MANUFACTURER NAME	R VENT Medikal Uretim A.S. Yazibasi Mah. Balkan Cad. No:33, Torbalı, 35860- Izmir, Turkey		Tel: +90 232 853 9500 E-mail: info@rventmedikal.com																				
REGULATORY APPROVALS AND CERTIFICATION	ISO 13485 – 31816401 CE Certificate – 2195-MED-1816401																						
CLASSIFICATION	Disposable Medical Device <u>MDD 93/42/EEC</u> Class IIa Rule 2 Annex V, Article 3																						
GMDN CODE/DESCRIPTION	37704 Anaesthesia breathing circuit, single-use An assembly of devices designed to conduct medical gases from an anaesthesia unit/workstation to a patient artificial airway/anaesthesia mask (not included) during general anaesthesia. It includes breathing tubes and a Y-piece connector, typically with a ventilator/ventilation bag and appropriate connectors, and may include a carbon dioxide (CO2) absorber, a one-way directional valve, or adjustable pressure limiting (APL) valve. This is a single-use device.																						
EMDN CODE/DESCRIPTION	R02010101 Breathing Circuits, w/o Water Trap																						
FEATURES	<ul style="list-style-type: none">• Disposable breathing circuits may help reduce cross-contamination.• Available in a wide variety of tubing styles, components and configurations to meet specific needs.																						
INTENDED USE	Disposable breathing circuit for conduction of respiratory gases between anesthesia machine or ventilator and patient and intended for single use only. Sterile and Non-sterile options are available. Breathing bag with connection hose (limb) intended for use with anesthesia delivery systems as a reservoir during automatic ventilation and as a manual breathing bag during manual ventilation.																						
TECHNICAL SPECIFICATIONS	<div><p>Drawing:</p></div> <div><p>Materials:</p><table><thead><tr><th>Components</th><th>Materials</th></tr></thead><tbody><tr><td>1 22M – 22F Straight Connector</td><td>Ethylene vinyl acetate (EVA)</td></tr><tr><td>2 22 Mm Extendible Tubing 180cm</td><td>Polyvinyl Chloride (PVC) (PHT FREE)</td></tr><tr><td>3 Y Connector W/Out Port</td><td>Polypropylene (PP)</td></tr><tr><td>4 Tethered Cap</td><td>Low-density polyethylene (LDPE)</td></tr><tr><td>5 Elbow Connector with CO2 Port</td><td>Polypropylene (PP)</td></tr><tr><td>6 22MM Extendible Tubing</td><td>Polypropylene (PP)</td></tr><tr><td>7 22M-22M/15F Straight Connector</td><td>Polypropylene (PP)</td></tr><tr><td>8 2lt Breathing Bag</td><td>Neoprene</td></tr><tr><td>9 Long Connector Red Cap</td><td>Ethylene vinyl acetate (EVA)</td></tr></tbody></table></div> <p>This product does not contain any metallic parts.</p>			Components	Materials	1 22M – 22F Straight Connector	Ethylene vinyl acetate (EVA)	2 22 Mm Extendible Tubing 180cm	Polyvinyl Chloride (PVC) (PHT FREE)	3 Y Connector W/Out Port	Polypropylene (PP)	4 Tethered Cap	Low-density polyethylene (LDPE)	5 Elbow Connector with CO2 Port	Polypropylene (PP)	6 22MM Extendible Tubing	Polypropylene (PP)	7 22M-22M/15F Straight Connector	Polypropylene (PP)	8 2lt Breathing Bag	Neoprene	9 Long Connector Red Cap	Ethylene vinyl acetate (EVA)
Components	Materials																						
1 22M – 22F Straight Connector	Ethylene vinyl acetate (EVA)																						
2 22 Mm Extendible Tubing 180cm	Polyvinyl Chloride (PVC) (PHT FREE)																						
3 Y Connector W/Out Port	Polypropylene (PP)																						
4 Tethered Cap	Low-density polyethylene (LDPE)																						
5 Elbow Connector with CO2 Port	Polypropylene (PP)																						
6 22MM Extendible Tubing	Polypropylene (PP)																						
7 22M-22M/15F Straight Connector	Polypropylene (PP)																						
8 2lt Breathing Bag	Neoprene																						
9 Long Connector Red Cap	Ethylene vinyl acetate (EVA)																						




















	BREATHING SYSTEMS TECHNICAL DATA SHEET	Document No	TDS.BS
		Release Date	09.08.2023
		Rev. No	01
		Rev. Date	05.09.2023
		Page No	2 / 3

	Appearance: As shown on drawing	
	Recommended Patient: Adult	
	Length of Circuit: 180 cm	
	Connection Port(s): 15mm ID & 22mm OD	
TESTS PERFORMED ON THE PRODUCT	-The Leakage Test -The Pull Test -The Gauge Test -The Routine Assembling And Packaging Process Controls	
APPLICABLE STANDARDS	Standard Number	Standard Name
	TS EN ISO 5356-1:2015	Anaesthetic and respiratory equipment — Conical connectors — Part 1: Cones and sockets
	TS EN ISO 11135:2014	Sterilization of health-care products — Ethylene oxide — Requirements for the development, validation and routine control of a sterilization process for medical devices
	TS EN ISO 10993-1:2021	Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process
	TS EN ISO 10993-5:2010	Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity
	TS EN ISO 10993-10:2014	Biological evaluation of medical devices — Part 10: Tests for irritation and skin sensitization
	TS EN ISO 10993-12:2021	Biological evaluation of medical devices — Part 12: Sample preparation and reference materials
	TS EN ISO 5362:2019	Anaesthetic reservoir bags
	TS EN ISO 5367:2015	Anaesthetic and respiratory equipment - Breathing sets and connectors
	ISO 13485:2016	Medical devices - Quality management systems - Requirements for regulatory purposes
	TS EN ISO 15223-1:2021	Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements
	TS EN ISO 20417:2021	Medical devices - Information to be supplied by the manufacturer
	TS EN ISO 14644-1:2016	Cleanrooms and associated controlled environments Part 1: Classification of air cleanliness
	TS EN ISO 11607-1: 2020	Packaging for terminally sterilized medical devices - Part 1: Requirements for materials, sterile barrier systems and packaging systems
	TS EN ISO 14971:2020	Medical devices - Application of risk management to medical devices
	TS EN ISO 24971:2021	Medical devices — Guidance on the application of ISO 14971
	TS EN ISO 10993-7:2010	Biological evaluation of medical devices part 7: Ethylene oxide sterilization residuals
	TS EN ISO 10993-11: 2018	Biological evaluation of medical devices - Part 11: Tests for systemic toxicity
	TS EN ISO 11737-1:2018	Sterilization of health care products — Microbiological methods — Part 1: Determination of a population of microorganisms on products
	TS EN ISO 11737-2: 2020	Sterilization of health care products - Microbiological methods - Part 2: Tests of sterility performed in the definition, validation and maintenance of a sterilization process
	TS EN 62366-1: 2015	Medical devices - Part 1: Application of usability engineering to medical devices
STERILIZATION STATUS	Non-sterile	
CLEANING	Device assembled within ISO 8 Cleanroom.	
PRODUCT SHELF LIFE	5 years from the date of manufacturing. Expiration date and date of production are detailed on the product labelling.	
PACKAGING	Pouch: Polyethylene (PE) Box material: Craft Box dimention: 400 mm x 800 mm x 560 mm Quantity per box: ...	
STORAGE CONDITIONS	Temperature: -20°C to +55°C Humidity: 0% to 95% Luminosity: Keep away from direct sunlight	
TRANSPORTATION CONDITIONS	Temperature: -20°C to +55°C Humidity: 0% to 95%	



	BREATHING SYSTEMS TECHNICAL DATA SHEET	Document No	TDS.BS
		Release Date	09.08.2023
		Rev. No	01
		Rev. Date	05.09.2023
		Page No	3 / 3

		Luminosity: Keep away from direct sunlight		
PRECAUTIONS		Keep away from sunlight		Sterilized with Ethylene Oxide *for sterile products
		Do not use if package is opened or damaged		CE Marking
		Do not re-use		Non-sterile *for non sterile products
		Phthalate-free		Lot number
		Consult instruction for use		Catalog Number
		Latex-free		Expiry Date
		Do not re sterilize *for sterile products		Contains Latex *for products made with Latex
		Storage conditions +55 °C -20 °C		
		Country of manufacture – Date of manufacture		
		Manufacturer		
WASTE METHOD	Local regulations and/or hospital waste management procedures of the relevant country should be followed when disposing of the used products.			
NOTES	-			

