

WATO EX-65

Anesthesia System

Physical Specifications

Dimensions and Weight

| | |
|--------|--------------------------------------------------------------------------------|
| Height | 1370 mm |
| Width | 780 mm (not including breathing system) 945 mm (including breathing system) |
| Depth | 690 mm |
| Weight | <145 kg (without vaporizers and cylinders) |

Top Shelf

| | |
|--------------|--------|
| Weight limit | 30 kg |
| Width | 305 mm |
| Length | 545 mm |

Work Surface

| | |
|--------|----------------------|
| Height | 850 mm |
| Area | 1635 cm ² |

Drawer (3Xdrawers, Internal Dimension)

| | |
|--------|--------|
| Height | 130 mm |
| Width | 415 mm |
| Depth | 320 mm |

Bag Arm

| | |
|------------|----------------------|
| Height | 1150 mm |
| Length | 312 mm |
| Connection | ISO 22mm OD, 15mm ID |

Casters

| | |
|----------|----------------------------------------------|
| Diameter | 125 mm |
| Brakes | Center brake system with Lock / Unlock icons |

Ventilator Specifications

Modes of Ventilation

Manual/Spontaneous Ventilation/Bypass
Volume Control Ventilation (VCV) with PLV function
Pressure Control Ventilation (PCV) with/without volume guarantee (VG)
Synchronized Intermittent Mandatory Ventilation (SIMV-Volume Controlled and SIMV-Pressure Controlled)
Pressure Support Ventilation (PS) with apnea backup
Synchronized Intermittent Mandatory Ventilation Volume Guarantee (SIMV-VG)
Continuous Positive Airway Pressure/Pressure Support Ventilation (CPAP/PS)

Compensation

Circuit gas leakage compensation and automatic compliance compensation

Ventilation Parameters Range

| | |
|------------------------------|-----------------------------------------------------------------|
| Patient Size | Adult, Pediatric, Neonate |
| Tidal volume | 10~1500 mL (Volume Mode) 5~1500 mL (Pressure Mode) |
| Pinsp | 5~80 cmH ₂ O |
| Plimit | 10~100 cmH ₂ O |
| ΔPsupp | 3~60 cmH ₂ O 0, 3~60 cmH ₂ O (CPAP/PS) |
| Rate | 2~100 bpm |
| I:E | 4:1 - 1:8 |
| Inspiratory pause (Tip:Ti) | OFF, 5% - 60% |
| Inspiratory time (Tinsp) | 0.2 - 10.0 s |
| Trigger window | 5% - 90% |
| Flow trigger | 0.2 ~ 15 L/min |
| Pressure trigger | -20~ -1 cmH ₂ O |
| Expiration termination level | 5% - 80% |
| Minimum Rate | 2 - 60 bpm |



| | |
|------------------------------------------------|-----------------------------------|
| Tslope | 0.0 - 2.0 s |
| Apnea I: E | 4:1~1:8 |
| ΔPapnea | 3 - 60 cmH ₂ O |
| Positive End Expiratory Pressure (PEEP) | |
| Type | Integrated, electronic controlled |
| Range | OFF, 3~30 cmH ₂ O |

Ventilator Performance

| | |
|------------------|----------------------------|
| Driving pressure | 280 kPa to 600 kPa |
| Peak gas flow | 120 L/min + Fresh Gas Flow |

Monitoring Parameters

| | |
|-------------------------------------|-------------------------------------------|
| Minute volume | 0 ~ 100 L/min |
| Tidal volume | 0~3000 ml |
| Inspired oxygen (FiO ₂) | 18% ~ 100% |
| Airway pressure | -20 ~ 120 cmH ₂ O |
| I:E | 50:1 ~ 1:50 |
| Rate | 0~120 bpm |
| PEEP | 0 ~ 70 cmH ₂ O |
| Resistance (R) | 0 ~ 600 cmH ₂ O/(L/s) |
| Compliance (C) | 0 ~ 300 ml/cmH ₂ O |
| Elasticity (E) | 0.003 to 10 hPa/mL(cmH ₂ O/mL) |

Control Accuracy

| | |
|------------------|------------------------------------------------------------------------------------------------------------------------------------------|
| Volume delivery | 5 mL to 60 mL: ±10 mL 60 mL to 210 mL: ±15 mL 210 mL to 1500 mL: ± 7% of the set value |
| Pressure | Pinsp, Plimit, ΔPsupp, ΔPapnea ± 2.5 cmH ₂ O or ± 7% of the set value, whichever is greater |
| PEEP | OFF: ± 3.0 cmH ₂ O 3 to 30 cmH ₂ O: ± 2.0 cmH ₂ O, or ± 8% of the set value, whichever is greater |
| Rate | ± 1bpm or ± 10% of the set value, whichever is greater |
| I:E | 2:1 to 1:4: ± 10% of the set value Other range: ± 25% of the set value |
| Tip:Ti | ± 8% |
| Tinsp | ± 0.2s |
| Trigger Window | ± 10% |
| Flow Trigger | ± 1L/min |
| Pressure Trigger | ± 2cmH ₂ O |
| Exp% | ± 10% |

Monitoring Accuracy

| | |
|---------------------|------------------------------------------------------------------------------------------|
| Volume monitoring | 0 to 60 mL: ± 10 mL 60 to 210 mL: ± 15 mL 210 to 3000 mL: ± 7% of the real reading |
| Pressure monitoring | ± 2.0 cmH ₂ O or ± 4% of the real reading, whichever is greater |
| Rate | ± 1bpm or ± 5% of the real reading, whichever is greater |

| | |
|------------------|-----------------------------------------------------------------|
| I:E | 2:1 to 1:4: ± 10% of the reading Other range: no defined. |
| MV | ± 0.1L/min or ± 8% of the real reading, whichever is greater |
| O2 concentration | ± (2.5% of volume percentage + 2.5% of gas concentration) |

Trend Graph

Continuous trend information with time discrete events for the latest 48 hours

Trend Table

Continuous trend information together with time discrete events for the latest 48 hours

Alarm Log Book

500 events storage, first in first out

Alarm setting

| | |
|----------------------------------|--------------------------------------------------------------------------|
| Tidal volume | Low: 0 ~ 1595 ml High: 5 ~ 1600 ml |
| Minute volume | Low: 0 ~ 99 L/min High: 0.2 ~ 100 L/min |
| Inspired oxygen | Low: 18% ~ 98% High: OFF, 20% ~ 100% |
| Apnea alarm | VTe < 10ml measured in 20s Paw < (PEEP + 3) cmH ₂ O in 20s |
| Airway pressure low | 0 ~ 98 cmH ₂ O |
| Airway pressure high | 2 ~ 100 cmH ₂ O |
| Sustained airway pressure alarm: | 15s |
| Subatmospheric pressure alarm: | Paw < -10 cmH ₂ O |
| Alarm silence countdown timer: | 120 to 0 seconds |

Lung Recruitment Tool

| | |
|------------------------|---------------------------------------------------------------------------------------------------------------------|
| Maneuver | Multi-Step and One-Step Recruitment |
| One-Step Recruitment | Pressure Hold: 20 to 60 cmH ₂ O Hold Time: 10 to 40s PEEP on Exit: Off, 3 to 30 cmH ₂ O |
| Multi-Step Recruitment | Increasing PEEP progressively (with a maximum of 7 stages) |

Ventilator Components

Flow Sensor

| | |
|----------|---------------------------------|
| Type | Variable orifice flow sensor |
| Location | Inspiratory and expiratory port |

Oxygen Sensor

| | |
|----------------------------|------------------------------------------------|
| Type | Galvanic fuel cell |
| FiO ₂ displayed | 18% to 100% |
| Accuracy | ± (volume fraction of 2.5 % + 2.5 % gas level) |
| Response Time | ≤ 20 seconds |

Ventilator Screen

| | |
|--------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Display type | Color capacitive touch screen |
| Display size | 15 inch |
| Pixel format | 1024 x 768 |
| Brightness | Adjustable |
| Screen display | configurable |
| Display parameters | All setting and alarm parameters (including Breath rate, I/E ratio, Tidal volume, Minute volume, PEEP, MEAN, PEAK, PLAT, and O ₂ concentration, EtCO ₂ , N ₂ O, Anesthesia gas concentration, BIS) |
| Display waveforms | P-T, F-T, V-T, CO ₂ , BIS, O ₂ , Anesthetic gas, N ₂ O |
| Spirometry loops | P-V, F-V and F-P |
| Timer | On screen timer |

Communication Ports

One RS-232C connector and one DB9 connector
Ethernet (RJ-45)
USB
VGA

Vaporizers

| | |
|----------------|---------------------------------------------------------------------------------|
| Vaporizer | Mindray V60 Anesthetic Vaporizer or Penlon Sigma Delta Anesthetic Vaporizer |
| Support agents | Halothane, Enflurane, Isoflurane, Sevoflurane |
| Position | MAX.2 |
| Mounting mode | Selectatec®, with interlocking function Plug-in®, with interlocking function |

Modules

Anesthesia Gas (AG) Module

| | |
|---------------|--------------------------------------------------------------------------------------------------------------------------------------------------|
| Monitor gases | CO ₂ , N ₂ O, Halothane, Enflurane, Isoflurane, Sevoflurane, Desflurane, MAC, Paramagnetic O ₂ (optional) |
| Warm-up time | 45 s (ISO accuracy mode) 10min (full accuracy mode) |
| Pump rate | Adu/Ped: 150, 180, 200 ml/min Neo: 100, 110, 120 ml/min |
| Range | CO ₂ : 0% ~ 10% Des: 0% ~ 18 % Sev: 0% ~ 8% Enf, Iso, Hal: 0% ~ 5% O ₂ /N ₂ O: 0% ~ 100% |

Carbon Dioxide (CO₂) Modules

| | |
|--------------------|-----------------------------------------------------------------------------------|
| Method | Infrared absorption |
| Module type | Mindray side-stream Capnostat mainstream Oridion micro-stream (optional) |
| Work mode | Standby or measurement |
| Displayed numerics | EtCO ₂ , FiCO ₂ |
| Waveform | Capnography |

Side-Stream Carbon Dioxide (CO₂) Module

| | |
|-------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Measurement range | 0 ~ 152 mmHg |
| Accuracy | ± 2 mmHg (0 ~ 40 mmHg) ± 5% of the real reading (41 ~ 76 mmHg) ± 10% of the real reading (77 ~ 152 mmHg) |
| Resolution | 1 mmHg |
| Pump rate | Neonatal: 100 mL/min or 120 mL/min Adult/children: 120 mL/min or 150 mL/min |
| Warming-up time | < 1 min, enter the ISO accuracy mode After 1 min, enters the full accuracy mode |
| Response time | < 5 s @ 100 mL/min < 5 s @ 120 mL/min Measured by using neonatal watertrap and 2.5 m neonatal sampling line < 6.5 s @ 120 mL/min < 6 s @ 150 mL/min Measured by using adult watertrap and 2.5 m adult sampling line |

Mainstream CO₂ Module

| | |
|-------------------|----------------------------------------------------------------------------------------------------------------------------------------------|
| Measurement range | 0 ~ 150 mmHg |
| Accuracy | ± 2 mmHg (0 ~ 40 mmHg) ± 5% of the reading (41 ~ 70 mmHg) ± 8% of the reading (71 ~ 100 mmHg) ± 10% of the reading (101 ~ 150 mmHg) |
| Resolution | 1 mmHg |
| Response time | < 2 s |
| Alarm limit | EtCO ₂ High: OFF, 2 ~ 150 mmHg EtCO ₂ Low: OFF, 0 ~ 148 mmHg FiCO ₂ High: OFF, 1 ~ 150 mmHg |

Micro-stream CO₂ Module

| | |
|-------------------|----------------------------------------------------------------------------------------------------------|
| Measurement range | 0 ~ 99 mmHg |
| Accuracy | 0 ~ 38 mmHg: ± 2 mmHg 39 ~ 99 mmHg: ± (5 % of the reading + 0.08 % of (the reading minus 38 mmHg)) |

| | |
|---------------------|---------------------------------------------------------------------------------------------------------------------------------|
| Sampling rate | 50 ml/min |
| Sampling accuracy | -7.5 ml/min ~ + 15 ml/min |
| Initialization time | 30s |
| Response time | ≤ 2.9s |
| Rising time | ≤ 190 ms |
| Alarm range | EtCO ₂ High: OFF, 2 ~ 99 mmHg EtCO ₂ Low: OFF, 0 ~ 97 mmHg FiCO ₂ High: OFF, 1 ~ 99 mmHg |

BIS Module

| | |
|-----------------------|-----------------------------------------------------------------------------------------------------------------------------------------|
| Measured parameters | EEG |
| BIS/BIS L, BIS R | 0 ~ 100 |
| Sweep speed | 6.25 mm/s, 12.5 mm/s, 25 mm/s or 50 mm/s |
| Alarm limit | BIS high: 2 ~ 100 BIS low: 0 ~ 98 |
| Calculated parameters | SQI/SQI L, SQI R; EMG/EMG L, EMG R; SR/SR L, SR R; SEF/SEF L, SEF R; TP/TP L, TP R; BC/BC L, BC R; sBIS L, sBIS R; sEMG L, sEMG R; ASYM |

Agent Consumption Calculation

| | |
|-------------------|-----------------------------------------------------------|
| Calculation range | 0 to 3000 ml |
| Accuracy | ± 2 mL, or ± 25% of the real reading, whichever is larger |

Electrical Specifications

Current Leakage

| | |
|------------|----------|
| 100 ~ 240V | < 500 µA |
|------------|----------|

Power and Battery Backup

| | |
|------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------|
| Power input | 220-240 V, 50/60 Hz, 6A 100-120 V, 50/60 Hz, 7A 100-240 V, 50/60 Hz, 7A |
| Auxiliary electrical outlets | Up to 4 outlets (3A for each, total 5A) |
| Battery backup | 90 minutes in case of one battery or 240 minutes in case of two batteries (powered by new fully-charged batteries with 25°C ambient temperature) |
| Battery type | Build-in Li-ion battery, 10.95 VDC, 4500 mAh |
| Safety feature | In case of electricity and battery failure, manual ventilation, gas delivery and agent delivery are possible |

Pneumatic Specifications

ACGO (Auxiliary Common Gas Outlet, Integrated)

| | |
|-----------|---------------------------|
| Connector | ISO 22 mm OD and 15 mm ID |
|-----------|---------------------------|

Pipeline Supply

| | |
|----------------------|-------------------------------------------|
| Gas type | O ₂ , N ₂ O and Air |
| Pipeline input range | 280 to 600 kPa |
| Pipeline connections | DISS or NIST |

Pipeline Supply Pressure Gauges

| | |
|--------------|-------------------------------------------------------------|
| Display type | Mechanical |
| Ranges | 0 to 1000kPa |
| Accuracy | ± (4% of the full scale reading + 8% of the actual reading) |

Cylinder Supply

| | |
|------------------------------|-----------------------------------------|
| Cylinder Supply | E Cylinder (American style or UK style) |
| O ₂ Input Range | 6.9 to 20 MPa |
| N ₂ O Input Range | 4.2 to 6 MPa |
| Air Input Range | 6.9 to 20 MPa |
| Cylinder Connections | Pin-Index Safety System (PISS) |
| Yoke Configuration | O ₂ , N ₂ O, Air |

Cylinder Supply Pressure Gauges

| | |
|------------------------|-----------------------------------------------------------|
| Display type | Mechanical |
| Air Range | 0 to 25 MPa |
| O ₂ Range | 0 to 25 MPa |
| N ₂ O Range | 0 to 10 MPa |
| Accuracy | ± (4% of the full scale reading+8% of the actual reading) |

O₂ Controls

| | |
|----------------------|----------------------------------------------------------------|
| Method | N ₂ O shut off with loss of O ₂ pressure |
| Supply failure alarm | ≤ 220.6 kPa ± 34.2kPa |
| O ₂ Flush | 25 ~ 75 L/min |

O₂-N₂O Link system

| | |
|-------|-------------------------------------------------|
| Type | Mechanical |
| Range | O ₂ concentration not lower than 25% |

Auxiliary O₂ Flowmeter

| | |
|-----------|--------------|
| Range | 0 ~ 15 L/min |
| Indicator | Flow tube |

Electronic Flow Meters

| | |
|-----------------------------|----------------------------------------------------------------------------------------------------------------------|
| O ₂ flow range | 0 to 15 L/min |
| Air flow range | 0 to 15 L/min |
| N ₂ O flow range | 0 to 10 L/min |
| Accuracy | between -10% and +10% of the indicated value (under 20°C and 101.3 kPa, for flow between 10% and 100% of full scale) |

Optimizer

Only available when AG or CO₂ Module is loaded

Environmental Specifications

Operating

| | |
|-------------------|---------------------------|
| Temperature | 10 ~ 40°C |
| Relative humidity | 15% ~ 95% (noncondensing) |
| Barometric (Kpa) | 70 ~ 106 kPa |

Storage

| | |
|-------------------|-------------------------------------------------------------------|
| Temperature | -20 ~ 60°C for main unit, -20 ~ 50°C for O ₂ sensor |
| Relative humidity | 10% ~ 95% (noncondensing) |
| Barometric | 50 ~ 106 kPa |

Electromagnetic Compatibility

| | |
|-----------|-------------------------------------------------|
| Immunity | Complies with all requirements of IEC 60601-1-2 |
| Emissions | Complies with all requirements of IEC 60601-1-2 |

Breathing System Specification

Breathing system volume (Pre-pak)

| | |
|-----------------------|---------|
| Automatic ventilation | 2850 ml |
| Manual ventilation | 1800 ml |

Breathing system volume (Non Pre-pak)

| | |
|-----------------------|---------|
| Automatic ventilation | 2600 ml |
| Manual ventilation | 1800 ml |

System Components

| | |
|---------------------------------------|-----------------------------|
| Carbon dioxide absorbent canister | Absorbent capacity: 1500 mL |
| Integrated expiratory limb water trap | Capacity: 6 mL |

Breathing Circuit Parameters

| | |
|------------------------|-------------------------------------------------------------------------------------------------------------------------------|
| System leakage | ≤ 60 mL/min at 3 kPa |
| Compliance | ≤4 mL/100Pa (Manual mode) Automatically compensates for compression losses within the breathing circuit in mechanical mode |
| Expiration resistance | < 6.0 cm H ₂ O @60 L/min |
| Inspiration resistance | < 6.0 cm H ₂ O @60 L/min |

System Pressure Gauge

| | |
|----------|-------------------------------------------------------------|
| Range | -20 ~ 100 cmH ₂ O |
| Accuracy | ± (2% of the full scale reading + 4% of the actual reading) |

Ports and Connectors

| | |
|-----------------|-----------------------------|
| Exhalation | 22 mm OD / 15 mm ID conical |
| Inhalation | 22 mm OD /15 mm ID conical |
| Manual bag port | 22 mm OD /15 mm ID conical |

Bag-to-Ventilator Switch

| | |
|------|-----------|
| Type | Bi-stable |
|------|-----------|

| | |
|---------|--------------------------------------------------|
| Control | Switch between manual and mechanical ventilation |
|---------|--------------------------------------------------|

Integrated Adjustable Pressure Limiting (APL) Valve

| | |
|--------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------|
| Range | SP, 5 ~ 70 cmH ₂ O |
| Tactile knob indication at above 30 cmH ₂ O | |
| Accuracy | ± 3 cmH ₂ O or ± 15% of the setting value, which is greater, but is not more than + 10 cmH ₂ O |

Anesthetic Gas Scavenging System (AGSS)

| | |
|-------------------------|--------------------------------------------------------|
| Size (H x W x D) | 430 x 132 x 114 mm |
| Type of disposal system | Active: High-flow or Low-flow Passive |
| Applicable standard | ISO 80601-2-13 |
| Pump rate | 75 ~ 105 L/min (High-flow) 25 ~ 50 L/min (Low-flow) |

Pressure relief device: Pressure compensation opening to the air
State indication of the disposal system: The float falls below the "MIN" mark on the sight glass when the disposal system does not work or the pump rate is lower than 25 L/min (Low-flow) or 75 L/min (high-flow).
Connector of the disposal system: ISO 9170-2

Materials

All materials in contact with exhaled patient gases are autoclavable, except flow sensors (being not capable of being autoclaved), O₂ sensor, and mechanical pressure gauge.
All materials in contact with patient gas are latex free.

Suction Device

Venturi Suction Regulator

| | |
|----------------|-----------------------------------------------------------------------------------------|
| Gas source | Air, from system gas source |
| Minimum flow | 20 L/min |
| Maximum vacuum | ≥72 kPa at supply gas pressure of 280 kPa; ≥73 kPa at supply gas pressure of 600 kPa |

Continuous Suction Regulator

| | |
|----------------|-----------------------------------------------------------------------------------------------------------|
| Supply | Negative Pressure Suction |
| Maximum vacuum | 517.5 mmHg to 540 mmHg (69 kPa to 72 kPa) with external vacuum applied of 540 mmHg and 40 L/min free flow |
| Maximum flow | 39 L/min to 40 L/min with external vacuum applied of 540mmHg and 40 L/min free flow |
| Minimum flow | 20 L/min |

Please contact your local Mindray sales representative for the most current information.

www.mindray.com

P/N: ENG-WATO EX-65 datasheet-210285X4P-20210330

©2021 Shenzhen Mindray Bio-Medical Electronics Co.,Ltd. All rights reserved.

mindray
healthcare within reach