














VG70-Service kits

No.	Part Name	Remarks	Part Number	Photo of Appearance	Quantity / Sets	Replacement Period
1	Expiratory diaphragm	Used to the expiratory valve	130000231		5	Need to be changed in every 20 times of disinfection
2	O-Ring 16*1.8 Seal Ring for Exp. Valve	Used to the expiratory valve	304000099		5	Need to be changed in every 20 times of disinfection
3	O-Ring 20*1.8 Seal Ring for Exp. Valve	Used to the expiratory valve	304000062		5	Need to be changed in every 20 times of disinfection
4	Sump tray	Used to the expiratory valve	130009661		5	Need to be changed in every 20 times of disinfection
5	PTFE Seal Ring	Used to the expiratory valve	130012614		5	Need to be changed in every 20 times of disinfection
6	Rubber o-sealing ring for 29x1.8	Used to the expiratory valve	304000105		5	Need to be changed in every 20 times of disinfection
7	Filter Element	Installed to the inhalation air port	130004358		4	3 months or as needed
8	O-ring Seal Ring 23.6*1.8 for Inspiratory Port	Installed to inspiratory port	304000024		2	A half year or as needed
9	One-way diaphragm	Installed to expiratory port	130001347		5	One years or as needed
10	Fan Filter	Installed to cooling fan	130010423		5	One years or as needed
11	Water trap	Used to the expiratory valve	130009651		3	One year or as needed
12	Filter	Installed the gas supply inlet	130003930		2	One year or as needed
13	Gas Supply Inlet filter	Installed the gas supply inlet	230000135		4	One year or as needed



Certificate

No. Q5 065725 0022 Rev. 02

Holder of Certificate:

Beijing Aeonmed Co., Ltd.
Room 405
Basement 1 to 4th Floor of 901 Unit
Building 9, No.26 Outer Ring West Road
Fengtai District
100070 Beijing
PEOPLE'S REPUBLIC OF CHINA

Certification Mark:



Scope of Certificate:

Design and Development, Production, Distribution, Installation and Servicing of Anaesthetic Workstation, Vaporizer, Ventilator, Medical Air Compressor, Infusion Pump, Ceiling Pendant, Operating Table, Surgical Light, Multi-Parameter Patient Monitor, Syringe Pump, Patient Warming System, Videoscope System.

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). See also notes overleaf.

Report No.: BJ1985904

Valid from: 2020-03-23

Valid until: 2022-12-31

Date, 2020-03-23

Christoph Dicks
Head of Certification/Notified Body

ZERTIFIKAT ◆ CERTIFICATE ◆ CERTIFICADO ◆ CERTIFIKAT ◆ 認證證書 ◆ CERTIFICATE ◆ CERTIFICATE

Certificate

No. Q5 065725 0022 Rev. 02

Applied Standard(s):

EN ISO 13485:2016
Medical devices - Quality management systems -
Requirements for regulatory purposes
(ISO 13485:2016)
DIN EN ISO 13485:2016

Facility(ies):

Beijing Aeonmed Co., Ltd.
Room 405, Basement 1 to 4th Floor of 901 Unit, Building 9, No.26
Outer Ring West Road, Fengtai District, 100070 Beijing, PEOPLE'S
REPUBLIC OF CHINA

Beijing Aeonmed Co.,Ltd.
No. 10 Chaobai Street, Yingbin Road West, Yanjiao Development
Zone, 065201 Langfang City, Hebei Province, PEOPLE'S
REPUBLIC OF CHINA

ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ CERTIFICADO ◆ CERTIFICAT



EC Certificate

Full Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)

(Devices in Class IIa, IIb or III)

No. G1 065725 0019 Rev. 04

Manufacturer:

Beijing Aeonmed Co., Ltd.

Room 405

Basement 1 to 4th Floor of 901 Unit

Building 9, No.26 Outer Ring West Road

Fengtai District

100070 Beijing

PEOPLE'S REPUBLIC OF CHINA

**Product Category(ies): Anaesthetic Workstation, Vaporizer,
Ventilator, Medical Air Compressor,
Infusion Pump, Ceiling Pendant,
Multi-Parameter Patient Monitor,
Videoscope System, Patient Warming System.**

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex II. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class III devices an additional Annex II (4) certificate is mandatory. All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:G10657250019Rev.04

Report No.: BJ19859071

Valid from: 2021-05-21

Valid until: 2024-05-26

Date, 2021-05-21

Christoph Dicks

Head of Certification/Notified Body