

# Certificate of Analysis

## CERTIFIED REFERENCE MATERIAL Organic substance

Ref No: SB39070.250MG

Lot No: XXXXXX

Certification Date: XXXXXX

Barcode: XXXXXXXXX

Description of the Reference Material (CRM): Sulfameter

CAS No: 651-06-9

Empirical formula:  $C_{11}H_{12}N_4O_3S$

MW: 280.303

Certified Purity/ Uncertainty: 99.9 +/- 0.1 %

Storage Conditions: Store in a refrigerator at temperatures between 2°C to 8°C

Expiry date: XXXXXXXXXXXX

Method of certification: CRM's calibration procedure (WQP 5.15.1/22)

The following methods of analysis are used to determine purity: HPLC (Purity=100% - Assay impurities)

### Analytical Data:

Column:	Chromolith® Performance RP-18e 4.6x100 mm, 2µm particle size	Gradient elution (time):	A%	B%
Mobile phase:	A: Water B: Acetonitrile	0	80	20
Flow rate:	2.0 ml/min	7	0	100
Column temperature:	20 °C	8	0	100
Injection volume:	10µl	9	80	20
Detector	DAD	11	80	20

### Concept of Certification and traceability statement:

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA 4/02.

Metrological traceability is established through in-house validated method.

The measurement results are traceable to SI.

### Intended use:

#### For Laboratory Use Only

This CRM is intended for:



CPAchem Ltd. is ISO 17034 (Cert No AR-1835)  
and ISO/IEC 17025 (Cert No AT-1836) accredited by ANAB

Calibration of TLC, GC/FID, GC/TCD, GC/ECD, GC/MS, GC/MS/MS, LC/UV, LC/MS and LC/MS/MS  
Validation of analytical methods  
Preparation of "working reference samples"  
Detection limit and linearity studies

This statement is not intended to restrict the use for other purposes.

**Instructions for the correct use of this reference material:**

This CRM can be used directly or can be diluted in an appropriate solvent. Only a clean glassware should be used.

**Stability and storage:**

This CRM is with a guaranteed purity +/- 2% deviation prior to the expiration date. Stability is guaranteed, provided that the material is kept in its original packaging, tightly closed stored, as written in the section: Storage Conditions.

**Hazardous situation:**

The normal laboratory safety precautions should be observed when working with this CRM. Further details for the handling of this chemical are available as safety data sheet.

**Level of homogeneity**

The material was tested for homogeneity by analyzing randomly selected samples according to an in-house procedure. The level of homogeneity proved satisfactory for a sample volume of min. 2 mg. The uncertainty incorporates the sample standard deviation combined with the uncertainty calculated from homogeneity and stability studies.

This certificate relates solely to the lot number given above.

All processes (including generating of this certificate) are completely controlled by the specialized Computer-Aided-Manufacturing (CAM) software.

This Certified Reference Material was produced under a quality management system that is:

- Registered to ISO 9001 Quality Management System (Lloyd's Register Quality Assurance Ltd Cert No 0039638)
- Accredited according to ISO/IEC 17025
- Accredited according to ISO 17034

This document is designed and the certified value and uncertainty are determined in accordance with ISO Guide 31, ISO Guide 35, and Eurachem / CITAC Guides

**Names of certifying officers:**

Laboratory:  Margarita Dimitrova

Manager:  Krassimira Taralova

End

## Peak Integration Report

Sample Name:	Sulfameter_41485842	Inj. Vol.:	10.00
Instrument Method:	Chromolith short		
Inj. Date / Time:	14-Oct-2022 / 13:55	Run Time:	11.00

No.	Time min	Peak Name	Peak Type	Area mAU*min	Height mAU	Rel.Area %
2	1.56	Sulfameter	BMb*	51.225	488.977	99.87
<b>TOTAL:</b>				51.22	488.98	99.87

