Downloaded by, SUDARAT KETSIRIKUN on date 01 Apr 2021

Test Report No. 7191205139-EEC19/01-WBH dated 27 Feb 2019

Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.



Add value. Inspire trust.

SUBJECT:

Testing of Latex Powdered Glove submitted by Sri Trang Gloves (Thailand) Public Company Limited on 13 Feb 2019.

TESTED FOR:

Sri Trang Gloves (Thailand) Public Company Limited 110 Kanchnawanit Road., T.Patong, A.Hatyai, Songkhla, 90230, Thailand.

TEST DATE:

27 Feb 2019

DESCRIPTION OF SAMPLES:

S/N	Product Description	Colour	Lot No.	Size	Sample received (pieces)	Manufacturer	
1	Latex Powdered Glove – (012296388)	White	(see remark 1)	M	13	Sri Trang Gloves (Thailand) Public Company Limited	

Lot size as specified by client: 35,001 to 150,000 pieces

METHOD OF TEST:

As requested by client, the test was conducted as follows:

EN 455-2:2015 Medical gloves for single use Part 2: Requirements and testing for physical properties - Clause 4 Dimensions



Laboratory: TÜV SÜD PSB Pte. Ltd. No.1 Science Park Drive Singapore 118221 Phone: +65-6885 1333 Fax: +65-6776 8670 E-mail: enquiries@tuv-sud-psb.sg www.tuv-sud-psb.sg Co. Reg: 199002667R

Regional Head Office: TÜV SÜD Asia Pacific Pte. Ltd. 1 Science Park Drive, #02-01 Singapore 118221

Page 1 of 4

Test Report No. 7191205139-EEC19/01-WBH dated 27 Feb 2019



RESULTS:

Sample: Latex Powdered Glove - (012296388), Size M

Table: Results for EN 455-2:2015 Clause 4

Clause	Tests	Requirements	Number tested (pieces)	Results	Median	Inferred results
	Dimensions a) Length (mm)	Median Length for size M: ≥ 240	13	247 242 240 242 242 245 241 250 245 244 242 248	244	Passed
4	b) Width (mm)	Median Width for size M: 95 ± 10	T ₁₃	94 94 94 94 94 94 94 94 94 94 94 95	94	Passed

REMARKS:

1. The manufacturing batch code was not provided by the client.

Yeo Poh Kwang Associate Engineer

Wong Bee Hui Product Manager Medical Health Services (NAM)

Test Report No. 7191205139-EEC19/01-WBH dated 27 Feb 2019



APPENDIX:



Photo: Latex Powdered Glove - (012296388), Size M



Downloaded by , SUDARAT KETSIRIKUN on date 01 Apr 2021

Test Report No. 7191205139-EEC19/01-WBH dated 27 Feb 2019



Please note that this Report is issued under the following terms :

- 1. This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SÜD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
- The sample/s mentioned in this report is/are submitted/supplied/manufactured by the Client. TÜV SÜD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
- 3. Nothing in this report shall be interpreted to mean that TÜV SÜD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
- 4. This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSB or to the report or results furnished by TÜV SÜD PSB in any advertisements or sales promotion.
- 5. Unless otherwise stated, the tests were carried out in TÜV SÜD PSB Pte Ltd, No.1 Science Park Drive Singapore 118221.



July 2011