



奥美医疗用品股份有限公司检测中心
ALLMED MEDICAL PRODUCT Co., Ltd, TESTING CENTER

检测报告
Test Report

报告编号 Report No: TC2203030035

样品信息 Sample information:

样品名称 Sample name	纱布分切卷	产品号/料号 Product code	N/A
生产单位 Manufactory	国际业务事业群事业 1 部脱漂车间 MA1031E	订 单 号 Order #	N/A
规格 Specification	40Sx40S/30x30-187cm-18X-3 幅 (155 进*2+157 进)	批 号 Lot #	22022513-8713-1
测试方法 Test method	EN14079:2003	测试周期 Testing period	2022-03-07 ~ 2022-03-08

检测结果 Test Result:

项目 Item	单位 Unit	标准 Standard	结果 Results	结论 Conclusion
水中可溶物 Water-soluble substances	%	总量不大于 0.50% No more than 0.50%.	0.26	OK
醚中可溶物 Ether-soluble substances	%	总量不大于 0.50% No more than 0.50%.	0.20	OK
硫酸盐灰分 Sulphated ash	%	总量不大于 0.40% No more than 0.40%.	0.19	OK
淀粉和糊精 Starch and dextrin	\	溶液不应显蓝色、紫色、淡红色或淡棕色。 Solution shall not show any blue,violet,red dish or brownish colour.	Complies	OK
酸碱度 Acidity and alkalinity	\	两者均不得显粉红色。 No pink color develops in either portion.	Complies	OK
可浸提的着色物质 Extractable colouring matter	\	滤液的颜色应不深于对照液 Y5 或 GY6 或蓝色 对照溶液。 Liquid obtained shall not be more intensely coloured than reference solution Y5,GY6 or blue reference solution.	Complies	OK
外来纤维 Foreign fibres	\	可以偶尔有少量孤立的外来纤维存在。 A few isolated foreign fibres may also be present.	Complies	OK
纤维鉴别 Fibre Identification	\	棉纤维应符合鉴别试验 A,B 和 C 的要求。 Cotton fibres shall conform to the ident tests A, B and C.	Complies	OK

备注 Remarks:

接收标准: EN14079:2003

结论 Conclusion:

The inspection results for the inspected items are in accordance with the relative specifications in the inspection reference, and the corresponding performance of the inspected samples is qualified.



编制人/日期: 刘奇 2022.03.08
Edited by/ Date:

审核人/日期: 杨玉莹 2022.03.09
Checked by/Date:

批准人/日期: 杨梦菊 2022.03.09
Approved by/Date:

本报告仅对检测样品负责 This report is only responsible for the provided sample.

表格编号 Form#: LTR-001-001

表格批准人 Approved By: 戴虹梅