

BILIRUBIN (TOTAL)

COD 12506 5 x 50 mL

BILIRUBIN (TOTAL)
DICHLOROPHENYL DIAZONIUM

INTENDED USE

Reagent for the quantitative measurement of total bilirubin concentration in human serum for monitoring the evolution of jaundice in adult population. For in vitro professional use in the clinical laboratory only.

This reagent is for use in the BioSystems A25 and A15 analyzers or in other analyzer with similar performance characteristics.

CLINICAL SIGNIFICANCE

A number of inherited and acquired diseases affect production, uptake, metabolism, and excretion of bilirubin, resulting in hyperbilirubinemia^{1,3}.

Unconjugated hyperbilirubinemia is seen in increased red cell destruction (hemolytic anemia, extensive hematoma), in ineffective erythropoiesis and in some rare genetic diseases (Gilbert's syndrome, Crigler-Najjar syndrome).

Conjugated hyperbilirubinemia is associated to a decreased excretion of bile due to liver diseases (hepatitis or cirrhosis) or to intrahepatic or extrahepatic cholestasis.

Jaundice is a clinical manifestation of hyperbilirubinemia, consisting of deposition of bile pigments in the skin, resulting in a yellowish staining of the skin and mucous membranes.

Based on clinical guidelines and textbooks, and when used in conjunction with other diagnostic technologies and options, this medical information is useful for the assessment of bilirubin variations. Clinical diagnosis should not be made on the findings of a single test result, but should integrate both clinical and laboratory data.

PRINCIPLE OF THE METHOD

Direct bilirubin in the sample reacts with 3,5-dichlorophenyl diazonium salt forming a coloured complex that can be measured by spectrophotometry at 535 nm⁴. Both direct and indirect bilirubin couple with diazo in the presence of cetrimide^{5,6}. The terms "direct" and "total" refer to the reaction characteristics of serum bilirubin in the absence or presence of solubilizing (accelerating) reagents. The "direct" and "indirect" bilirubin are only approximately equivalent to the conjugated and unconjugated fractions.

CONTENTS AND COMPOSITION

- A. Reagent: 5 x 40 mL. Phosphoric acid 188 mmol/L, cetrimide 40 mmol/L, pH 0,9.
DANGER: H314: Causes severe skin burns and eye damage. P260: Do not breathe dust/fume/gas/mist/vapours/spray. P280: Wear protective gloves/protective clothing/eye protection/face protection. P303+P361+P353: IF ON SKIN (or hair): Remove/Take off immediately all contaminated clothing. Rinse skin with water/shower. P305+P351+P338: IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.
- B. Reagent: 5 x 10 mL. 3,5-dichlorophenyl diazonium 1,5 mmol/L.
DANGER: H314: Causes severe skin burns and eye damage. P260: Do not breathe dust/fume/gas/mist/vapours/spray. P280: Wear protective gloves/protective clothing/eye protection/face protection. P303+P361+P353: IF ON SKIN (or hair): Remove/Take off immediately all contaminated clothing. Rinse skin with water/shower. P305+P351+P338: IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

For further warnings and precautions, see the product safety data sheet (SDS).

STORAGE AND STABILITY

Store at 2-8 °C.

Components are stable once opened until the expiry date marked in the label if they are stored at the recommended temperature, well closed and care is taken to prevent contamination during their use.

On board stability: Reagents open and kept in the refrigerated compartment of the analyzer are stable 3 months.

Indications of deterioration: Absorbance of the blank over the limit indicated in "Test Parameters".

ADDITIONAL MATERIALS REQUIRED (NOT PROVIDED)

Biochemistry Calibrator (BioSystems cod. 18011) or Biochemistry Calibrator Human (BioSystems cod. 18044).

REAGENT PREPARATION

Reagents are provided ready to use.

SAMPLES

Serum and plasma collected by standard procedures. Heparin and EDTA may be used as anticoagulants.

Bilirubin concentration in serum and plasma is stable for 1 day at 20-25°C, 7 days at 4-8°C and 6 months at -20°C if protected from light⁷.

CALIBRATION

A reagent blank should be done every day and a calibration at least every 3 months, after reagent lot change or as required by quality control procedures.

QUALITY CONTROL

It is recommended to use the Biochemistry Control Serum level I (cod. 18005, 18009 and 18042) and II (cod. 18007, 18010 and 18043) to verify the accuracy of the measurement procedure.

Each laboratory should establish its own internal Quality Control scheme and procedures for corrective action if controls do not recover within the acceptable tolerances.

REFERENCE VALUES

Adults¹:

Total:	Up to 1.2 mg/dL = 21 µmol/L
--------	-----------------------------

Newborns²:

Age	premature	full-term
Up to 24 h	1.0-8.0 mg/dL = 17-137 µmol/L	2.0-6.0 mg/dL = 34-103 µmol/L
Up to 48 h	6.0-12.0 mg/dL = 103-205 µmol/L	6.0-10 mg/dL = 103-171 µmol/L
3-5 days	10-14 mg/dL = 171-239 µmol/L	4.0-8.0 mg/dL = 68-137 µmol/L

These ranges are given for orientation only; each laboratory should establish its own reference ranges.

METROLOGICAL CHARACTERISTICS

– The metrological characteristics described below have been obtained using an A25 analyzer and following the guidelines of the Clinical & Laboratory Standards Institute (CLSI). Results are similar with A15.

– Detection limit: 0.172 mg/dL = 2.95 µmol/L.

– Linearity limit: 38 mg/dL = 650 µmol/L.

– Precision:

Mean concentration	Repeatability (CV)	Within-laboratory (CV)
2.02 mg/dL = 34.6 µmol/L	2.2 %	4.4 %
4.49 mg/dL = 76.8 µmol/L	1.4 %	2.8 %

– Trueness: Results obtained with this reagent did not show systematic differences when compared with reference reagents. Details of the comparison experiments are available on request.

LIMITATIONS OF THE PROCEDURE

– Interferences: Hemolysis (hemoglobin 250 mg/dL) do not interfere. Lipemia (triglycerides 1300 mg/dL) interfere. Other drugs and substances may interfere⁸.

BIBLIOGRAPHY

1. Rifai N, Chiu RWK, Young I, Burnham CD, Wittwer CT, Tietz Textbook of Laboratory Medicine. 7th ed. Elsevier, 2023.
2. Tietz Textbook of Clinical Chemistry and Molecular Diagnostics, 4th ed. Burtis CA, Ashwood ER, Bruns DE. WB Saunders Co, 2005.
3. Friedman and Young. Effects of disease on clinical laboratory tests, 4th ed. AACC Press, 2001.
4. Thaler M, Luppa PB and Schlebusch H. Bilirubin measurement – an updated survey. J Lab Med 2008; 32:1-9.
5. Zoppi F, Peracino A, Fenili D, Marcovina S and Ramella C. Metodo per la determinazione della bilirubina totale e coniugata. Uso di un tensioattivo cationico come agente solubilizzante. Giorn It Chim CI 1976; 1:343-359.
6. Pearman FC and Lee RTY. Detection and measurement of total bilirubin in serum, with use of surfactants as solubilizing agents. Clin Chem 1974; 20: 447-453.
7. WHO/DIL/LAB/99.1, Rev.2; 2002.
8. Young DS. Effects of drugs on clinical laboratory tests, 5th ed. AACC Press, 2000.

TEST PARAMETERS

R1: Use Reagent A

R2: Use Reagent B

	A25	A15
GENERAL		
Name	BILIRUBIN TOTAL	BILIRUBIN TOTAL
Sample type	serum / plasma	serum / plasma
Analysis mode	bireagent differential	bireagent differential
Units	mg/dL	mg/dL
Turbidimetry test	no	no
Decimals	2	2
Type of reaction	increasing	increasing
PROCEDURE		
Reading mode	bichromatic	bichromatic
Main filter	535	535
Reference filter	670	670
Sample	6	6
Vol. R1	240	240
Vol. R2	60	60
Washing	1.2	1.2
Reading 1 (cycle)	5	5
Reading 2 (cycle)	27	19
Reagent 2 (cycle)	7	7
Predilution factor	-	-
CALIBRATION AND BLANK		
Calibration type	multiple	multiple
Number of calibrators	1	1
Calibration curve	-	-
OPTIONS		
Blank absorbance limit	0.050	0.050
Kinetic blank limit	-	-
Linearity limit	38	38
Substrate depletion	-	-

