

CERTIFICATE

Issued to:
Applicant:
ESPECIALIDADES MÉDICAS MYR, S.L.
C/ Lleida Num. 17-23 Pol. Ind. L'Empalme
43712 Llorenc Del Penedes, Spain

Licensee:
ESPECIALIDADES MÉDICAS MYR, S.L.
C/ Lleida Num. 17-23 Pol. Ind. L'Empalme
43712 Llorenc Del Penedes, Spain

Product : Microtome
Trade name(s) : MYR
Type(s)/model(s) : M-240 and M-250

The product and any acceptable variation thereto as specified in the Annex to this certificate and the documents therein referred to.

DEKRA hereby declares that the above-mentioned product has been certified on the basis of:

- an evaluation according to the standard(s) CAN/CSA C22.2 No. 61010-1:2012/A1:2018 and UL 61010-1:2012
- an inspection of the factory location according to DEKRA's North American Certification Scheme
- a DEKRA certification agreement with the number 6063705


DEKRA hereby grants the right to use the DEKRA certification mark as shown below.

The DEKRA certification mark may be applied to the product as specified in this certificate for the duration and under the conditions of the DEKRA certification agreement.

This certificate is issued on 28 March 2023 and expires upon withdrawal of one of the above mentioned standards.

Certificate number: 71-127714

DEKRA Certification B.V.



B.T.M. Holtus
Managing Director



H.R.M. Barends
Certification Manager

© Integral publication of this certificate is allowed

ACCREDITED BY THE
STANDARDS COUNCIL
OF CANADA



SPECIFICATION OF THE CERTIFIED PRODUCT**Product data**

Product : Microtome
Trade name(s) : MYR
Type(s)/model(s) : M-240 and M-250
Voltage : 100-240 VAC
Frequency : 50/60 Hz

Product data – type M-240

Power consumption : 40 VA

Product data – type M-250

Power Consumption : 100VA

TESTS**Test requirements**

CAN/CSA C22.2 No. 61010-1:2012/A1:2018
UL 61010-1:2012

Test result

The test results are laid down in DEKRA test file 226231400.

Additional information

In addition to the standards mentioned above, the products were investigated and found in compliance with the following standards:

UL 61010-1, Edition 3, revision date 2019-07-19
CAN/CSA-C22.2 NO. 61010-1-12/A1:18 (R2022)
UL-61010-2-101, Edition 3, revision date 2019-07-31
CAN/CSA-C22.2 NO. 61010-2-101:19

The list of components is laid down in test report 2262314.52B.

Conclusion

The examination proved that all requirements were met.

Factory location

ESPECIALIDADES MÉDICAS MYR, S.L.
C/ Lleida Num. 17-23 Pol. Ind. L'Empalme
43712 Llorenc Del Penedes, Spain