

Declaration of Conformity



Manufacturer: Shenzhen Mindray Bio-Medical Electronics Co., Ltd.
Mindray Building, Keji 12th Road South, Hi-tech Industrial
Park, Nanshan, Shenzhen, 518057, P. R. China

EC-Representative: Shanghai International Holding Corp. GmbH (Europe)
Eiffestraße 80
20537 Hamburg, Germany

Product: Hematology Control

Model: BC-3D

Applied Hematology BC-3200, BC-3000CT, BC-3000 Plus, BC-2800,

Analyzer Model: BC-2600, BC-2900, BC-1800, BC-2300, BC-2100
BC-3600 ,BC-3300, BC-30s, BC-31s, BC-20s
BC-21s, BC-30, BC-31

Classification: The device not in IVDD annex II and not for self
testing/performance evaluation

Conformity Assessment Route: IVDD Annex III(excluding Section 6)

We herewith declare that the above mentioned products meet the provisions of the Directive 98/79/EC on In Vitro Diagnostic Medical Devices. All supporting documentations are retained under the premises of the manufacturer.

Standards Applied:

List of (harmonized) standards for which documented evidence for compliance can be provided as attachment.

Start of CE-Marking: 2008-04-29

Place, Date of Issue: Shenzhen, 2017-12-18

Signature:

Name of Authorized Signatory: Mr.WangXinBing

Position Held in Company: Manager ,Technical Regulation

Applied Standards List

Applied Standards:

| | |
|----------------------------|--|
| EN ISO 18113-1:2011 | In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 1: Terms, definitions and general requirements |
| EN ISO 18113-2:2011 | In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 2: In vitro diagnostic reagents for professional use |
| EN ISO15223-1:2012 | Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied —Part 1: General requirements |
| EN 13612:2002 | Performance evaluation of in vitro diagnostic medical devices |
| EN 13640: 2002 | Stability testing of in vitro diagnostic reagents |
| EN 13641: 2002 | Elimination or reduction of risk of infection related to in vitro diagnostic medical devices |
| ISO14971: 2012 | Medical devices – Application of risk management to medical devices |