

## Power SYBR Green PCR Master Mix 5mL

Product No. **4367659**  
 Lot No. **2601676**  
 Date of Manufacture **02FEB2026**  
 Expiration Date **13NOV2026**

TEST	RESULT
------	--------

### ABI Prism SDS Functional Test

The Power SYBR Green PCR Master Mix is tested for functional performance on an Applied Biosystems Sequence Detection System, using the B-Actin Control Reagents and human genomic DNA as set forth:

At least 3 of 4 NTC (No Template Control) reactions must yield Ct values = > 35	Pass
Average Ct for 10 copies of human genomic DNA must be = < 35	Pass
The correlation coefficient for the standard curve must be = > 0.99	Pass
The slope of the standard curve must be between -3.2 and -3.8	Pass

### Physical Inspection

Correct label and lot number	Pass
Caps are tight with no signs of leakage.	Pass

***For Research Use Only.  
 Not for use in diagnostic procedures.***

***Manufactured under a Quality Management System that meets the requirements of ISO 13485 and ISO 9001. Site: Warrington, UK***

Corporate Legal Entity  
 Life Technologies  
 5781 Van Allen Way  
 Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
 For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

**Electronically generated, no signature  
 required  
 Issued 10FEB2026**