

**ACON BIOTECH (HANGZHOU) CO.,LTD**  
**CERTIFICATE OF ANALYSIS**

**PRODUCT NAME:** Hb Hemoglobin Test Strips

**PURCHASE NO.:** 37740068A

**BATCH NO.:** Hb0070049

**EXPIRATION DATE:** 2022-07-18

**QUANTITY:** 4,000pcs

ANALYSIS			CRITERION	TEST RESULT
SYSTEM ACCURACY (Bias with manufacture control)	Level 0	4.5-5.5g/dL	$\leq \pm 1.0\text{g/dL}$	Complies
	Level 1	6~10g/dL or 10~25.6g/dL	$\leq \pm 0.4\text{g/dL}$ or $\leq \pm 4\%$	Complies
	Level 2	10~25.6g/dL	$\leq \pm 4\%$	Complies
	Level 3	10~25.6g/dL	$\leq \pm 4\%$	Complies
SYSTEM PRECISION	Level 0	4.5-5.5g/dL	$SD \leq 0.8 \text{ g/dL}$	Complies
	Level 1	6~10g/dL or 10~25.6g/dL	$SD \leq 0.4 \text{ g/dL}$ or $CV \leq 3\%$	Complies
	Level 2	10~25.6g/dL	$CV \leq 3\%$	Complies
	Level 3	10~25.6g/dL	$CV \leq 3\%$	Complies

Quality Department: Hedy Zhang

Date: Sep. 23, 2020