

Polypropylene Mesh

The monofilament Polypropylene Mesh is bio-compatible, non-absorbable, sterile and porous material. Knitted to provide single directional elasticity while allowing the mesh to be cut for any shape without unravelling structure.

Fabric has soft design flexible, strong with permeable pore use in hernia repair as well as other soft tissue reinforcement. Ideal properties for high visibility & colonization. The weaving has ideal mesh properties. Sterilized by EtO. Paha® Polypropylene Mesh having properties of;

- **Uniformly Strong**

Strong and durable polypropylene mesh material maintains consistent of strength.

- **Transparent**

Transparent open pore structure in polypropylene mesh facilitate fast incorporation and visualization of underlying tissue structures.

- **Smooth**

In thickness and a low coefficient of friction over predicate polypropylene mesh devices improving ease of use and trocar deployment in laparoscopic procedures.

- **Structure**

Thin wall structure with less surgical mesh material for reducing scar tissue build up and minimizing patient discomfort.

- **Bio-compatible**

Surface area and void area reduction for improving healing and bio-compatibility. ISO 10993 Bio-compatibility results for Polypropylene Mesh are available to customers upon request.

- **Cutting**

Polypropylene Mesh re-sizing done by laser cutting technique due to its versatility and ability to seal the cut edges. Laser cutters can cut any shape and seal the edges while doing so. Sealed edges prevent shedding of cut loops at the fabric edges and do not generate debris in the cutting and post processes.

- **Sterility**

Polypropylene based products can be sterilized using steam autoclave and ethylene oxide (EtO). Polypropylene can not be sterilized using irradiation techniques, such as gamma or beta irradiation as the polymers will be significantly degraded.














- **Classification**

The polypropylene mesh is classified as Class II-b according to Medical Device Directives.

- **We Offer:**

OEM / ODM / Bulk
Standard sizes and shapes
Custom sizes and shapes
Sterile and non-sterile
Custom Labeling and packaging
Standard / Soft polypropylene Mesh
Pre-cut and Pre-shaped Polypropylene Mesh



POLYPROPYLENE MESH RECTANGLE AND SQUARE SHAPES				
CODE	SIZE	INDICATION	SHAPE	
P0220	2 cm * 20 cm	Prolapse		
P0520	5 cm * 20 cm	Prolapse		
P0510	5 cm * 10 cm	Open & Laparoscopic Hernia Repair		
P0611	6 cm * 11 cm	Open & Laparoscopic Hernia Repair		
P0614	6 cm * 14 cm	Open & Laparoscopic Hernia Repair		
P7515	7,5 cm * 15 cm	Open & Laparoscopic Hernia Repair		
P0815	8 cm * 15 cm	Open & Laparoscopic Hernia Repair		
P0914	9 cm * 14 cm	Open & Laparoscopic Hernia Repair		
P1010	10 cm * 10 cm	Open & Laparoscopic Hernia Repair		
P1013	10 cm * 13 cm	Open & Laparoscopic Hernia Repair		
P1015	10 cm * 15 cm	Open & Laparoscopic Hernia Repair		
P1515	15 cm * 15 cm	Open & Laparoscopic Hernia Repair		
P1530	15 cm * 30 cm	Open & Laparoscopic Hernia Repair		
P2020	20 cm * 20 cm	Open & Laparoscopic Hernia Repair		
P2030	20 cm * 30 cm	Open & Laparoscopic Hernia Repair		
P2525	25 cm * 25 cm	Open & Laparoscopic Hernia Repair		
P2535	25 cm * 35 cm	Open & Laparoscopic Hernia Repair		
P3030	30 cm * 30 cm	Open & Laparoscopic Hernia Repair		
P5050	50 CM*50 MTRS	POLYPROPYLENE MESH ROLL		
POLYPROPYLENE MESH PRE - CUT SHAPES				
CODE	SIZE	INDICATION	SHAPE	
PP0505	5 cm * 5 cm	Indirect Inguinal Hernia		
PP0707	7 cm * 7 cm	Indirect Inguinal Hernia		
PP0505-H	5 cm * 5 cm	Indirect Inguinal Hernia		
PP0707-H	7 cm * 7 cm	Indirect Inguinal Hernia		
PP4510	4,5 cm * 10 cm	Open Inguinal Hernia		
PP0611	6 cm * 11 cm	Open Inguinal Hernia		
PP4510-H	4,5 cm * 10 cm	Open Inguinal Hernia		
PP0611-H	6 cm * 11 cm	Open Inguinal Hernia		
PP75125	7,5 cm * 12,5 cm	Rectum Surgery		
PP8515	8,5 cm * 15 cm	Rectum Surgery		
PP1515	15 cm * 15 cm	Hernia Repair		



Confirmation Letter

To:

Center for Public Procurement in Healthcare,
Republic of Moldova

WHEREAS,


We, **ALTAYLAR MEDICAL**, who are manufacturers of **Surgical Polypropylene Mesh**, having factories at **Malikoy MH, Baskent OSB, 19. CD, No: 54 06909 Sincan, Ankara**,

confirm that **Surgical Polypropylene Mesh (codes P0815, P1515, P3030)**, offered by **SOGNO SRL** company on the **Tender no. 21060804 / ocds-b3wdp1-MD-1659013693951**, meet the following tender requirements:

- **The diameter of the Propylene wires is 0.15 mm;**
- **The mesh has bidirectional elasticity.**

For ALTAYLAR MEDICAL,
25.08.2022

Pasa Altay
General Manager



ALTAYLAR
ALTAYLAR MEDİKAL TIBBİ MALZ. İNŞ.
YEKS. GIDA İTH. İHR. SAN. VE TİC. LTD. ŞTİ.
Malikoy Mh, Başkent OSB, No: 54 Sincan ANKARA
Tef:0312 312 04 81-02+Fax:0312 312 04 89+Tic. Sic. No: 23E 351
Polizm V.D. TED 033 4743 711 315 No: 0050 334743 00611

ALTAYLAR MEDİKAL TIBBİ MALZ. İNŞ. TEKS. GIDA. İTH. İHR. SAN. VE TİC. LTD. ŞTİ
Malıköy Mh, Başkent OSB, 19. Cd. No: 54, Malıköy, Sincan, 06909 ANKARA; TURKEY
Phone: +90 444 8 172 Fax: +90 312 502 04 89
Web: www.altaylarmedikal.com E-Mail: info@altaylarmedikal.com