

Product Specifications

CL03.0213

HYDREX

Buffer pH 4.00

Tampão pH 4.00
Puffer pH 4.00
Buffer pH 4.00
Tampón pH 4.00
Tampon pH 4.00
Tampone pH 4.00

**For laboratory use, pH measurements, traceable to SRM from NIST
pH 4.00 ($\pm 0.02/20^{\circ}\text{C}$) Citrate-NaCl-NaOH**

**Traceability: Reference standards acc. NIST SRM 189c, 188, 185i,
186lg, 186llg, 187e**

Mol weight:
Density: 1,01 g/ml
HS Nr: 38229000
CAS Nr:
UN:
ADR:

Specifications:

| | |
|----|---------------|
| pH | = 4.03 @ 0°C |
| pH | = 4.02 @ 10°C |
| pH | = 4.00 @ 20°C |
| pH | = 4.00 @ 25°C |
| pH | = 4.00 @ 30°C |
| pH | = 4.00 @ 40°C |
| pH | = 4.00 @ 50°C |
| pH | = 4.00 @ 60°C |
| pH | = 4.00 @ 70°C |
| pH | = 4.00 @ 80°C |
| pH | = 4.00 @ 90°C |