

# EC Certificate

**mdc medical device certification GmbH**

Notified Body 0483  
herewith certifies that

**REGER Medizintechnik GmbH  
Gewerbstraße 10  
78667 Villingendorf  
Germany**

for the scope

**Instruments and accessories for surgery and electro surgery  
(see attachment)**

has introduced and applies a

**Quality System**

for the design, manufacture and final inspection.

The mdc audit has proven that this quality system  
meets all requirements according to

**Annex II – excluding Section 4  
of the Council Directive 93/42/EEC**

of 14 June 1993 concerning medical devices.

The surveillance will be held as specified in Annex II, Section 5.

Valid from	2018-07-30
Valid until	2023-07-29
Registration no.	D1228500011
Report no.	P18-00423-117158
Stuttgart	2018-07-27



Head of Certification Body



**Attachment of the certificate**

**No. D1228500011**

Date 2018-07-27

Page 1 of 1

Product category	Product	Class	Product code
Instruments and accessories for surgery and electro surgery	HF-Handles (controllable)	IIb	11-499
	HF-Handles (non-controllable)	IIa	11-499
	Bipolar forceps	IIb	11-502
	HF-Electrodes, monopolar and bipolar	IIb	15-579
	Micropump (CapnoPen)	IIa	16-722
	Bipolar Scissors/ Clamp-Scissors	IIb	16-860
	HF-Neutral Electrodes	IIb	11-500



A handwritten signature in blue ink, appearing to read "R. L. C.", is written over a light blue rectangular background.

Head of Certification Body