

## CERTIFICATE OF EC According to EU Directive 93/42/EEC

We herewith declare that the under-mentioned products are in conformity with the essential requirements and provision of Council directive 93/42/EEC as amended by Directive 2007/47/EC. All supporting documentation is retained under the premises of the manufacturer. This declaration of conformity is issued under the sole responsibility of the manufacturer.

Manufacturer Registered Name : **RENOSEM Co., Ltd** Product Name : **Sterilization Agent** 

① GMDN Code No : 44835, Hydrogen peroxide device sterilant

Model Name	: RENO-SA,RENO-SA10,RENO-SA20,RENO-SA90,
	FINO-SA30, FINO-SA130, FINO-SA20
Address	: 2nd~5th Floor.,54,Samjak-ro 133beon-gil,Bucjeon-si,
	Gyeonggi-do, Republic of Korea

Classification : Class IIb (MDD 93/42/EEC, Annex IX Rule 15) Conformity Assessment Route : Annex II, Excluding Section 4, MDD 93/42/EEC As amended by Directive 2007/47/EC Notified Body : SGS Belgium NV

SGS House Noorderlaan 87 2030 Antwerp Belgium

Standards applied : ISO13485:2016, EN ISO14971:2012, EN1041:2008, EN ISO 15223-1:2016, ISO14937:2009

EC Representative : CMC Medical Devices & Drugs S.L. C/Horacio Lengo N° 18, CP 29006, Málaga, Spain Certification No. : KR19/ 81826213 Date of issue : 2022. 2. 22.

**CE**<sub>1639</sub>

Signature : Kump ank Lee

Kwang-Sik Lee / CEO