

EC Certificate

FULL QUALITY ASSURANCE SYSTEM

Directive 93/42/EEC on Medical Devices, Annex II (3)

Certificate Number
41314824

Initial Certification Date
November 1, 2004

Certificate Valid from
August 18, 2009

Certificate Expiry Date
August 18, 2014

We hereby declare that an examination of the under mentioned full quality assurance system has been carried out following the requirements of the Swedish national legislation to which the undersigned is subjected, transposing Annex II (with the exemption of section 4) of the Directive 93/42/EEC on medical devices. We certify that the full quality assurance system conforms with the relevant provisions of the aforementioned directive, and the result entitles the organization to use the CE 0413 marking on those products listed below.

The certification is subject to the organization maintaining their system in compliance with the regulations stated in this certificate, allowing regular assessments and following the contracted requirements of the Notified Body.

Intertek Semko AB is a Notified Body according to Directive 93/42/EEC on medical devices, with identification number 0413.

*Intertek Semko AB
Box 1103, SE-164 22 Kista,
Sweden
Telephone +46 8 750 00 00
medtechsweden@intertek.com*

Organization:

Getinge Disinfection AB

Ljungadalsgatan 11, Box 1505, 351 15 Växjö, Sweden


Product Category:

- Washer Disinfectors

For further identification of the products covered, see the MDD product list/product schedule.

August 18, 2009

Signed date


Marie Olsson, Certification Manager MDD
Intertek Semko AB, Kista, Sweden

Products included in the certificate no: 41314824
 Issued to: **Getinge Disinfection AB**
 Ljungadalsgatan 11
 Box 1505
 SE-351 15 Växjö
 Sweden


Product category	Type/Model designation	Class	Sterile	GMDN code <small>(not mandatory)</small>	Date added
Washer Disinfector	Getinge 46-series, 46-2, 46-3, 46-4, 46-5	IIb	No	17671	* ¹
	Getinge 46-series, 46-4T, 46-5T	IIb	No	17671	Sep 15, 2010
	Getinge 88-Series, 88-5	IIb	No	17671	* ¹
	Getinge 8666, S-8666	IIb	No	17671	* ¹
	Getinge 8668, S-8668	IIb	No	17671	* ¹
	Getinge 9000-Series, 9027	IIa	No	17671	*
	Getinge 9100-Series, 9120, 9122, 9125, 9128	IIb	No	17671	* ¹
	Getinge CM320, CM320	IIb	No	17671	Feb 1, 2011
	Getinge WD 14 Tablo, GE14	IIb	No	17671	Nov 12, 2009 ¹
Getinge WD 15 Claro, GE15	IIb	No	17671	Aug 18, 2009 ¹	
Washer Disinfectors for human waste containers	Getinge FD 1600, Ninjo, FD 1600, FD 1605, FD 1610, FD 1615	IIa	No	35318	May 31, 2011
	Getinge FD 1800, Tornado, FD 1800, FD 1810	IIa	No	35318	May 31, 2011
	Getinge 607, 608, Amigo, S-607, S-608	IIa	No	35318	*
	Getinge 2000, Tornado, SP-1000	IIa	No	35318	*
	Getinge 1200, Ninjo, Ecomat, SP-1200	IIa	No	35318	*
	Getinge 6000, Typhoon, SP-6000	IIa	No	35318	*

* Product added before August 18, 2009.

¹ Product class change from IIa to IIb on 21 March 2010

Date of Issue: May 31, 2011

Intertek Semko AB
 Notified Body MDD



Marie Olsson
 Certification Manager MDD

This product list is only valid together with the referenced, valid EC certificate.

The GMDN codes are assigned by the manufacturer and are only provided for convenience.

Intertek Semko AB is a Notified Body according to the Directive 93/42/EEC on medical devices, with identification number 0413.

Product list for certificate no: 41314824

Date: May 31, 2011

Page 1 of 1

Intertek Semko AB

Torshamnsgatan 43, Box 1103, SE-164 22 Kista, Sweden

Telephone +46 8 750 00 00, Fax +46 8 750 60 30, www.sweden.intertek-etlsemko.com

Registered in Sweden: No SE556024059901, Registered office: As address