



Bio-Rad Laboratories, Inc.

Certificate of Analysis

Material Description: iQ-Check Salmonella II Kit
Material Number: 3578123
Batch Number: 64578603
Expiration Date: 2025-05-07

<u>Characteristic</u>	<u>Results</u>
Reagent A, Lysis Reagent Reagent tested for nucleic acid extraction	Pass
Reagent B, Salmonella Probe Fluorescent probes mix compared to a reference	Pass
Reagent C, Amplification Mix Amplification tested on artificially contaminated samples. Sensitivity compared to a reference amplification solution.	Pass
Reagent D, Negative Control PCR negative control compared to a reference	Pass
Reagent E, Salmonella Pos Ctrl PCR positive control compared to a reference	Pass
Final QC of the Kit Validation on contaminated food samples	Pass

Ranin L Canlas

This certificate has been verified and electronically approved by an authorized Quality representative. This e-signature is performed within a secure system.

2023-11-06 / 20:45:00 UTC

Date/Time