

CERTIFICATE

Full Quality Assurance System

Medical Devices Directive 93/42/EEC Annex II (Excluding Section 4)

Company Name

: Huaian Pingan Medical InstrumentCo., Ltd.

Company Address

: No:128 West Meigao Road, Huaian, Jiangsu, China

Related Directives and Annex

: 93/42/EEC Medical Devices Directive - Annex II

(Excluding Section 4)

Product

- : Sterile, Non-Absorbable Silk Braided Suture with or
- without needle Class IIb - Sterile, Non-Absorbable Polyester Braided Suture
- with or without needle Class Ilb - Sterile, Non-Absorbable Polypropylene monofilament Suture
- with or without needle Class Ilb
- Sterile, Non-Absorbable Nylon / Polyamide monofilament Suture with or without needle - Class Ilb

GMDN

: 13910, 13906, 13909, 13905

Product Types are attached.

Certificate Number

: M.2019.106.11952

Report Number

: MD.3625.IB

Initial Assessment Date

: 09.04.2018

Registration Date

: 14.05.2019

Revision Date /No

Expiry Date

: 13.05.2024

UDEM hereby declares that the requirements of Annex II, excluding section 4 of the 93/42/EEC Directive have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance audits, defined by Annex II, section 5 of the forementioned directive. According to Annex II, section 4 an EC design-examination certificate is required for placing the Class III devices on the market. UDEM's responsibility for class I devices covered by the EC sertificate is limited to manufacturing issues related to safeguarding and maintaining sterile conditions, if the dev,ce is sterile; and manufacturing issues related to product's conformity with metrological requirements, if it has measurement function. This certificate remains as the property of DEM International Certification Auditing Training Centre Industry and Trade Inc. Co. to whom it must be returned upon request. The above named company and UDEM must keep a copy of this certificate for 5 years from the registration of the certificate. Usage of the CE mark is under the responsibility of the manufacturer with the completion of EC Declaration of Conformity. The above mentioned company must notify all changes related with the approved product to UDEM. If UDEM will not renew the expiry date of this certificate in question, thementioned company should stop placing the product on the market. The validity of the certificate can be checked through www.udem.com. tr.

Address: Mutlukent Mahallesi 2073 Sokak (Eski 93 Sokak) No:10 Çankaya – Ankara – TURKEY

Phone: +90 0312 443 03 90 Fax: +90 0312 443 03 76 E-mail: info@udemltd.com.tr www.udem.com.tr

UDEM International Certification **Auditing Training Centre Industry** and Trade Inc. Co.





This document containing 1 (one) pages is the Annex of the Certificate with the number M.2018.106.11952 and with the registration date of 14.05.2019 issued for "Huaian Pingan Medical InstrumentCo., Ltd." by UDEM Uluslararası Belgelendirme Denetim Eğitim Merkezi San. ve Tic. A.Ş. that is giving service as Notified Body with the ID No: 2292 according to 93/42/EEC Medical Devices Directive

Silk braided (SK) - GMDN NO: 13910

Sterile, Non Absorbable, Multifilament, Braided, white or black colored Surgical Sutures are single or loop, with single, double needle or without needle, and in the following USP, EP, Suture Length, Needle Length, Needle Shape and Needle properties:

USP: 8/0, 7/0, 6/0, 5/0, 4/0, 3/0, 2/0, 0, 1, 2, 3, 4

EP: 0.4, 0.5, 0,7, 1,1.5, 2, 3,3.5,4,5,6,7

Suture Lengths: Variety of lengths, from 13 cm to 250 cm

Needle Lengths: Variety of lengths, from 4.5 mm to 90 mm

Needle Shapes: 1/2, 3/8, 1/4, 5/8, Straight, J shape

Needle Properties: Round bodied needle, cutting needle, reverse cutting needle, tapercut needle, blunt needle, micro-point spatula needle

Polyester braided (PB) - GMDN NO: 13906

Sterile, Non Absorbable, Multifilament, Braided, white or green colored Surgical Sutures are single or loop, with single, double needle or without needle, and in the following USP, EP, Suture Length, Needle Length, Needle Shape and Needle properties:

USP: 8/0, 7/0, 6/0, 5/0, 4/0, 3/0, 2/0, 0, 1, 2, 3, 4

EP: 0.4, 0.5, 0,7, 1,1.5, 2, 3,3.5,4,5,6,7

Suture Lengths: Variety of lengths, from 13 cm to 250 cm

Needle Lengths: Variety of lengths, from 4.5 mm to 90 mm

Needle Shapes: 1/2, 3/8, 1/4, 5/8, Straight, J shape

Needle Properties: Round bodied needle, cutting needle, reverse cutting needle, tapercut needle, blunt needle, micro-point spatula needle

Polypropylene monofilament (PM) - GMDN NO: 13909

Sterile, Non Absorbable, Monofilament, white or blue colored Surgical Sutures are single or loop, with single, double needle or without needle, and in the following USP, EP, Suture Length, Needle Length, Needle Shape and Needle properties:

USP: 10/0, 9/0, 8/0, 7/0, 6/0, 5/0, 4/0, 3/0, 2/0, 0, 1, 2, 3, 4

EP: 0.2, 0.3, 0.4, 0.5, 0,7, 1,1.5, 2, 3,3.5,4,5,6,7

Suture Lengths: Variety of lengths, from 13 cm to 250 cm

Needle Lengths: Variety of lengths, from 4.5 mm to 90 mm

Needle Shapes: 1/2, 3/8, 1/4, 5/8, Straight, J shape

Needle Properties: Round bodied needle, cutting needle, reverse cutting needle, tapercut needle, blunt needle, micro-point spatula needle

Nylon / Polyamide monofilament (NL) - GMDN NO: 13905

Sterile, Non Absorbable, Monofilament, black or blue colored Surgical Sutures are single or loop, with single, double needle or without needle, and in the following USP, EP, Suture Length, Needle Length, Needle Shape and Needle properties:

USP: 10/0, 9/0, 8/0, 7/0, 6/0, 5/0, 4/0, 3/0, 2/0, 0, 1, 2, 3, 4

EP: 0.2, 0.3, 0.4, 0.5, 0,7, 1,1.5, 2, 3,3.5, 4, 5, 6, 7

Suture Lengths: Variety of lengths, from 13 cm to 250 cm

Needle Lengths: Variety of lengths, from 4.5 mm to 90 mm

Needle Shapes: 1/2, 3/8, 1/4, 5/8, Straight, J shape

Needle Properties: Round bodied needle, cutting needle, reverse cutting needle, tapercut needle, blunt needle, micro-point spatula needle

Sayfa 1/1

UDEM International Certification Auditing Training Centre Industry and Trade Inc. Co.