

Certificate of Conformity

VWR Catalogue Number	525-0666
Description	VWR®, Centrifuge Tubes, Conical Bottom, 5 ml, natural, With attached lid, sterile
Lot Number	2501033274A
Date of Manufacture (yyyy-mm-dd)	2025-01-03
Date of Sterilization (yyyy-mm-dd)	2025-01-05, 2025-01-06
Expiry Date (yyyy-mm-dd)	2028-01-03
Country of Origin	Manufactured in China
Date of Issue (yyyy-mm-dd)	2026-02-25

Quality System Compliance

This certificate confirms that the product has been sourced from a supplier or manufacturer who has declared their Quality Management System (QMS) complies with **ISO 9001:2015 & ISO 13485:2016**.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Max. Relative Centrifuge force: Products labeled Max. Relative Centrifuge force has been validated on centrifuge force testing. The acceptance level for product with Conical-bottom is 25,000×g.

Autoclave Test: This product has been tested at 121°C for 20 min in a autoclave and it is integrate, no distortion and no leak with cap after autoclave.

Low temperature test: This product has been tested at -80°C for 24 hours in freezer and it is integrate, no distortion and no leak with cap after freezing.

Product Specifications

Material	Tube/Cap - Polypropylene (PP), Compliance with USP Class VI
Sterilization	Product labeled as sterile is EB (Electron beam) irradiated and dose released upon ISO11137 recommended practices in effect at the time of validation. Products labeled sterile meets a minimum requirement of 10 ⁻⁶ SAL with a specified dose range of 15–30 kGy
ATP Assay	Not Applicable.
DNase & RNase Free	This product is free of any detectable DNase/RNase contamination.
BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE) according to EMEA 410/01 revision 3.
Cytotoxicity	Testing is conducted to quality all material resins using ISO 10993 standards for cytotoxicity and have been shown to be non-toxic.
Non-Pyrogenic Statement	The acceptance level for product is 0.05 EU/ml or less than 20 EU/device. (TAL Gel Clot Method). Test passed.
Latex Statement	This product is not made from natural rubber latex.
BPA Statement	Bisphenols are not used in the manufacture of the raw material and are not expected to be present.
DEHP Statement	Not Applicable
RoHS	Raw materials comply with the requirements of the Directive 2011/65/EU, as amended, and (EU) 2015/863 concerning the limits of cadmium, lead, mercury, hexavalent chromium, dibutylphthalate(DBP), Diisobutyl phthalates(DIBP), Butyl benzyl phthalate(BBP), Bis(2-ethylhexyl)phthalate(DEHP), Polybrominated Biphenyls (PBB) and Polybrominated Diphenyl Ethers (PBDE).
REACH Statement	Not Applicable.
Storage Conditions	Store at room temperature.
Shelf Life	3 years.

Disclaimer: VWR's Quality Management System (QMS) is ISO 9001:2015 certified as a wholesale distributor of finished products, ensuring effective oversight of finished product handling, traceability, and distribution processes. VWR is not a manufacturer and does not engage in manufacturing, co-manufacturing or processing activities such as design, packaging, labeling, testing or quality release. Consequently, the supplier or manufacturer's QMS certification may differ from VWR's.

VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. This product is intended for research use only. It is not for use in diagnostic procedures, therapeutic applications, or any human or clinical use.

This document has been produced electronically and is valid without a signature.