

Certificate of Analysis

nerbe plus GmbH & Co. KG hereby certifies that the product group:

GreenTip

Filter tip PP, premium surface

are manufactured under strictly defined and monitored conditions as per product specifications. This here by certifies that the supplied product complies with following standards of production.

Quality System Compliance: The product conforms to written material specifications and was manufactured under the ISO 9001/2015 quality system.

Our raw materials comply with Regulation (EC) No. 1907/2006 (REACH), free of Substances of Very High Concern (SVHC)

Bisphenol A, Phthalates, BSE / TSE, PFAS or Latex are not used in the manufacture process. Our raw materials do not contain Phthalates above 0.1% (w / w).

The Product complies the Regulation (EC) No. 1935/2004 (food contact)

nerbe plus is certified according to EN ISO 13485.

Material Polypropylene		
Test	Limit	Test Result
Humane DNA	< 2 pg; of template copy	Product is free from Human DNA
Endotoxin	≤ 0.03 EU/ ml	Product is free from Endotoxin
DNase	≤ 1.0 X 10 ⁻⁴ U /μl	Product is free from DNase
RNase	≤ 1.0 X 10 ⁻⁸ U / μl	Product is free from RNase
PCR Inhibitor	Less than 2 CT variation	Product is free of PCR Inhibitor

06.03.2025

Quality Management

nerbe plus GmbH & Co. KG

This certificate was issued digitally and is valid without a signature.