



3600.102

Volumetric Flasks, Clear, Class 'A' with Batch Certificate

- Compliance with DIN/ISO 1042 standard.
- Class 'A' with Batch Certificate.
- Printed in Blue.
- With Glass Stopper.
- Calibrated at 20°C.



| Cat. No. | Capacity (mL) | Neck Size | Tolerance (±mL) | Pack of |
|-------------|---------------|-----------|-----------------|---------|
| 3600.102.01 | 1 | 10/19 | 0.025 | 2 |
| 3600.102.02 | 2 | 10/19 | 0.025 | 2 |
| 3600.102.03 | 5 | 10/19 | 0.025 | 2 |
| 3600.102.04 | 10 | 10/19 | 0.025 | 2 |
| 3600.102.05 | 20 | 10/19 | 0.040 | 2 |
| 3600.102.06 | 25 | 10/19 | 0.040 | 2 |
| 3600.102.07 | 50 | 12/21 | 0.060 | 2 |
| 3600.102.08 | 100 | 14/23 | 0.100 | 2 |
| 3600.102.09 | 150 | 14/23 | 0.100 | 2 |
| 3600.102.10 | 200 | 14/23 | 0.150 | 2 |
| 3600.102.11 | 250 | 14/23 | 0.150 | 2 |
| 3600.102.12 | 500 | 19/26 | 0.250 | 2 |
| 3600.102.13 | 1000 | 24/29 | 0.400 | 2 |
| 3600.102.14 | 2000 | 29/32 | 0.600 | 1 |
| 3600.102.15 | 5000 | 34/35 | 1.200 | 1 |

Blue ring optional.

Also available with QR Code and Color Code.



3600.103

Volumetric Flasks, Clear, Class 'B'

- Compliance with DIN/ISO 1042 standard.
- Accuracy as per Class 'B'.
- Printed in Blue.
- With Glass Stopper.
- Calibrated at 20°C.



| Cat. No. | Capacity (mL) | Neck Size | Tolerance (±mL) | Pack of |
|-------------|---------------|-----------|-----------------|---------|
| 3600.103.01 | 1 | 10/19 | 0.05 | 2 |
| 3600.103.02 | 2 | 10/19 | 0.05 | 2 |
| 3600.103.03 | 5 | 10/19 | 0.05 | 2 |
| 3600.103.04 | 10 | 10/19 | 0.05 | 2 |
| 3600.103.05 | 20 | 10/19 | 0.08 | 2 |
| 3600.103.06 | 25 | 10/19 | 0.08 | 2 |
| 3600.103.07 | 50 | 12/21 | 0.12 | 2 |
| 3600.103.08 | 100 | 14/23 | 0.20 | 2 |
| 3600.103.09 | 150 | 14/23 | 0.20 | 2 |
| 3600.103.10 | 200 | 14/23 | 0.30 | 2 |
| 3600.103.11 | 250 | 14/23 | 0.30 | 2 |
| 3600.103.12 | 500 | 19/26 | 0.50 | 2 |
| 3600.103.13 | 1000 | 24/29 | 0.80 | 2 |
| 3600.103.14 | 2000 | 29/32 | 1.20 | 1 |
| 3600.103.15 | 5000 | 34/35 | 2.50 | 1 |

Blue ring optional.

Also available with QR Code and Color Code.