WUXI WEIGHER FACTROY CO., LTD Majiaqiao, Guangyi, Wuxi city, China

CERTIFICATE OF ANALYSIS

	Testing QTY		Contract No.		Testing Result						
Model					Tested QTY		Eligible	Inferior	Not Eligible	Eligible Rate %	Conclusion
Infant Measuring Mat MR-100-A	400		LZ-CBF-119-WHW		400		\checkmark			100%	Permitted to leave factory
Inspection item		r	Testing result	-	ality ass	Remark					Checker
Coating imperfections		no	scratches, cracks or blisters	Eligible							Wang lingbao
Brown/black smundges, Coating pimples,		no	scratches, cracks or blisters	Elig	gible						Wang lingbao
Optical Testing		no	scratches, cracks or blisters	Elig	gible						Wang lingbao
Height range			10-100cm	Elig	gible						Wang lingbao
Min value of height per division			0.5cm	Elig	Eligible						Wang lingbao
Outer packing		Soli	d, no damaged or smundges	Eligible						Wang lingbao	
Handling issues in general			Ok	Eligible						Wang lingbao	
Testing conclusion of Local products									Testing	Date	Aug, 2020
		The products are satisfied with the technical requirements of Checker							Wang lingbao		
		manufacturing drawing, so permitted to leave factory.						Testing	Chief	Deng yuanpei	