



Bio-Rad
Certificate of Analysis

Material Description:	CEFEPIME-CLAV. ACID 30-10 G 4X50 DISKS
Material Number:	12012123
Batch Number:	64577475
Manufacture Date:	2023-11-22
Expiration Date:	2025-12-21

Characteristic

Results

Labeling	Specifications see above	Conforms
----------	--------------------------	----------

Acronym	CFE 40	Conforms
---------	--------	----------

Ink colour	Green	Conforms
------------	-------	----------

.....
Acceptable increase in zone
Diameter for ATB tested alone
versus its zone when tested in
combination with clavulanic acid
.....

Escherichia coli ATCC_25922	2 mm increase	Conforms
-----------------------------	---------------	----------

Klebsiella pneumoniae ATCC_700603	2 mm increase	Conforms
-----------------------------------	---------------	----------

Internal guidelines



Bio-Rad
Certificate of Analysis

Material Description:	CEFEPIME-CLAV. ACID 30-10 G 4X50 DISKS
Material Number:	12012123
Batch Number:	64577475
Manufacture Date:	2023-11-22
Expiration Date:	2025-12-21

Characteristic

Results

Herve Gamba

2023-12-11 / 09:36:33 UTC

This certificate has been verified and electronically approved by an authorized Quality representative. This e-signature is performed within a secure system.

Date/Time