

Certificate of Analysis

Brand: medaSEPT latex PF
 Product: Powder Free Latex Examination Gloves, Textured
 AQL: $\leq 1,5$
 Shape: Ambidextrous
 Color: natural latex
 Packaging: 10 x 100 gloves
 Sizes Assortment: XS,S, M, L, XL

1. WATERTIGHT TEST- EN455:00 Part 1

Inspection level: G1

AQL:1.5

Size	Sample size (pcs)	Acceptance (pcs)	Defects (pcs)	Result
XS	200	7	2	Pass
S	315	10	3	Pass
M	315	10	5	Pass
L	315	10	4	Pass
XL	200	7	3	Pass

2. Dimension EN 455:15 Part 2

Size	XS		S		M		L		XL	
Dimension	Length	Width	Length	Width	Length	Width	Length	Width	Length	Width
Median	240	76	240	84	240	94	240	104	240	112

Size	XS			S			M			L			XL		
Thickness	Cuff	Palm	Finger	Cuff	Palm	Finger	Cuff	Palm	Finger	Cuff	Palm	Finger	Cuff	Palm	Finger
Median	0.07	0.09	0.12	0.07	0.10	0.12	0.07	0.10	0.12	0.07	0.10	0.12	0.07	0.09	0.11

3. Force at break & Elongation at break before accelerated ageing and after accelerated ageing

Size	Median Value														
	Before Accelerated Aging (N)									After Aging (N)					
	Force at break (Min 6 N)									Force at break (Min 6 N)					
XS	7.3									6.7					
S	7.0									6.5					
M	7.2									6.6					
L	8.1									6.9					
XL	8.0									6.8					

4. Protein content test- EN 455:15 PART 3

Requirements: Max 50 $\mu\text{g/g}$

Rozmiar	XS	S	M	L	XL
protein content $\mu\text{g/g}$	49.67	49.23	49.80	49.21	49.75
Result	Pass				

5. Powder residue

Requirements: Max 2mg/glove

Size	Powder content mg/glove	Result
XS	1.0	Pass
S	1.1	
M	1.3	
L	1.2	
XL	1.4	

6. According to EN 455 part 4 above mentioned parameters (after ageing) are guaranteed during whole period of glove shelf life

Dyrektor ds. Zakupów i Produktów
Jowita Boryta-Szymańska
Purchasing & Product Director

MEDASEPT
Spółka z ograniczoną odpowiedzialnością
61-362 Poznań, Forteczna 19
NIP 7822557229 REGON 302347543