



People's Republic of China

Measuring instrument type approval certificate

Hongqi Instrument (Changxing) Co., Ltd.

According to the relevant provisions of Article 13 of the Measurement Law of the People's Republic of China and the Implementation Rules of the Measurement Law of the People's Republic of China, the new measuring instrument products that your unit has applied for type approval have been reviewed and passed, and are now approved and can use the following marks and numbers:



2016F664-33

Approved by: Zhao Duojin

Approved new measuring instrument products (name, model):

The test prototype meets the requirements of the type evaluation program. It is recommended that the following types of measuring

instruments be approved:

Measuring instrument name	model	Specification	level of accuracy
vacuum gauge	Y-60	$(-0.1 \sim 0)$ MPa	Level 2.5
	Y-100	$(-0.1 \sim 0)$ MPa	Level 1.6
	Y-150	$(-0.1 \sim 0)$ MPa	Level 1.6
pressure gauge	Y-60	$(0 \sim 0.1)$ MPa to $(0 \sim 60)$ MPa	Level 2.5
	Y-100	$(0 \sim 0.1)$ MPa to $(0 \sim 60)$ MPa	Level 1.6
	Y-150	$(0 \sim 0.1)$ MPa to $(0 \sim 60)$ MPa	Level 1.6

Blank below

