

Lipase Kit (Enzymatic Colorimetric Assay Method)



Order Information

| Cat. No. | Package size |
|---------------|---------------------------|
| 105-002171-00 | R1: 1×35 mL + R2: 1×9 mL |
| 105-002172-00 | R1: 1×40 mL + R2: 1×10 mL |
| 105-002173-00 | R1: 2×40 mL + R2: 2×10 mL |

Intended Purpose

In vitro test for the quantitative determination of Lipase (LIP) activity in human serum and plasma on Mindray BS series chemistry analyzers. It is intended to be used for aiding to diagnose and monitoring the therapeutic effect of pancreas related diseases.

Summary¹⁻³

Lipases are glycoproteins with a molecular weight of 48000 Da. They are defined as triglyceride hydrolases which catalyze the cleavage of triglycerides to diglycerides with subsequent formation of monoglycerides and fatty acids.

Lipase are produced in the pancreas, also secreted in small amounts by the salivary glands as well as by gastric, pulmonary and intestinal mucosa. The activity of Lipase has been measured as an important marker for diagnosis and monitoring of diseases of the pancreas such as acute and chronic pancreatitis and obstruction of the pancreatic duct.

Assay Principle

Enzymatic Colorimetric Assay Method

1,2-O-dilauryl-rac-glycero-3-glutaric acid-(6-methylresorufin) ester $\xrightarrow{\text{Lipase}}$

1,2-O-dilauryl-rac-glycerol + glutaric acid-(6-methylresorufin) ester

glutaric acid-(6-methylresorufin) ester $\xrightarrow{\text{Spontaneous decomposition}}$ glutaric acid + methylresorufin

The method for the determination of lipase is based on the cleavage of specific chromogenic lipase substrate (1,2-O-dilauryl-rac-glycero-3-glutaric acid-(6-methylresorufin) ester) emulsified in stabilized micro-partides. In the presence of specific activators of pancreatic lipase as colipase, calcium ions and bile acids, the substrate is converted to 1,2-O-dilauryl-rac-glycerol and glutaric acid-(6-methylresorufin) ester which decomposes spontaneously to glutaric acid and methylresorufin. The increase of absorbance due to methylresorufin formation, is proportional to the activity of lipase in the sample.

Reagents Components

| | | |
|------------|--|--------------|
| R1: | Tris | 40 mmol/L |
| | Desoxycholate | 1.8 mmol/L |
| | Taurodesoxycholate | 7.2 mmol/L |
| | Colipase | >1mU/L |
| R2: | Tartrate buffer | 15 mmol/L |
| | Calcium chloride | 0.13 mmol/L |
| | 1,2-O-dilauryl-rac-glycero-3-glutaric acid-(6-methylresorufin) ester | ≥ 0.7 mmol/L |
| | | |

Storage and stability

Up to expiration date indicated on the label, when stored unopened at 2-8°C and protected from light.

On board in use, the reagents are stable for 28 days when refrigerated on the analyzer.

Contamination must be avoided.

Do not freeze the reagents.

Specimen collection and preparation**■ Specimen types**

Serum, lithium heparin plasma are suitable for samples.

■ Preparation for Analysis

1. Use the suitable tubes or collection containers and follow the instruction of the manufacturer; avoid effect of the materials of the tubes or other collection containers.
2. Centrifuge samples containing precipitate before performing the assay.
3. Specimens should be tested as soon as possible after sample collection and pre-analytical treatment.

■ Sample Stability

7 days at 15-25°C

7 days at 2-8°C

2 months at (-25)-(-15)°C (Only frozen once)

For longer storage periods, samples should be frozen at (-20°C)⁴. Sample stability claims were established by manufacturer and/or based upon references, each laboratory should establish its own sample stability criteria.

Reagent Preparation

R1 and R2 are ready to use.

Please perform scheduled maintenance and standard operation including calibration and analysis to assure the performance of measurement system.

Materials required but not provided

1. General laboratory materials: NaCl solution 9 g/L (saline), distilled/deionized water.
2. Calibrator and Control: Please check the section of reagent instruction of Calibration and Quality Control.
3. Mindray BS series chemistry analyzers and General laboratory equipment.

Assay procedure

| Parameters Item | BS-800 chemistry analyzers |
|--|----------------------------|
| Assay type | Kinetic |
| Wavelength (Primary/Secondary) | 570/700 nm |
| Reaction direction | Increase |
| R1 | 200 μ L |
| Sample or Calibrator | 2 μ L |
| Mix, incubate at 37°C for 3 min, then add: | |
| R2 | 40 μ L |
| Mix thoroughly, after incubate at 37°C for 2 min, measure the absorbance value continuously within further 2 min. Then calculate $\Delta A/\text{min}$ | |

Parameters may vary in different chemistry analyzers, may adjust in proportion if necessary. For Mindray BS series chemistry analyzers, Reagent Parameters is available on request. Please refer to the appropriate operation manual for the analyzers.

Calibration

1. It is recommended to use the Mindray Calibrator (Multi Sera Calibrator: 105-001144-00 or other suitable calibrators) and 9 g/L NaCl (saline) for two-point calibration. Traceability of the Mindray Multi Sera Calibrator can refer to the calibrator instructions for use of Mindray Company.

2. Calibration frequency

Calibration is stable for approximately 7 days on BS-800 chemistry analyzers. The calibration stability may vary in different instruments, each laboratory should set a calibration frequency in the instrument parameters appropriate to their usage pattern.

Recalibration may be necessary when the following occur:

- As changed reagent lot.
- As required following quality control procedures or out of control.
- As executes specific maintenance or troubleshooting procedure of chemistry analyzers.

3. The calibrator values are lot-specific with the matched models listed in the value sheet.

Quality control

1. It is recommended to use the Mindray Control (ClinChem Multi Control: 105-009119-00, 105-009120-00 or other suitable controls) to verify the performance of the measurement procedure; other suitable control material can be used in addition.
2. Two levels of control material are recommended to analyze each batch of samples. In addition, the control should be run with each new calibration, with each new reagent cartridge, and after specific maintenance or trouble shooting procedures as detailed in the appropriate system manual.
3. Each laboratory should establish its own internal quality control scheme and procedures for corrective action if control doesn't recover within the acceptable tolerances.

Calculation

The BS series chemistry analyzer detects the change of absorbance ($\Delta A/\text{min}$) and calculates the LIP activity of each sample automatically with a specified calibration curve from calibration process.

Dilution

If the value of sample exceeds 250 U/L, the sample should be diluted with 9 g/L NaCl solution (saline) (e.g. 1+ 9) and rerun; the result should be multiplied by 10.

Expected values

| Sample Type | Units |
|--------------|---------------|
| Serum/Plasma | ≤ 60 U/L |

The expected value is established by Mindray company using 174 serum samples of people from China.

Each laboratory should establish its own reference intervals based upon its particular locale and population characteristics since expected values may vary with geography, race, sex and age.

Performance Characteristics

■ Analytical Sensitivity

The Lipase kit has an analytical sensitivity of 5 U/L on BS-800. Analytical sensitivity is defined as the lowest concentration of analyte that can be distinguished from a sample that contains no analyte. It is calculated as the value lying 3 standard deviations above that of the mean from 20 replicates of an analyte-free sample.

■ Measuring range

The Mindray BS series systems provide the following linearity range:

| Sample Type | Units |
|----------------|-----------|
| Serum / Plasma | 5-250 U/L |

LIP

A high LIP activity sample (approximately 250 U/L) is mixed with a low activity sample (<5 U/L) at different ratios, generating a series of dilutions. The LIP activity of each dilution is determined using Mindray System, the linearity range is demonstrated with the correlation coefficient $r \geq 0.990$. The reportable range is 5-2500 U/L.

■ Precision

Precision was determined by following CLSI Approved Guideline EP05-A3⁵, each sample was assayed 2 times per run, 2 runs per day, a total of 20 days. The precision data of controls on BS-800 are summarized below*.

| Specimen Type (N=80) | Mean (U/L) | Repeatability | | Within-Lab | |
|-------------------------|---------------|---------------|------|------------|------|
| | | SD (U/L) | CV % | SD (U/L) | CV % |
| Control Level 1 | 53.97 | 1.22 | 2.25 | 2.22 | 4.12 |
| Control Level 2 | 87.37 | 1.80 | 2.06 | 2.82 | 3.22 |

*Representative data, results in different instruments, laboratories may vary.

■ Analytical Specificity

The samples with different concentration interfering substance were prepared by addition of interferent to human serum pools, and recovers are within $\pm 10\%$ of the corresponding control value to be considered as no significant interference.

No significant interference was observed when the following substances were tested for interference with this methodology. The data of interference studies on BS-800 are summarized below.

| Interfering Substance | Interferent Concentration (mg/dL) | Analyte Concentration (U/L) | Relative Deviation (%)* |
|-----------------------|-----------------------------------|-----------------------------|-------------------------|
| Ascorbic acid | 30 | 91.27 | +0.55 |
| Hemoglobin | 150 | 92.77 | -0.25 |
| Bilirubin | 40 | 94.30 | -2.12 |
| Intralipid | 500 | 95.13 | +4.55 |

*Representative data, results in different instruments, laboratories may vary.

Acetaminophen metabolite N-acetyl-p-benzoquinone imine (NAPQI) and Acetylcysteine that is frequently used as an antidote to Acetaminophen intoxication may cause falsely low results independently.

In very rare cases gammopathy, in particular type IgM, may cause unreliable results⁶.

■ Method Comparison

Correlation studies were performed using CLSI Approved Guideline EP09-A3⁷. The Mindray System (Mindray BS-2800M/Mindray LIP Reagent)

(y) was compared with comparison system (Mindray BS-2000 / Mindray LIP Reagent) (x) using the same serum specimens. The statistical data obtained by linear regression are shown in the table below *:

| Regression Fit | Correlation Coefficient (r) | Sample (N) | Concentration Range (U/L) |
|--------------------|-----------------------------|------------|---------------------------|
| $y=1.0248x+0.0742$ | 0.9997 | 116 | 5.93-241.83 |

*Representative data, results in different instruments, laboratories may vary.

Result interpretation

The results could be affected by drugs, disease, or endogenous substances^{6,8}. When the reaction curve is abnormal, it is recommended to retest and check the result.

Warnings and precaution

1. For in vitro diagnostic use only. For laboratory professional use.
2. Please take the necessary precautions for handling all laboratory reagents.
3. Please confirm the integrity of the package before use. Do not use the kits with damaged packages. The reagents avoid direct exposure to sunlight and freezing. The results can't be assured when stored at inappropriate condition.
4. If unintentionally opened before used, store the reagents tightly capped at 2-8°C and protected from light, and the stability is equally to in-use stability.
5. Do not mix reagents with different lots and bottles.
Do not use the reagents beyond the expiration date and the in-use date.
Do not mix fresh reagents with in-use reagents.
Avoid the formation of foam.
6. Instability or deterioration should be suspected if there are visible signs of leakage, precipitates or microbial growth, or if calibration or controls do not meet the insert and/or the Mindray System criteria.
7. Reliability of assay results cannot be guaranteed if the instructions in this package insert are not followed.
8. Do not swallow. Avoid contact with skin and mucous membranes.
9. When the reagents accidentally enter the eyes and mouth, or contact with the skin, immediately wash with plenty of water. If necessary, visit the doctor for further medical treatment.
10. Safety data sheet is available for professional user on request.
11. Disposal of all waste material should be in accordance with local guidelines.
12. All human material should be considered potentially infectious.

13. All identified risks have been reduced as far as possible without adversely affecting the benefit-risk ratio, and the overall residual risk is acceptable.
14. Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

References

1. Burtis CA, Ashwood ER, Bruns DE, eds. Tietz Textbook of Clinical Chemistry and Molecular Diagnostics, 5th ed. Elsevier Saunders 2012;585-587.
2. Thomas L. Clinical Laboratory Diagnostics: Use and Assessment of Clinical Laboratory Results. 1st ed. Frankfurt: TH-Books Verlagsgesellschaft 1998; 95-97.
3. McPherson RA, Pincus MR. Henry's Clinical Diagnosis and Management by Laboratory Methods, 22nd ed. Elsevier Saunders 2007;314-315, 440-441.
4. CLSI. Procedures for the handling and processing of blood specimens; Approved Guideline-Third Edition. CLSI document H18-A3. Wayne, PA: Clinical and Laboratory Standards Institute; 2004.
5. CLSI. Evaluation of Precision of Quantitative Measurement Procedures; Approved Guideline-Third Edition. CLSI document EP05-A3. Wayne, PA: Clinical and Laboratory Standards Institute; 2014.
6. Bakker AJ, Mucke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. Clin Chem Lab Med, 2007,45(9):1240-1243.
7. CLSI. Measurement Procedure Comparison and Bias Estimation Using Patient Sample; Approved Guideline-Third Edition. CLSI document EP09-A3. Wayne, PA: Clinical and Laboratory Standards Institute; 2013.
8. Young DS. Effects of Drugs on Clinical Laboratory Tests, 5th ed. Washington, DC: AACC Press; 2000:3-526,3-528.

Graphical symbols



In Vitro Diagnostic
medical device



Unique device
identifier



European
Conformity



Consult Instructions
For use



Use-by
date

LIP

mindray



Authorized representative in the European Community



Batch Code



Temperature limit



Manufacturer



Catalogue number



Keep away from sunlight

Indicates a medical device that needs protection from light sources

© 2022-2024 Shenzhen Mindray Bio-Medical Electronics Co., Ltd. All rights Reserved

Manufacturer: Shenzhen Mindray Bio-Medical Electronics Co., Ltd.

Address: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P. R. China

E-mail Address: service@mindray.com

Website: www.mindray.com

Tel: +86-755-81888998; **Fax:** +86-755-26582680

EC-Representative: Shanghai International Holding Corp. GmbH(Europe)

Address: Eiffestraße 80, Hamburg 20537, Germany

Tel: 0049-40-2513175; **Fax:**0049-40-255726

LIP

Набор для определения активности липазы (LIP), ферментативный колориметрический метод



Информация для оформления заказа

| № кат. | Размер упаковки |
|---------------|---------------------------|
| 105-002171-00 | R1: 1×35 мл + R2: 1×9 мл |
| 105-002172-00 | R1: 1×40 мл + R2: 1×10 мл |
| 105-002173-00 | R1: 2×40 мл + R2: 2×10 мл |

Целевое назначение

Анализ *in vitro* для количественного определения активности липазы (LIP) в сыворотке и плазме крови человека на биохимических анализаторах Mindray серии BS. Данный анализ предназначен для использования в диагностике и мониторинга эффекта лечения заболеваний, связанных с поджелудочной железой.

Краткая справка¹⁻³

Липазы — это гликопротеины с молекулярной массой 48 000 Да. Они определяются как гидролазы триглицерида, которые катализируют расщепление триглицеридов в диглицериды с последующим образованием моноглицеридов и жирных кислот.

Липаза вырабатывается в поджелудочной железе, и, кроме того, выделяется в малых количествах слюнными железами, а также слизистой оболочкой желудка, легких и кишечника. Активность липазы измеряется в качестве важного маркера для диагностики и наблюдения за течением заболеваний поджелудочной железы, например, острого и хронического панкреатита, и непроходимости протока поджелудочной железы.

Принцип анализа

Энзиматический колориметрический метод определения

1,2-О-дилаурил-рац-глицеро-3-глутаровая кислота-(6-метилрезорурфин)-эфир $\xrightarrow{\text{Липаза}}$

1,2-О-дилаурил-рац-глицерин + глутаровая кислота-(6-метилрезорурфин)-эфир

глутаровая кислота-(6-метилрезорурфин)-эфир $\xrightarrow{\text{Саморазложение}}$
глутаровая кислота + метилрезорурфин

Метод определения липазы основан на расщеплении специфического хромогенного субстрата липазы

(1,2-О-дилаурил-рац-глицеро-3-глутаровая кислота-(6-метилрезорурфин)-эфир), эмульгированного в стабилизированные микрочастицы. В присутствии специфических активаторов липазы поджелудочной железы, таких как колипаза, ионы кальция и жёлчные кислоты, этот субстрат преобразуется в 1,2-О-дилаурил-рац-глицерин и глутаровую кислоту (6 метилрезорурфин)-эфир, который самопроизвольно разлагается на глутаровую кислоту и метилрезорурфин. Повышение оптической плотности, обусловленное образованием метилрезорурфина, пропорционально активности липазы в пробе.

Реагенты и компоненты

| | | |
|------------|--|---------------|
| | Трис | 40 ммоль/л |
| R1: | Дезоксихолат | 1,8 ммоль/л |
| | Тауродезоксихолат | 7,2 ммоль/л |
| | Колипаза | > 1 МЕд/л |
| | Тартратный буфер | 15 ммоль/л |
| R2: | Хлорид кальция | 0,13 ммоль/л |
| | 1,2-О-дилаурил-рац-глицеро-3-глутаровая кислота-(6-метилрезорурфин)-эфир | ≥ 0,7 ммоль/л |

Хранение и стабильность

До истечения срока годности, указанного на этикетке, при хранении в нераспечатанном виде при 2-8°C в защищенном от света месте.

При использовании в анализаторе реагенты стабильны в течение 28 суток при хранении в охлажденном состоянии на анализаторе.

Необходимо избегать загрязнения.

Запрещается замораживать реагенты.

Отбор и подготовка образцов

■ Типы образцов

В качестве проб можно использовать сыворотку, плазму с гепарином.

■ Подготовка перед анализом

1. Необходимо использовать подходящие пробирки или контейнеры для сбора проб и следовать инструкциям изготовителя; избегать воздействия материалов пробирок или других контейнеров для сбора проб.
2. Пробы, содержащие осадок, перед проведением анализа необходимо центрифугировать.
3. Образцы после сбора и предварительной обработки следует проанализировать как можно скорее.

■ Стабильность проб

7 дней при 15-25°C

7 дней при 2-8°C

2 месяца при (-25)-(-15)°C (при однократной заморозке)

Для более длительного хранения проб необходимо замораживать их при температуре (-20°C)⁴. Требования к стабильности образцов были установлены изготовителем и/или основаны на эталонах, каждая лаборатория должна устанавливать свои собственные критерии стабильности образцов.

Подготовка реагентов

Реагенты R1 и R2 готовы к использованию.

Выполняйте плановое техническое обслуживание и стандартные операции, включая калибровку и анализ системы, чтобы гарантировать работоспособность аналитической системы.

Необходимые материалы, которые не представлены в наборе

1. Обычные лабораторные материалы: Раствор NaCl в концентрации 9 г/л (физиологический раствор), дистиллированная/деионизованная вода.
2. Калибратор и Контроль: Ознакомьтесь с разделом инструкции по применению реагентов «Калибровка и контроль качества».
3. Химические анализаторы Mindray серии BS и лабораторное оборудование общего назначения.

Методика количественного анализа

| Параметры | Химические анализаторы BS-800 |
|--|-------------------------------|
| Тип анализа | Кинетическая |
| Длина волны (первичная/вторичная) | 570/700 нм |
| Направление реакции | Увеличение |
| R1 | 200 мкл |
| Проба или калибратор | 2 мкл |
| Перемешайте, инкубируйте при 37°C в течение 3 минут, затем добавьте: | |
| R2 | 40 мкл |
| Тщательно перемешайте, затем инкубируйте при 37°C в течение 2 минут, и в течение следующих 2 минут непрерывно измеряйте значение оптической плотности. Затем рассчитайте $\Delta A/\text{мин}$ | |

Параметры могут отличаться на разных химических анализаторах, при необходимости их можно пропорционально корректировать. Для химических анализаторов Mindray серии BS параметры для реагентов предоставляются по запросу. Обратитесь к соответствующему руководству по эксплуатации этих анализаторов.

Калибровка

1. Рекомендуется использовать калибратор Mindray (Сывороточный мультикалибратор: 105-001144-00 или другие подходящие калибраторы) и 9 г/л NaCl (физраствор) для двухточечной калибровки. Информацию о прослеживаемости Сывороточного мультикалибратора Mindray см. в инструкции по использованию калибратора компании Mindray.

2. Частота калибровки

Калибровка химических анализаторов BS-800 стабильна в течение приблизительно 7 суток. Для разных приборов стабильность калибровки может отличаться, каждой лаборатории следует установить частоту калибровки в параметрах прибора в соответствии со своим режимом использования.

Может понадобиться повторная калибровка при возникновении следующих обстоятельств:

- При изменении партии реагента.
- Согласно требованию соблюдаемых процедур контроля качества или при выходе значений для контрольного материала за допустимые пределы.
- При выполнении конкретной процедуры по техническому обслуживанию или устранению неисправности биохимических анализаторов.

3. Значения для калибратора зависят от партии и соответствующих моделей, указанных в таблице значений.

Контроль качества

1. Рекомендуется использовать контроль Mindray (Клинико-химический мультиконтроль: 105-009119-00, 105-009120-00 или другой подходящий контрольный материал) для проверки эффективности процедуры измерения; дополнительно можно использовать другие подходящие контрольные материалы.

2. Рекомендуется использовать два уровня контрольных материалов для анализа каждой партии проб. Кроме того, контрольный материал следует анализировать для каждой новой калибровки, каждого нового картриджа с реагентами и после определенного технического обслуживания или устранения неисправностей, как подробно описано в соответствующем руководстве по эксплуатации системы.

3.В каждой лаборатории следует установить собственную схему контроля качества и порядок выполнения корректирующих действий, если контрольные материалы не восстанавливаются в пределах допустимых отклонений.

Расчет

Химический анализатор серии BS определяет изменение поглощения (ДА/мин) и автоматически рассчитывает активность LIP для каждой пробы с указанной калибровочной кривой, полученной в процессе калибровки.

Разведение

Если значение пробы превышает 250 ЕД/л, пробу необходимо разбавить раствором NaCl в концентрации 9 г/л (физраствором) (например, 1+9) и повторить анализ; полученный результат следует умножить на 10.

Предполагаемые значения

| Тип пробы | Единицы измерения |
|------------------|-------------------|
| Сыворотка/плазма | ≤ 60 ед/л |

Предполагаемое значение установлено компанией Mindray с помощью 174 образцов сыворотки, взятых у людей из Китая.

Каждая лаборатория должна установить свои собственные референсные интервалы в зависимости от ее конкретного расположения и популяционных характеристик, поскольку предполагаемые значения могут отличаться в зависимости от географии, расы, пола и возраста.

Рабочие характеристики

■ Аналитическая чувствительность

Аналитическая чувствительность набор для определения активности липазы (LIP) на анализаторе BS-800 составляет 5 ед./л. Аналитическая чувствительность определяется как наименьшая концентрация аналита, по которой можно отличить образец, не содержащий этот аналит. Она рассчитывается как значение, на 3 стандартных отклонения превышающее среднее значение, полученное из 20 повторных анализов пробы, не содержащего анализируемого вещества.

■ Диапазон измерений

Системы Mindray серии BS обеспечивают следующий диапазон линейности:

| Тип пробы | Единицы измерения |
|------------------|-------------------|
| Сыворотка/плазма | 5-250 ед./л |

Проба с высокой активностью LIP (приблизительно 250 Ед/л) смешивается с пробой с низкой активностью (< 5 Ед/л) в разных соотношениях, в результате получается серия разведений. LIP каждого разведения определяется с помощью системы Mindray, диапазон линейности демонстрируется с коэффициентом корреляции $r \geq 0,990$. Регистрируемый диапазон составляет 5-2500 ед./л.

■ Прецизионность

Прецизионность определялась с помощью следующего одобренного CLSI EP05-A3⁵, каждая проба анализировалась по 2 раза за одну обработку, 2 обработки в сутки, всего 20 суток.

Данные прецизионности контрольных материалов на BS-800 приведены ниже*.

| Тип образцов (N=80) | Среднее (Ед/л) | Воспроизводимость | | Внутрилабораторная | |
|-----------------------|----------------|-------------------|------|--------------------|------|
| | | SD (Ед/л) | CV % | SD (Ед/л) | CV % |
| Контрольный уровень 1 | 53,97 | 1,22 | 2,25 | 2,22 | 4,12 |
| Контрольный уровень 2 | 87,37 | 1,80 | 2,06 | 2,82 | 3,22 |

*Репрезентативные данные, результаты, полученные на разных приборах и в разных лабораториях могут отличаться.

■ Аналитическая специфичность

Пробы с различной концентрацией интерферирующего вещества готовили путем добавления мешающего вещества к пулам человеческой сыворотки, и отсутствием значимого мешающего воздействия считалось восстановление в пределах $\pm 10\%$ от соответствующего контрольного значения.

Не наблюдали значимого мешающего воздействия указанных далее веществ при исследованиях вместе с ними с использованием данной методологии. Данные исследования мешающих материалов на BS-800 приведены ниже.

| Мешающее вещество | Мешающая концентрация (мг/дл) | Концентрация аналита (Ед/л) | Относительное отклонение (%)* |
|----------------------|-------------------------------|-----------------------------|-------------------------------|
| Аскорбиновая кислота | 30 | 91,27 | +0,55 |
| Гемоглобин | 150 | 92,77 | -0,25 |
| Билирубин | 40 | 94,30 | -2,12 |
| Интралипид | 500 | 95,13 | +4,55 |

**Репрезентативные данные, результаты, полученные на разных приборах и в разных лабораториях могут отличаться.*

Ложные низкие результаты независимо друг от друга могут быть вызваны метаболитом ацетаминофена N-ацетил-п-бензохинониминим (NAPQI), а также N-ацетилцистеином, часто используемым в качестве антидота при интоксикации ацетаминофеном.

В очень редких случаях гаммопатия, в частности тип IgM, может привести к ненадежным результатам⁶.

■ Сравнение методов

Исследования корреляции выполнялись с использованием одобренного CLSI руководства EP09-A3⁷. Система Mindray (анализатор Mindray BS-2800M/реагент Mindray LIP) (y) была сопоставлена с системой сравнения (анализатор Mindray BS-2000/реагент Mindray LIP) (x) с использованием одинаковых образцов сыворотки. Статистические данные, полученные линейной регрессией, показаны в таблице ниже*:

| Уравнение регрессии | Коэффициент корреляции (r) | Проба (N) | Диапазон активности (Ед/л) |
|------------------------|----------------------------|-----------|----------------------------|
| $y = 1,0248x + 0,0742$ | 0,9997 | 116 | 5,93-241,83 |

**Репрезентативные данные, результаты, полученные на разных приборах и в разных лабораториях могут отличаться.*

Интерпретация результатов

На результаты могут влиять лекарственные средства, заболевания или эндогенные вещества^{6,8}. Если график реакции отклоняется от нормального, рекомендуется провести повторный анализ и проверить результат.

Предупреждения и меры предосторожности

1. Только для диагностики *in vitro*. Для профессионального лабораторного применения.
2. Необходимо соблюдать меры предосторожности при работе со всеми лабораторными реагентами.
3. Убедитесь в целостности упаковки перед использованием набора. Не используйте наборы с поврежденной упаковкой. Исключите воздействие на реагенты прямого солнечного света и их замораживание. При несоблюдении соответствующих условий хранения реагентов получение корректных результатов анализов не гарантируется.
4. Если реагенты были непреднамеренно открыты до использования, храните их плотно закрытыми при температуре 2-8°C и защищенными от света. В этом случае их стабильность не отличается

- от стабильности при использовании.
5. Не допускайте смешивания реагентов из разных партий и флаконов. Не используйте реагенты после истечения их срока годности и даты использования. Не допускайте смешивания свежих реагентов с уже используемыми.
Избегайте образования пены.
 6. Следует ожидать потерю стабильности или ухудшение качества при наличии видимых признаков утечки, выпадения осадков или роста микроорганизмов, а также если калибровка или контрольные материалы не соответствуют критериям, указанным в листке-вкладыше и/или для системы Mindray.
 7. Надежность результатов анализа не гарантируется в случае несоблюдения инструкций, приведенных в данном листке-вкладыше.
 8. Запрещается проглатывать. Избегайте контакта с кожей и слизистыми оболочками.
 9. При случайном попадании реагента в глаза или ротовую полость, а также на кожу немедленно обильно промойте пораженные участки большим количеством воды. При необходимости обратитесь к врачу для получения медицинской помощи.
 10. Паспорт безопасности материала предоставляется профессиональному пользователю по запросу.
 11. Утилизация всех отходов должна производиться в соответствии с местными правилами.
 12. Весь человеческий материал следует считать потенциально инфекционным.
 13. Все выявленные риски были снижены настолько, насколько это возможно без негативного влияния на соотношение пользы и риска, и общий остаточный риск является приемлемым.
 14. О любом серьезном инциденте, произошедшем в связи с этим устройством, необходимо сообщать производителю и компетентному органу страны, к которой относится пользователь и/или пациент.

Литература

1. Burtis CA, Ashwood ER, Bruns DE, eds. Tietz Textbook of Clinical Chemistry and Molecular Diagnostics, 5th ed. Elsevier Saunders 2012;585-587.
2. Thomas L. Clinical Laboratory Diagnostics: Use and Assessment of Clinical Laboratory Results. 1st ed. Frankfurt: TH-Books Verlagsgesellschaft 1998; 95-97.
3. McPherson RA, Pincus MR. Henry's Clinical Diagnosis and Management by Laboratory Methods, 22nd ed. Elsevier Saunders 2007;314-315, 440-441.

4. CLSI. Procedures for the handling and processing of blood specimens; Approved Guideline-Third Edition. CLSI document H18-A3. Wayne, PA: Clinical and Laboratory Standards Institute; 2004.
5. CLSI. Evaluation of Precision of Quantitative Measurement Procedures; Approved Guideline-Third Edition. CLSI document EP05-A3. Wayne, PA: Clinical and Laboratory Standards Institute; 2014.
6. Bakker AJ, Mucke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. Clin Chem Lab Med, 2007;45(9):1240-1243.
7. CLSI. Measurement Procedure Comparison and Bias Estimation Using Patient Sample; Approved Guideline-Third Edition. CLSI document EP09-A3. Wayne, PA: Clinical and Laboratory Standards Institute; 2013.
8. Young DS. Effects of Drugs on Clinical Laboratory Tests, 5th ed. Washington, DC: AACC Press; 2000:3-526,3-528.

Условные обозначения



In Vitro Diagnostic
medical device



Unique device
identifier



European
Conformity



Consult Instructions
For use



Use-by
date



Authorized representative in
the European Community



Batch Code



Temperature
limit



Manufacturer



Catalogue
number



Keep away from sunlight

Indicates a medical device that needs protection from light sources

© 2022-2024 Shenzhen Mindray Bio-Medical Electronics Co., Ltd. Все права защищены.

Производитель: Shenzhen Mindray Bio-Medical Electronics Co., Ltd.

Адрес: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P. R. China

Адрес электронной почты: service@mindray.com

Веб-сайт: www.mindray.com

Тел.: +86-755-81888998; **Факс:** +86-755-26582680

Представитель в ЕС: Shanghai International Holding Corp. GmbH
(Европа)

Адрес: Eiffestraße 80, Hamburg 20537, Германия

Тел.: 0049-40-2513175; **Факс:** 0049-40-255726