

HiMedia Laboratories Private Limited

23, Vadhani Industrial Estate, L.B.S. Marg, Mumbai - 400086

Website : www.himedialabs.com, Email : info@himedialabs.com

Certificate of Analysis , Quality and Conformity

Material Code : SD028	Material Name : Penicillin-G P 10 units	Lot No : 0000418174
AR No.: 040000984538	Date of Report : 2020-01-08	Exp. Date : 2021-06

TEST	SPECIFICATIONS	RESULTS
Appearance 1 Appearance	Filter paper discs of 6 mm diameter with printed "P 10" on centre of each side of the disc.	Complies
Cultural response 1 Cultural response	Average diameter of zone of inhibition observed on Mueller Hinton Agar (M173) after 18 hours incubation at 35-37°C for standard cultures.	
Staphylococcus aureus ATCC 25923 1 Zone of inhibition (mm)	26mm - 37mm	35mm

The performance of this batch has been tested and standardised as per the CLSI (formerly NCCLS) Standards.

STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current HiMedia manual or product sheets.

This document has been produced electronically and is valid


Sukanya Deshmukh

Microbiologist/Sr.Executive Microbiologist


Ujwala M. Kokate

Asst./Dy/QC Manager


Dr. Santosh Kaul

Dy/QA Manager