

Specificații tehnice

[Acest tabel va fi completat de către ofertant în coloanele 2, 3, 4, 6, 7, iar de către autoritatea contractantă – în coloanele 1, 5.]

Numărul procedurii de achiziție : ocds-b3wdp1-MD-1664371848792 din 28.09.2022						
Obiectul achiziției: Echipamente medicale						
Denumirea serviciului	Denumirea modelului serviciului	Țara de origine	Producătorul	Specificarea tehnică deplină solicitată de către autoritatea contractantă	Specificarea tehnică deplină propusă de către ofertant	Standarde de referință
1	2	3	4	5	6	7
Bunuri						
Lot I. Respirator, FFP2/N95	Respirator FFP2	ASTEK METAL MEDİKAL YAPI TEK. SAN. TİC	Turcia	<p>Single use, disposable, non-fluid resistant, for medical purposes respirator mask protecting against airborne pathogens (such as mycobacterium tuberculosis, SARS). Good breathability with design that does not collapse against the mouth (e.g. duckbill or cup-shaped). With elastic ear loops, easy to put on and provide a secure fit. Adjustable nose clip. Respirator mask fits all face shapes, without inspiration / expiration air-leakage. Without valve. Color: white. Non-sterile. Made from non-woven fabric. Mask comes with clear indication on applicable standard and instructions for use. <i>(Pack of maximum 30-50 units/pack)</i></p>	<p>Single use, disposable, non-fluid resistant, for medical purposes respirator mask protecting against airborne pathogens (such as mycobacterium tuberculosis, SARS). Good breathability with design that does not collapse against the mouth cup-shaped. With elastic ear loops, easy to put on and provide a secure fit. Adjustable nose clip. Respirator mask fits all face shapes, without inspiration / expiration air-leakage. Without valve. Color: white. Non-sterile. Made from non-woven fabric. Mask comes with clear indication on applicable standard and instructions for use. <i>(Pack 10 units/pack)</i></p>	<p>Declarație de conformitate ISO 9001 ISO 13485 ISO 45001 ISO14001 ISO10002</p>

				<p>The quality, performance characteristics and related standards according to the WHO Technical specifications of personal protective equipment for COVID-19: Good particle filtration, minimum 94% or 95%. Air permeability: > 2 mm H₂O. Approved to standards N95/FFP2. EN 149 minimum FFP2 (European Standard EN149:2001)</p>	<p>The quality, performance characteristics and related standards according to the WHO Technical specifications of personal protective equipment for COVID-19: Good particle filtration, minimum 95%-98% Air permeability: > 2 mm H₂O. Approved to standards N95/FFP2. EN 149 minimum FFP2 (European Standard EN149:2001)</p>	
--	--	--	--	--	--	--

Semnat: _____ Numele, Prenumele: Cobzari-Turcan Rodica

În calitate de: Administrator Ofertantul: "Aelo Grup" SRL

Adresa: Mun.Chisinau, str. Petru Movila 23/5, ap (of)3